Title: **ATRIAL SENSING AND MULTIPLE SITE STIMULATION AS INTERVENTION FOR ATRIAL FIBRILLATION**

Inventor: Morton M. Mower, M.D.

Field of the Invention

The present invention relates generally to electronic stimulation devices to control the beating of hearts, especially hearts with pathologies that interfere with normal rhythmicity, electrical conduction, and/or contractility. In particular, the present invention relates to pacemakers used to overcome atrial fibrillation by use of 1) atrial sensing; 2) electrical test stimulation of the atria; and 3) multiple site stimulation in which the various atrial areas are slowly entrained to a common beating rate to produce electrical/functional conformity, i.e., cardioversion, with each case either eventuating in spontaneous reversion to a normal atrial rhythm, or reduced energy requirement for reversion by electrical countershock.

Background of the Invention

Morbidity associated with malfunctions of the atria, while not immediate, is high. Atrial malfunctions of rhythmicity (e.g., atrial fibrillation, various atrial arrhythmias, A-V block and other conduction abnormalities, etc.) can contribute to thrombosis, emboli, stroke and/or heart failure, each of which can place a patient in significant peril.

Atrial Sensing. A variety of approaches have been developed which use pacemakers to counter atrial malfunctions of rhythmicity, as well as attendant effects on ventricular function. In addition, sophisticated approaches have been developed for pacemaker systems to determine the nature of any particular ventricular malfunction, and whether a malfunction originates in the atria or in the ventricles. One such approach uses ventricular sensing to measure/determine the probability density function (pdf) on a moment-to-moment basis. For example, U.S. Patent No. 5,163,429 to Cohen discloses the

use of narrow window pdf data as but one criterion among several for assessing ventricular cardiac function. The use of pdf data to determine ventricular fibrillation also is disclosed in *Implantable Cardioverter-Defibrillators* (N.A. Estes III, A. Manolis & P. Wang, ed.). U.S. Patent No. 5,421,830 to Epstein, et al. (discussed further below) also discloses the use of pdf data as one set among a variety of data types that collectively are also used to assess cardiac function. The use of probability density function data for assessing atrial cardiac function has not been disclosed and presents its own unique difficulties as will be further discussed.

Electrical Test Stimulation of Atria. In a few limited cases, pacemaker protocols have been employed in which electrical test stimuli are applied to the atria, and the physiological responses thereto are monitored to aid in the determination of the best or most appropriate protocol to initiate, cure, or ameliorate the existing cardiac malfunction. For example, U.S. Patent No. 5,620,471 to Duncan discloses three basic protocols for determining whether observed ventricular irregularities are actually caused by atrial arrhythmias. One protocol includes atrial electrical test stimulation, and all three protocols monitor both atrial and ventricular rhythms for three parameters: rates of atrial and ventricular firing, stability of firing/beating in atria and ventricles, and whether or not ventricular firing tracks atrial firing. In the first protocol, when the ventricular firing rate is less than the atrial firing rate (indicating no ventricular tracking of atrial beats), and firing rates are stable, then ventricular tachycardia is presumed, and ventricular stimulation is applied. On the other hand (second protocol), if the ventricular firing rate is not stable, then atrial arrhythmia is presumed, and atrial stimulation is applied. The third protocol is based on the fact that, when the ventricular firing rate equals the atrial firing rate, there

may or may not be ventricular tracking of atrial firing. Whether or not there is ventricular tracking is determined by the presence or not of ventricular tracking following premature atrial stimulation by the pacemaker. If there is ventricular tracking of atrial firing, the arrhythmic mechanism is presumed to be atrial tachycardia. However, if there is no ventricular tracking of atrial firing, then ventricular tachycardia is presumed, and ventricular stimulation is performed.

U.S. Patent No. 5,421,830 to Epstein, et al. discloses a general means for recording, testing, and analyzing cardiac function based on data from -- and electrical test stimulation via -- a patient's pacemaker, as well as data from additional sensors detecting hemodynamic or other body functions. Total intracardiac electrograms (reflecting both atrial and ventricular functional status) or just selected data (e.g., P-P or R-R intervals, heart rate, arrhythmia duration, slew rate, probability density function, etc.) may be recorded and analyzed. The patient's atrial and ventricular responses to electrical test pulses may also be recorded. In sum, this system provides a means to more easily tailor settings for pacemakers to achieve optimal settings for the specific patient or for the specific situation (e.g., during exercise or exertion) of a given patient.

U.S. Patent No. 5,215,083 to Drane, et al. also discloses the use of electrical test stimulation to aid in the fine tuning and evaluation of different possible stimulation protocols for a patient's heart. In particular, electrical test pulses are employed to induce ventricular fibrillation or tachycardia for use in evaluating the effectiveness of alternative programmed therapies.

Multiple Site Atrial Stimulation. The use of multiple site atrial stimulation has been disclosed for various purposes, such as defibrillation, cardioversion, pacing, and dc

field production. One example is provided by U.S. Patent No. 5,562,708 to Combs, et al., which discloses the employment of large surface electrodes (each effectively comprising multiple electrodes) that are implanted to one or both atria for providing extended, low energy electrical impulses. The electrical impulses are applied simultaneously at multiple sites over atrial surfaces, and atrial fibrillation is interrupted by gradually entraining greater portions of atrial tissue. These pacemaker electrodes may be used for various purposes in addition to pacing, such as conventional defibrillation and cardioversion.

U.S. Patent No. 5,649,966 to Noren, et al. discloses the use of multiple electrodes for the purpose of applying a subthreshold dc field to overcome fibrillation. The rate of application of the dc field is sufficiently low so that no action potential is triggered. Polarity may also be changed periodically. In one embodiment, four electrodes are positioned within a single plane in the heart, which permits a dipole field in virtually any direction within that plane.

U.S. Patent No. 5,411,547 to Causey, III discloses the use of sets of complex mesh patch electrodes, in which each electrode comprises an anode patch and a cathode patch, for purposes of cardioversion-defibrillation. Bidirectional cardiac shocking is permitted by these electrodes.

U.S. Patent No. 5,391,185 to Kroll discloses the use of multiple electrodes to effect atrial defibrillation. The possibility of inducing ventricular fibrillation during the course of atrial defibrillation is greatly reduced by synchronizing the atrial stimulation to fall within the QRS phase of the ventricular cycle.

U.S. Patent No. 5,181,511 to Nickolls, et al. discloses the use of multiple electrodes in antitachycardia pacing therapy. The electrodes not only each serve an

electrical sensing role (to locate the site of an ectopic focus), but also function in concert to create a virtual electrode for stimulating at the site of an ectopic focus.

Existing Needs. In the area of atrial malfunctions of rhythmicity what is needed is a means to entrain multiple atrial sites, but also in combination with an atrial sensing/ measurement capability that is coupled with atrial test stimulation and analysis capability. Atrial test stimulation and analysis capability is needed to provide better determination of the nature of the malfunction and the most probable or efficacious corrective therapy to undertake. Furthermore, the use of atrial test stimulation is critically needed for the fundamental reason that the physician cannot know *a priori* how a given heart (or a given heart under a particular medical or pathological condition) will respond to a selected stimulation regime, even if that selected stimulation regime would work generally for other cardiac patients. Thus, a trial-and-error testing capability needs to be available for pacemakers whose traditional stimulation regimes do not work for the occasional refractory patient. The multiple site stimulation capability is needed in order to more quickly and efficiently cardioconvert the atria in the face of arrhythmia, fibrillation, etc. Atrial sensing and use of measurement data are needed to better provide the physician and/or the circuit logic of the pacemaker with information as to the physiological state of the heart; i.e., whether there is atrial arrhythmia or fibrillation, where an ectopic focus is located, etc. Thus, what is needed is a pacemaker that combines all three of these elements: atrial sensing and measurement capability, atrial electrical test stimulation and analysis capability, and multiple site stimulation capability.

Lastly, a need also exists for a stimulation protocol which can travel more quickly across the myocardium and which provides improved cardiac entrainment along with the

ability to entrain portions of the heart from a greater distance.

Summary of the Invention

It therefore is an object of the present invention to provide a pacemaker that is capable of pacing atria from multiple sites.

It is another object of the present invention to provide a pacemaker that is capable of slowly entraining atria by stimulating the atria at multiple sites to produce electrical and functional conformity of the atria, with resulting increased pumping efficiency of the heart.

It is yet another object of the present invention to provide a pacemaker that is capable of detecting the presence of atrial fibrillation and atrial arrhythmias by stimulating the atria and observing and measuring the consequent effects on atrial and ventricular function.

It is a further object of the present invention to provide a pacemaker that is capable of obtaining and analyzing probability density function data from atria in order to determine atrial rates of beating and to assess atrial physiological function.

It is a further object of the present invention to provide an electronic stimulation device, for stimulating the atria from multiple sites, where the electrodes of the electronic stimulation device can be inserted intravenously.

It is a further object of the present invention to provide an electronic stimulation device, for stimulating the atria from multiple sites, where each electrode of the device has an independent generator.

It is a further object of the present invention to provide an electronic stimulation device for stimulating the atria from multiple sites, where each site is entrained separately and quickly brought to the same phase.

It is a further object of the present invention to provide an electronic stimulation device for stimulating the atria from multiple sites, to sequence the sites to mimic a normal heart beat.

It is a further object of the present invention to determine cardiac capture by monitoring cardiac activity and noting when the baseline of such activity is off zero.

It is a further object of the present invention to decrease threshold rises due to a build up of fibrous tissue.

The present invention accomplishes the above objectives by providing a cardiac pacemaker with a unique constellation of features and capabilities. In particular, a means for entraining multiple atrial sites is provided by the use of multiple electrodes. The multiple electrodes not only permit multi-site stimulation capability, but also multi-site sensing (including pdf measurement) capability, which, by triangulation, essentially provides the ability to determine the site(s) of any atrial ectopic focus. The multi-site stimulation capability inherently provides a system poised for more efficient entrainment and/or cardioconversion of the atria in the face of arrhythmia, fibrillation, etc. Combined with this multi-site stimulation/sensing capability is the means to execute trial-and-error testing and analysis to determine the best general stimulation protocol, to fine tune a given protocol, or to adjust a protocol in response to changes in the physiological/pathological status of the patient in general and/or the patient's heart in particular.

Incorporating the use of biphasic stimulation with the present invention provides the additional benefits of reducing cardiac inflammation damage, reducing or eliminating threshold rises due to the buildup of fibrous tissue and extending battery life of the electrodes.

In addition, the ability to conduct trial-and-error testing, including the analysis of the data derived therefrom, permits more thorough and more definitive determination of the physiological status of the heart; this determination can practically approach a moment-to-moment basis when analysis is automated by appropriate software for the purpose.

In sum, the present invention provides a cardiac pacemaker that has greater functional capabilities for the patient's atria than current technologies allow. The greater atrial "coverage" from the strategic placement of multiple electrodes permits faster correction of atrial arrhythmia, fibrillation, etc. Similarly, the use of multi-site electrodes permits more accurate sensing, including the capability of locating the site(s) of any atrial ectopic focus so as to better apply corrective stimulation procedures. In addition, the ability to apply trial-and-error testing/analytical procedures permits quicker analysis and correction of malfunctions of electrical conduction, cardiac contractility, rhythmicity, etc. Thus, the present invention constitutes an advance in cardiac care procedures as they relate to atrial pacemakers. The end result for the patient is better treatment, and, hence, a better prognosis from the better and faster treatment.

The method and apparatus relating to biphasic pacing comprises a first and second stimulation phase, with each stimulation phase having a polarity, amplitude, shape, and duration. In a preferred embodiment, the first and second phases have differing polarities. In one alternative embodiment, the two phases are of differing amplitude. In a second alternative embodiment, the two phases are of differing duration. In a third alternative embodiment, the first phase is in a chopped wave form. In a fourth alternative embodiment, the amplitude of the first phase is ramped. In a fifth alternative embodiment

the first phase is administered over 200 milliseconds after completion of a cardiac beating/pumping cycle. In a preferred alternative embodiment, the first phase of stimulation is an anodal pulse at maximum subthreshold amplitude for a long duration, and the second phase of stimulation is a cathodal pulse of short duration and high amplitude. It is noted that the aforementioned alternative embodiments can be combined in differing fashions. It is also noted that these alternative embodiments are intended to be presented by way of example only, and are not limiting.

Enhanced myocardial function is obtained through the biphasic stimulation of the present invention. The combination of cathodal with anodal pulses of either a stimulating or conditioning nature, preserves the improved conduction and contractility of anodal stimulation while eliminating the drawback of increased stimulation threshold. The result is a depolarization wave of increased propagation speed. This increased propagation speed results in increased synchronization and reduced heterogenicity of myocardial depolarization resulting in superior blood flow and contraction. Improved stimulation at a lower voltage level also results in: 1/ reduction in scar tissue buildup thereby reducing the tendency of the capture threshold to rise; 2/ reduction in power consumption leading to increased life for pacemaker batteries; and 3/ decreased potential for patient discomfort due to stimulation of the phrenic or diaphragmatic plexus or due to intercostal muscle pacing.

Brief Description of the Drawings

Figure 1 illustrates the location of leads and electrodes in relation to a human heart.

Figure 2 illustrates an alternative location of leads and electrodes in relation to a human heart.

Figure 2A illustrates a block diagram of the major functional components of the implanted pacemaker.

Figure 3 is a schematic representation of leading anodal biphasic stimulation.

Figure 4 is a schematic representation of leading cathodal biphasic stimulation.

Figure 5 is a schematic representation of leading anodal stimulation of low level and long duration, followed by conventional cathodal stimulation.

Figure 6 is a schematic representation of leading anodal stimulation of ramped low level and long duration, followed by conventional cathodal stimulation.

Figure 7 is a schematic representation of leading anodal stimulation of low level and short duration, administered in series followed by conventional cathodal stimulation.

Figure 8 illustrates the practice of the present invention.

Description of the Preferred Embodiments

Electrical stimulation is delivered via lead(s) or electrode(s). These leads can be epicardial (external surface of the heart) or endocardial (internal surface of the heart) or any combination of epicardial and endocardial. Leads are well known to those skilled in the art. Lead systems can be unipolar or bipolar. A unipolar lead has one electrode on the lead itself, the cathode. Current flows from the cathode, stimulates the heart, and returns to the anode on the casing of the pulse generator to complete the circuit. A bipolar lead has two poles on the lead a short distance from each other at the distal end, and both electrodes lie within the heart.

Figure 1 illustrates a plan view of implantable electronic stimulation device **102** and its associated lead and electrode system, in conjunction with human heart **104**. As illustrated, the device includes right atrial appendage lead **106**, right atrial septal lead **108**,

first coronary sinus lead **110** and second coronary sinus lead **112**. Each of these multiple small electrodes can be inserted intravenously and includes an independent generator.

Figure 2 illustrates a plan view of implantable electronic stimulation device **102** illustrating an alternative location of leads and electrodes in relation to human heart **104**. As illustrated, the device includes right atrial appendage lead **106**, right atrial septal lead **108**, first coronary sinus lead **110**, second coronary sinus lead **112** and left free wall lead **204**. Each of these multiple small electrodes can be inserted intravenously and includes an independent generator. Because of the use of independent generators, each electrode can be timed differently. In a preferred embodiment, left free wall lead **204** is placed by piercing septum **206** and passing left free wall lead **204** through the septum to the left side of the heart. The aforementioned placement of leads is for illustration purposes only, and is not intended as a limitation. It is contemplated that multiple leads placed in a variety of locations could be used.

Each site (area of lead placement) can be entrained separately, and then brought to the same phase. In a preferred embodiment each site is gradually brought to the same phase; however, certain situations could require that each site is quickly brought to the same phase. In an alternative embodiment, the sites can be sequenced to mimic a normal heart beat. In addition to allowing multi-site stimulation capability, the sensing circuits of each electrode also allow for multi-site sensing. Through triangulation the multi-site sensing provides a means for determining the site(s) of any atrial ectopic focus.

Referring to **Figure 2A**, a block diagram shows the major functional components of the implanted pacemaker **102**. Pacing/control circuitry **500**, in conjunction with microprocessor **501** detects the occurrence of tachycardia (and/or bradycardia) and in

response thereto controls the delivery of the various pacing therapies available via control bus **512**. The microprocessor **501** also detects the occurrence of atrial fibrillation. Detection of atrial fibrillation may be accomplished by the microprocessor **501** using any of the various detection methodologies known to the art. Generally, atrial fibrillation may be detected in response to an extended series of high rate atrial depolarizations. If greater specificity for atrial fibrillation is desired, analysis of regularity of rate waveform morphology may also be employed. Termination of atrial fibrillation may be detected in response to a decrease in the rate of atrial depolarizations and/or an increase in their regularity.

The operation of the microprocessor **501** is controlled by programming stored in a read only memory **505** and in a random access memory **503.** The operation of the device may be altered by the physician by altering the programming stored in the memory **503,** using control and telemetry circuitry conventional in implantable stimulators. Communication to and from the microprocessor **501,** the memories **503, 505,** and the control logic **500** is accomplished using an address/data bus **507.**

The atrial sensing circuit **509** can be any conventional cardiac sense amplifier circuits equivalent to any atrial cardiac sensing circuits employed in previous devices known in the art.

The implanted pacemaker **102** has a switch matrix **516** that allows selective delivery of pacing pulses from the atrial pacing driver **514** to the electrodes. The matrix **516** may be embodied as simply a collection of one or more FET and/or SCR switches activated under control of the pacing/control circuitry **500** to selectively pacing circuitry **516** to electrodes **106** and **108,** or to electrodes **110** and **112,** or other combinations of the electrodes. Thus, atrial anti-tachycardia (or anti-bradycardia pacing) is performed using any combination of the deployed pacing electrodes.

In a preferred embodiment, stimulation is administered at threshold until capture has occurred, at which time stimulation is administered at a subthreshold level. In alternative embodiments, stimulation is: (1) initiated at threshold and remains at threshold; (2) initiated subthreshold and remains subthreshold; (3) conventional prior to capture and then biphasic; (4) biphasic prior to capture and then conventional or (5) biphasic throughout.

Threshold refers to the minimum voltage level (or pulse width using a fixed voltage) which succeeds in stimulating (capturing) the myocardium. To capture is to produce a driven beat because of the stimulus given. Thus, in the absence of the pulse, the beat would not have been produced. Pulses which do not capture are subthreshold, (even though they may be shown to perturb the membrane potential somewhat, and transiently). Subthreshold pulses thus may affect subsequent conduction, but not by the mechanism of initiating a driven beat. Generally, to determine threshold, voltage (or pulse width) is varied (upward or downward) until capture is gained or lost.

Conventional stimulation is well known to those skilled in the art and comprises monophasic waveforms (cathodal or anodal) as well as multiphasic waveforms wherein the nonstimulating pulses are of a minimal magnitude and are used, for example, to dissipate a residual charge on an electrode.

Figures 3 through 7 depict a range of biphasic stimulation protocols. These protocols have been disclosed in United States Patent Application No. 08/699,552 to Mower, which is herein incorporated by reference in its entirety.

Figure 3 depicts biphasic electrical stimulation wherein a first stimulation phase comprising anodal stimulus **302** is administered having amplitude **304** and duration **306**.

This first stimulation phase is immediately followed by a second stimulation phase comprising cathodal stimulation **308** of equal intensity and duration.

Figure 4 depicts biphasic electrical stimulation wherein a first stimulation phase comprising cathodal stimulation **402** having amplitude **404** and duration **406** is administered. This first stimulation phase is immediately followed by a second stimulation phase comprising anodal stimulation **408** of equal intensity and duration.

Figure 5 depicts a preferred embodiment of biphasic stimulation wherein a first stimulation phase, comprising low level, long duration anodal stimulation **502** having amplitude **504** and duration **506**, is administered. This first stimulation phase is immediately followed by a second stimulation phase comprising cathodal stimulation **508** of conventional intensity and duration. In differing alternative embodiments, anodal stimulation **502** is: 1) at maximum subthreshold amplitude; 2) less than three volts; 3) of a duration of approximately two to eight milliseconds; and/or 4) administered over 200 milliseconds post heart beat. Maximum subthreshold amplitude is defined for purposes of this application as the maximum stimulation amplitude that can be administered without eliciting a contraction. In a preferred embodiment, anodal stimulation is approximately two volts for approximately three milliseconds duration. In differing alternative embodiments, cathodal stimulation **508** is: 1) of a short duration; 2) approximately 0.3 to 1.5 milliseconds; 3) of a high amplitude; 4) in the approximate range of three to twenty volts; and/or 5) of a duration less than 0.3 millisecond and at a voltage greater than twenty volts. In a preferred embodiment, cathodal stimulation is approximately six volts administered for approximately 0.4 millisecond. In the manner disclosed by these embodiments, as well as those alterations and modifications which can become obvious

upon the reading of this specification, a maximum membrane potential without activation is achieved in the first phase of stimulation.

Figure 6 depicts an alternative preferred embodiment of biphasic stimulation wherein a first stimulation phase, comprising anodal stimulation **602**, is administered over period **604** with rising intensity level **606**. The ramp of rising intensity level **606** can be linear or non-linear, and the slope can vary. This anodal stimulation is immediately followed by a second stimulation phase comprising cathodal stimulation **608** of conventional intensity and duration. In alternative embodiments, anodal stimulation **602**: (1) rises to a maximum subthreshold amplitude less than three volts; (2) is of a duration of approximately two to eight milliseconds; and/or (3) is administered over 200 milliseconds post heart beat. In yet other alternative embodiments, cathodal stimulation **608** is: (1) of a short duration; (2) approximately 0.3 to 1.5 milliseconds; (3) of a high amplitude; (4) in the approximate range of three to twenty volts; and/or (5) of a duration less than 0.3 milliseconds and at a voltage greater than twenty volts. In the manner disclosed by these embodiments, as well as those alterations and modifications which can become obvious upon the reading of this specification, a maximum membrane potential without activation is achieved in the first phase of stimulation.

Figure 7 depicts biphasic electrical stimulation wherein a first stimulation phase, comprising series **702** of anodal pulses, is administered at amplitude **704**. In one embodiment, rest period **706** is of equal duration to stimulation period **708**, and is administered at baseline amplitude. In an alternative embodiment, rest period **706** is of a differing duration than stimulation period **708**, and is administered at baseline amplitude. Rest period **706** occurs after each stimulation period **708**, with the exception that a second

stimulation phase, comprising cathodal stimulation **710** of conventional intensity and duration, immediately follows the completion of series **702**. In alternative embodiments: (1) the total charge transferred through series **702** of anodal stimulation is at the maximum subthreshold level; and/or (2) the first stimulation pulse of series **702** is administered over 200 milliseconds post heart beat. In yet other alternative embodiments, cathodal stimulation **710** is: (1) of a short duration; (2) approximately 0.3 to 1.5 milliseconds; (3) of a high amplitude; (4) in the approximate range of three to twenty volts, and/or (5) of a duration less than 0.3 milliseconds and at a voltage greater than twenty volts.

Figure 8 illustrates the practice of the present invention. Sensing is used to determine the existence of atrial fibrillation **802**. Sensing can be direct or indirect. For example, direct sensing can be based on data from multiple atrial sensing electrodes. The sensing electrodes sense the cardiac activity as depicted by electrical signals. For example, as is known in the art, R-waves occur upon the depolarization of ventricular tissue and P-waves occur upon the depolarization of atrial tissue. By monitoring these electrical signals the control/timing circuit of the ICD can determine the rate and regularity of the patient's heart beat, and thereby determine whether the heart is undergoing arrhythmia. This determination can be made by determining the rate of the sensed R-waves and/or P-waves and comparing this determined rate against various reference rates.

Direct sensing can be based upon varying criteria; such as, but not limited to, primary rate, sudden onset, and stability. The sole criteria of a primary rate sensor is the heart rate. When applying the primary rate criteria, if the heart rate should exceed a predefined level, then treatment is begun. Sensing electronics set to sudden onset criteria ignore those changes which occur slowly, and initiate treatment when there is a sudden

change such as immediate paroxysmal arrhythmia. This type of criteria would thus discriminate against sinus tachycardia. Stability of rate can also be an important criteria. For example, treatment with a ventricular device would not be warranted for a fast rate that varies, here treatment with an atrial device would be indicated.

In alternative embodiments, sensing can be indirect. Indirect sensing can be based on any of various functional parameters such as arterial blood pressure, rate of the electrocardiogram deflections or the probability density function (pdf) of the electrocardiogram. While it has been known in the art to apply pdf to the global electrocardiogram and/or to the R wave, it has been unexpectedly discovered that pdf of the baseline is also indicated for the determination of atrial abnormalities. Here, the electrodes are specific to the atrium and data related to the R wave is canceled out. Thus, whether or not to administer treatment can also be affected by pdf monitoring of the time the signal spends around the baseline.

Lastly, to determine whether an arrhythmia comes from the atria or the ventricles, a test impulse(s) can be given to one chamber to see if capture occurs and perturbs the rhythm. For example, in a ventricular rhythm, an atrial test impulse can capture the atrium, but the ventricular rhythm will continue unchanged afterwards. In an atrial rhythm, (or Sinus rhythm), if the atrial test pulse captures, the timing of all subsequent beats is changed. To determine if a pulse captures, the baseline immediately after the beat can be examined to determine if it is different from zero (or from a baseline template). If so, the beat can be inferred to have captured. In addition, the pdf pattern of the rhythm can be shown to have changed, inferring capture.

Thus, in a preferred embodiment, sensing electronics are based upon multiple

criteria. In addition, the present invention envisions devices working in more than one chamber such that appropriate treatment can be administered to either the atrium or the ventricle in response to sensing electronics based upon a variety of criteria, including those described above as well as other criteria known to those skilled in the art.

If atrial fibrillation occurs, a baseline of cardiac activity or a template can be recorded **804**. The template can be based on parameters such as electrocardiogram data, mechanical motion and/or probability density function data. In an alternative embodiment, the template is established after capture has occurred.

Pacing is initiated **806**. In a preferred embodiment, stimulation is administered at threshold until capture has occurred, at which time stimulation is administered at a subthreshold level. In alternative embodiments, stimulation is: (1) initiated at threshold and remains at threshold; (2) initiated subthreshold and remains subthreshold; (3) conventional prior to capture and then biphasic; (4) biphasic prior to capture and then conventional or (5) biphasic throughout.

The atrium is monitored throughout this initial pacing period to determine the status of capture **808**. Capture can be determined by multiple means. First, capture or the loss thereof, can be determined by monitoring cardiac rhythm. Loss of capture can result in a change in timing of the heart beat.

Second, capture or the loss thereof, can be determined through monitoring the previously described template. Where the template is established pre-stimulation, a change in the baseline signifies capture. Where the template is established after capture has occurred, a change in the template characteristics signifies loss of capture. The templates can be established and/or updated at any time.

Once capture occurs the stimulation protocol of the entrained sites is adjusted **810**. In a first embodiment, the stimulation rates of the entrained sites are slowed simultaneously, and then stopped. In a second embodiment, the spread of conduction is slowed. In a third embodiment, the stimulation speed is increased and stimulation is then stopped. In addition to adjusting stimulation rates upon the occurrence of capture, the stimulation protocol can also be adjusted such that (1) if stimulation of a conventional nature was administered prior to capture, biphasic stimulation is administered post-capture; (2) if biphasic stimulation was administered prior to capture, conventional stimulation is administered post-capture or (3) if biphasic stimulation was administered prior to capture, biphasic stimulation continues to be administered post-capture.

Having thus described the basic concept of the invention, it will be readily apparent to those skilled in the art that the foregoing detailed disclosure is intended to be presented by way of example only, and is not limiting. Various alterations, improvements and modifications will occur and are intended to those skilled in the art, but are not expressly stated herein. These modifications, alterations and improvements are intended to be suggested hereby, and within the scope of the invention. Further, the pacing pulses described in this specification are well within the capabilities of existing pacemaker electronics with appropriate programming. Accordingly, the invention is limited only by the following claims.

WHAT IS CLAIMED IS:

1 . A method of electrical cardiac pacing comprising:

sensing atrial fibrillation;

recording a baseline of cardiac activity;

stimulating the atrium using a pre-capture stimulation protocol;

determining status of capture; and

stimulating the atrium using a post-capture stimulation protocol, wherein the pre-capture stimulation protocol and the post-capture stimulation protocol comprise a procedure and wherein the procedure is selected from the group consisting of pre-capture stimulation at threshold with post-capture stimulation at threshold, pre-capture stimulation subthreshold with post-capture stimulation subthreshold and pre-capture stimulation at threshold with post-capture stimulation subthreshold.

2 . The method of electrical cardiac pacing of claim 1 , wherein the procedure is further selected from the group consisting of conventional stimulation pre-capture with biphasic stimulation post-capture, biphasic stimulation pre-capture with conventional stimulation post-capture, and biphasic stimulation pre-capture with biphasic stimulation post-capture.

3 . An apparatus for electrical cardiac pacing having:

a plurality of electrodes (106, 108) adapted to be disposed proximate atrial tissue;

an electrical stimulation driver (514), connected to at least one of the plurality of electrodes, to stimulate atrial tissue; and

processor circuitry (501) programmed to determine status of pacing capture;

wherein the apparatus is characterized by:

a sense amplifier (509) connected to at least one of the plurality of electrodes (106, 108) to sense atrial fibrillation;

a memory (503) in electrical communication with the sense amplifier (509), for recording a baseline of cardiac activity;

wherein, in the event that atrial fibrillation is sensed, the electrical stimulation driver (514) uses a pre-capture stimulation protocol,

wherein, in the event that capture status is determined, the electrical stimulation driver (514) uses a post-capture stimulation protocol, and

wherein the pre-capture stimulation protocol and the post-capture stimulation protocol comprise a procedure, and wherein the procedure is selected from the group consisting of: pre-capture stimulation at threshold with post-capture stimulation at threshold, pre-capture stimulation subthreshold with post-capture stimulation subthreshold, and pre-capture stimulation at threshold with post-capture stimulation subthreshold.

4 . The apparatus for electrical cardiac pacing of claim 3 , wherein the procedure uses biphasic stimulation post-capture.

5 . The apparatus for electrical cardiac pacing of claim 3 , wherein the procedure uses biphasic stimulation pre-capture.

6 . The apparatus for electrical cardiac pacing of claim 3 , wherein the procedure uses biphasic stimulation pre-capture with biphasic stimulation post-capture.

7 . The apparatus for electrical cardiac pacing of claim 4 , wherein biphasic stimulation comprises: applying a first stimulation phase and a second stimulation phase in sequence to cardiac tissue, wherein the first stimulation phase has a first phase polarity, a first phase amplitude, a first phase shape and a first phase duration, and wherein the second stimulation phase has a polarity opposite to the first phase polarity, a second phase amplitude, a second phase shape and a second phase duration.

8 . The apparatus for electrical cardiac pacing of claim 7 , wherein the first phase polarity is positive.

9 . The apparatus for electrical cardiac pacing of claim 5 , wherein biphasic stimulation comprises: applying a first stimulation phase and a second stimulation phase in sequence to cardiac tissue, wherein the first stimulation phase has a first phase polarity, a first phase amplitude, a first phase shape and a first phase duration, and wherein the second stimulation phase has a polarity opposite to the first phase polarity, a second phase amplitude, a second phase shape and a second phase duration.

10. The apparatus for electrical cardiac pacing of claim 9, wherein the first phase polarity is positive.

11. The apparatus for electrical cardiac pacing of claim 6., wherein wherein biphasic stimulation comprises: applying a first stimulation phase and a second stimulation phase in sequence to cardiac tissue, wherein the first stimulation phase has a first phase polarity, a first phase amplitude, a first phase shape and a first phase duration, and wherein the second stimulation phase has a polarity opposite to the first phase polarity, a second phase amplitude, a second phase shape and a second phase duration.

12. The apparatus for electrical cardiac pacing of claim 11, wherein the first phase polarity is positive.

13. The apparatus for electrical cardiac pacing of claim 11, wherein the first phase amplitude is less than the second phase amplitude.

14. The apparatus for electrical cardiac pacing of claim 11, wherein the first phase amplitude is ramped from a baseline value to a second value.

15. The apparatus for electrical cardiac pacing of claim 14, wherein the second value is equal to the second phase amplitude.

16. The apparatus for electrical cardiac pacing of claim 14, wherein the second value is at a maximum subthreshold amplitude.

17. The apparatus for electrical cardiac pacing of claim 16, wherein the maximum subthreshold amplitude is about 0.5 to 3.5 volts.

18. The apparatus for electrical cardiac pacing of claim 13, wherein the first phase duration is at least as long as the second phase duration.

19. The apparatus for electrical cardiac pacing of claim 13, wherein the first phase duration is about one to nine milliseconds.

20. The apparatus for electrical cardiac pacing of claim 13, wherein the second phase duration is about 0.2 to 0.9 milliseconds.

21. The apparatus for electrical cardiac pacing of claim 13, wherein the second phase amplitude is about two volts to twenty volts.

22. The apparatus for electrical cardiac pacing of claim 13, wherein the second phase duration is less than 0.3 milliseconds and the second phase amplitude is greater than 20 volts.

23. The apparatus for electrical cardiac pacing of claim 11, wherein the first stimulation phase further comprises a series of stimulating pulses of a predetermined amplitude, polarity, and duration.

24. The apparatus for electrical cardiac pacing of claim 23, wherein the first stimulation phase further comprises a series of rest periods.

25. The apparatus for electrical cardiac pacing of claim 24, wherein applying the first stimulation phase further comprises applying a rest period of a baseline amplitude after at least one stimulating pulse.

26. The apparatus for electrical cardiac pacing of claim 25, wherein the rest period is of equal duration to the stimulating pulse.

27. The apparatus for electrical cardiac pacing of claim 11, wherein the first phase amplitude is at a maximum subthreshold amplitude.

28. The apparatus for electrical cardiac pacing of claim 27, wherein the maximum subthreshold amplitude is about 0.5 to 3.5 volts.

29. The apparatus for electrical cardiac pacing of claim 11, wherein the first phase duration is at least as long as the second phase duration.

30. The apparatus for electrical cardiac pacing of claim 11, wherein the first phase duration is about one to nine milliseconds.

31. The apparatus for electrical cardiac pacing of claim 11, wherein the second phase duration is about 0.2 to 0.9 milliseconds.

32 . The apparatus for electrical cardiac pacing of claim 11, wherein the second phase amplitude is about two to twenty volts.

33 . The apparatus for electrical cardiac pacing of claim 11, wherein the second phase duration is less than 0.3 milliseconds and the second phase amplitude is greater than 20 volts.

34 . The apparatus for electrical cardiac pacing of claim 11, wherein the first stimulation phase is initiated greater than 200 milliseconds after heart beat.

35 . The apparatus for electrical cardiac pacing of claim 3, wherein sensing atrial fibrillation comprises:

monitoring parameters selected from the group consisting of: arterial blood pressure, rate of electrocardiogram deflections, and probability density function of the electrocardiogram.

36. An implantable cardiac stimulator to perform atrial defibrillation, the cardiac stimulator comprising:

sensor adapted to sense the onset of atrial fibrillation;

recorder adapted to record a baseline of cardiac activity;

processor adapted to determining whether capture has occurred; and

electrodes adapted to stimulate the atrium after the onset of atrial fibrillation has been sensed,

wherein, in the event it is determined that capture has not occurred, the atrium is stimulated using a pre-capture stimulation protocol,

wherein, in the event it is determined that capture has occurred, the atrium is stimulated using a post-capture stimulation protocol, and

wherein the pre-capture stimulation protocol and the post-capture stimulation protocol comprise a procedure, the procedure being selected from the group consisting of: pre-capture stimulation at threshold with post-capture stimulation at threshold, pre-capture stimulation subthreshold with post-capture stimulation subthreshold, and pre-capture stimulation at threshold with post-capture stimulation subthreshold.

37. The implantable cardiac stimulator of claim36, wherein the procedure utilizes biphasic stimulation pre-capture with biphasic stimulation post-capture.

38. The implantable cardiac stimulator of claim37, where at least two of the electrodes are adapted to be inserted intravenously into a patient, and at least two of the electrodes are adapted to be placed in conjunction with cardiac tissue.

39. The implantable cardiac stimulator of claim38, wherein at least one of the electrodes is adapted to be located in the right atrial appendage, at least one of the electrodes is adapted to be located in the right atrial septum, and at least one of the electrodes is adapted to be located in the coronary sinus.

40. The implantable cardiac stimulator of claim39, wherein at least one of the electrodes is adapted to be located in the left free wall.

41 . The implantable cardiac stimulator of claim 38 wherein each of the electrodes has an independent generator.

42 . The implantable cardiac stimulator of claim 38 wherein the stimulation of the electrodes is sequenced so as to mimic a normal heart beat.

43 . The implantable cardiac stimulator of claim 37 wherein the processor determines status of capture by monitoring for changes in the baseline of cardiac activity established by the recorder.

44 . The implantable cardiac stimulator of claim 37 wherein the baseline of cardiac activity comprises a template of parameters selected from the group consisting of: electrocardiogram data, mechanical motion, and probability density function data.

45 . The implantable cardiac stimulator of claim 37 wherein biphasic stimulation comprises:

defining a first stimulation phase with a first phase polarity, a first phase amplitude, a first phase shape and a first phase duration;

defining a second stimulation phase with a polarity opposite to the first phase polarity, a second phase amplitude, a second phase shape and a second phase duration; and

applying the first stimulation phase and the second stimulation phase in sequence to cardiac tissue.

46 . The implantable cardiac stimulator of claim 45 wherein the first phase polarity is positive.

47 . The implantable cardiac stimulator of claim 45 wherein the first phase amplitude is less than the second phase amplitude.

48 . The implantable cardiac stimulator of claim 45 wherein the first phase amplitude is ramped from a baseline value to a second value.

49. The implantable cardiac stimulator of claim 48 wherein the second value is equal to the second phase amplitude.

50. The implantable cardiac stimulator of claim 48 wherein the second value is at a maximum subthreshold amplitude.

51. The implantable cardiac stimulator of claim 50 wherein the maximum subthreshold amplitude is about 0.5 to 3.5 volts.

52. The implantable cardiac stimulator of claim 48 wherein the first phase duration is at least as long as the second phase duration.

53. The implantable cardiac stimulator of claim 48 wherein the first phase duration is about one to nine milliseconds.

54. The implantable cardiac stimulator of claim 48 wherein the second phase duration is about 0.2 to 0.9 milliseconds.

55. The implantable cardiac stimulator of claim 48 wherein the second phase amplitude is about two volts to twenty volts.

56. The implantable cardiac stimulator of claim 48 wherein the second phase duration is less than 0.3 milliseconds and the second phase amplitude is greater than 20 volts.

57. The implantable cardiac stimulator of claim 45 wherein the first stimulation phase further comprises a series of stimulating pulses of a predetermined amplitude, polarity and duration.

58. The implantable cardiac stimulator of claim 57, wherein the first stimulation phase further comprises a series of rest periods.

59. The implantable cardiac stimulator of claim 58 wherein applying the first stimulation phase further comprises applying a rest period of a baseline amplitude after at least one stimulating pulse.

60. The implantable cardiac stimulator of claim 59 wherein the rest period is of equal duration to the stimulating pulse.

61. The implantable cardiac stimulator of claim 45 wherein the first phase amplitude is at a maximum subthreshold amplitude.

62. The implantable cardiac stimulator of claim 61 wherein the maximum subthreshold amplitude is about 0.5 to 3.5 volts.

63. The implantable cardiac stimulator of claim 45 wherein the first phase duration is at least as long as the second phase duration.

64. The implantable cardiac stimulator of claim 45, wherein the first phase duration is about one to nine milliseconds.

65. The implantable cardiac stimulator of claim 45, wherein the second phase duration is about 0.2 to 0.9 milliseconds.

66. The implantable cardiac stimulator of claim 45, wherein the second phase amplitude is about two volts to twenty volts.

67. The implantable cardiac stimulator of claim 45, wherein the second phase duration is less than 0.3 milliseconds and the second phase amplitude is greater than 20 volts.

68. The implantable cardiac stimulator of claim 45, wherein the first stimulation phase is initiated greater than 200 milliseconds after heart beat.

69. The implantable cardiac stimulator of claim 45, wherein the first stimulation phase comprises anodal stimulation.

70. The implantable cardiac stimulator of claim 37, wherein the sensor senses atrial fibrillation by monitoring parameters selected from the group consisting of: arterial blood pressure, rate of electrocardiogram deflections, and probability density function of the electrocardiogram.

71. The implantable cardiac stimulator of claim 37, further comprising:

sensing circuits, each of the sensing circuits being connected to a respective one of the electrodes and being adapted to provide sensing data concerning the site of one or more atrial ectopic foci.

72. The implantable cardiac stimulator of claim 71, wherein the processor is connected to receive the sensing data from the sensing circuits and determines the site of one or more atrial ectopic foci by triangulating the sensing data.

73. An apparatus for electrical cardiac pacing comprising:

a plurality of electrodes adapted to be disposed proximate atrial tissue;

a sense amplifier connected to at least one of the plurality of electrodes to sense atrial fibrillation;

a memory in electrical communication with the sense amplifier, for recording a baseline of cardiac activity;

an electrical stimulation driver, connected to at least one of the plurality of electrodes, to stimulate atrial tissue; and

processor circuitry programmed to determine status of pacing capture;

wherein, in the event that atrial fibrillation is sensed, the electrical stimulation driver uses a pre-capture stimulation protocol,

wherein, in the event that capture status is determined, the electrical stimulation driver uses a post-capture stimulation protocol, and

wherein the pre-capture stimulation protocol and the post-capture stimulation protocol comprise a procedure, and wherein the procedure is selected from the group consisting of: pre-capture stimulation at threshold with post-capture stimulation at threshold, ~~pre capture~~

stimulation subthreshold with post-capture stimulation subthreshold, and pre-capture stimulation at threshold with post-capture stimulation subthreshold.

74. The apparatus for electrical cardiac pacing of claim 73 wherein the procedure uses biphasic stimulation post-capture.

75. The apparatus for electrical cardiac pacing of claim 74 wherein the biphasic stimulation has a first phase that comprises anodal stimulation.

76. The apparatus for electrical cardiac pacing of claim 73 wherein the procedure uses biphasic stimulation pre-capture.

77. The apparatus for electrical cardiac pacing of claim 76 wherein the biphasic stimulation has a first phase that comprises anodal stimulation.

78. The apparatus for electrical cardiac pacing of claim 73 wherein the procedure uses biphasic stimulation pre-capture with biphasic stimulation post-capture.

79. The apparatus for electrical cardiac pacing of claim 78 wherein the biphasic stimulation has a first phase that comprises anodal stimulation.