

WHAT IS CLAIMED IS:

1. A PAP kit comprising an interface with a first seal having a first size and a second seal having a second size that is different than the first size, wherein the second size is between 1.9% to 22.9% larger in width than the first size when measured horizontally between edges of an opening of the seal, and 5.4% to 18.5% larger in height than the first size when measured along a center line of the seal between a midpoint in a chin region and a midpoint in a nasal bridge region, wherein the width of the second size is between 58.1% to 67.0% of the height of the second size.

2. The kit of Claim 1, further comprising a frame, wherein the first seal and the second seal are configured to be removably and interchangeably connectable to the frame; and

further comprising at least one headgear, wherein the at least one headgear is configured to be removably and interchangeably connectable to the first seal and the second seal, or the at least one headgear comprises a first headgear configured to be removably connectable to the first seal and a second headgear configured to be removably connectable to the second seal.

3. The kit of Claim 1 or 2, wherein the first size has a height of about 81mm when measured along a center line of the seal between a midpoint in a chin region and a midpoint in a nasal bridge region and a width of about 50.5mm when measured horizontally between edges of an opening of the seal,

wherein the height and width of the first and second size may be within a range of ± 2.5 mm of the disclosed dimensions.

4. The kit of any one of Claims 1-3, wherein the second size has a height of about 90.5mm when measured along a center line of the seal between a midpoint in a chin region and a midpoint in a nasal bridge region and a width of about 56.5mm when measured horizontally between edges of an opening of the seal,

wherein the height and width of the first and second size may be within a range of ± 2.5 mm of the disclosed dimensions.

5. The kit of any one of Claims 1-4, wherein a width of an opening of the second size is about 11.9% greater than a width of an opening of the first size.

6. The kit of any one of Claims 1-5, wherein a width of an opening of the second

size is approximately 1mm to 11mm greater than a width of an opening of the first size, when measured horizontally between edges of the openings.

7. The kit of Claim 6, wherein a width of an opening of the second size is approximately 6mm greater than a width of an opening of the first size.

8. The kit of any one of Claims 1-7, wherein a height of the second size is approximately 4.5mm to 14.5mm greater than a height of the first size, when measured along a center line of the seal between a midpoint in a chin region and a midpoint in a nasal bridge region.

9. The kit of any one of Claims 1-8, wherein a difference in height is approximately twice a difference in width between the first size and the second size, the height being measured along a center line of the seal between a midpoint in a chin region and a midpoint in a nasal bridge region and the width being measured horizontally between edges of an opening of the seal.

10. The kit of any one of Claims 1-9, wherein a difference in height between the first size and the second size is equal to or less than 9.5mm, the height being measured along a center line of the seal between a midpoint in a chin region and a midpoint in a nasal bridge region.

11. The kit of any one of Claims 1-10, wherein a difference in height between the first size and the second size is equal to or less than 11.7%, the height being measured along a center line of the seal between a midpoint in a chin region and a midpoint in a nasal bridge region.

12. The kit of any one of Claims 1-11, wherein a width of the first size is between 57.5% to 67.5% of a height of the first size, the height being measured along a center line of the seal between a midpoint in a chin region and a midpoint in a nasal bridge region and the width being measured horizontally between edges of an opening of the seal.

13. The kit of Claim 12, wherein the width of the first size is 62.3% of the height of the first size.

14. The kit of Claim 13, wherein the width of the second size is 62.4% of the height of the second size.

15. The kit of any one of Claims 1-14, wherein the first size and the second size have equal width versus height proportions.

16. The kit of any one of Claims 1-15, wherein an upper portion of the first and second seals each include a seal having a forwardly-deflectable upper portion, wherein the heights are measured in a relaxed position of the forwardly-deflectable upper portions.

17. The kit of Claim 16, wherein the forwardly-deflectable upper portion rolls in a forward direction relative to a lower portion of the seal.