



Abstract:

With regard to leak test devices with multiple testing stations (1) it is suggested that the measuring signals that are relevant to the tightness be transmitted from the chambers via multiplexers (5) to a central evaluation (9).

(Fig. 1)