

Use of signal enhancing compounds in electrochemiluminescence detection

Field of the invention

The invention concerns methods for the detection of an analyte in a sample by electrochemiluminescence using new reagent compositions. New reagent compositions, reagent kits for measuring electrochemiluminescence (ECL) and electrochemiluminescence detection methods using the new reagent compositions are disclosed. In particular, the invention relates to the use of novel combinations of compounds which can be used in said measurements to provide improved assay performance.

Background and prior art

Methods for measuring electrochemiluminescent phenomena have been known for some years. Such methods make use of the ability of special metal complexes to achieve, by means of oxidation, an excited state from which they decay to ground state, emitting electrochemiluminescence. For review see Richter, M.M., Chem. Rev. 104 (2004) 3003-3036.

At this time, there are a number of commercially available instruments that utilize electrochemiluminescence (ECL) for analytical measurements. Species that can be induced to emit ECL (ECL-active species) have been used as ECL labels. Examples of ECL labels include organometallic compounds such as the tris-bipyridyl-ruthenium (RuBpy) moiety where the metal is from, for example, the metals of group VII and VIII, including Re, Ru, Ir and Os. Species that react with the ECL label in the ECL process are referred to herein as ECL coreactants. Commonly used coreactants for ECL include tertiary amines (e.g. tripropylamine (TPA)), oxalate, and persulfate. The light is generated by ECL labels or ligands; the participation of the binding reagent in a binding interaction can be monitored by measuring ECL emitted from the ECL label. Alternatively, the ECL signal from an ECL-active compound may be indicative of the chemical environment (see, e.g., U.S. Pat. No. 5,641,623 and 5,643,713, which describes ECL assays that monitor the presence or destruction of special ECL coreactants). For more background on ECL, ECL labels, ECL assays and instrumentation for conducting ECL assays see U.S. Pat. Nos. 5,093,268; 5,147,806; 5,240,863; 5,308,754; 5,324,457; 5,591,581; 5,597,910; 5,679,519; 5,705,402; 5,731,147; 5,786,141; 5,846,485; 5,866,434; 6,066,448; 6,136,268 and 6,207,369, and EP 0 441 875, and published PCT

Nos. WO97/36931; WO98/12539; WO99/14599; WO99/32662; WO99/58962; WO99/63347; WO00/03233 and WO98/57154.

5 Commercially available ECL instruments have demonstrated exceptional performance. They have become widely used for reasons including their excellent sensitivity, dynamic range, precision, and tolerance of complex sample matrices. The commercially available instrumentation uses flow cell-based designs with permanent reusable flow cells.

10 Available sample volumes for the determination of analytes are often limited and more and more different analytes have to be determined out of one sample. Also the development of faster laboratory equipment for assay automation and more sensitive methods for the detection of analytes are required. This leads to the need for high sensitive and specific electrochemiluminescent assays and methods for performing them. In addition improvements associated with safety hazards or environmental concerns should be considered.


15 However, even more sensitive detection of analytes would be of great advantage. Thus, the object of the present invention was to improve said known methods and reagent compositions especially with respect to enhancement of the ECL signal and an improved analyte detection in combination with electrochemiluminescent procedures. It would be desirable to find novel signal enhancing reagents and/or
20 compounds with improved performance in electrochemiluminescent assays.

Summary

The present disclosure in one aspect concerns a method for measuring an analyte in a sample via electrochemiluminescent detection, comprising the steps of a) incubating the sample with a detection reagent labeled with an
25 electrochemiluminescent group, b) separating analyte-bound and analyte-free labeled detection reagent, c) incubating the separated labeled detection reagent with a reagent composition comprising i) at least one coreactant, and ii) at least one compound selected from the group of carbonic acid amides of Formula I and Formula II,

30

Formula I




with R₁ = CH₃, CH₂F, CH₂Cl, CH₂CH₃, CHClCH₃, CH₂CH₂Cl, C(CH₃)₂CH₃, CH₂CH₂CH₃, CClHCH₂CH₃ or CH₂CH₂CH₂CH₃,

with R₂ = H, and

with R₃ = H,

Formula II,



d) electrochemically triggering the release of luminescence, and e) determining the electrochemiluminescence (ECL) signal thereby measuring the analyte.

The present disclosure also concerns a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II and ii) at least one coreactant.

In another aspect, there is provided a reagent composition for determining ECL, comprising

i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, and

ii) at least one coreactant,

wherein the coreactant is selected from the group consisting of tertiary amines (e.g. tripropylamine (TPA)), oxalate, and persulfate.

The present disclosure further relates to a reagent mixture, comprising a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II and ii) at least one coreactant, a sample to be investigated, and at least one detecting reagent labeled with an electrochemiluminescent group.

In another aspect, there is provided a reagent mixture for determining ECL, comprising a reagent composition as defined above, a sample to be investigated and at least one detection reagent labeled with an electrochemiluminescent group.

5 In another aspect, there is provided the use of a reagent composition as defined above in the determination of ECL.

In another aspect, there is provided the use of a carbonic acid amide selected from the group of Formula I and Formula II for performing an electrochemiluminescent detection procedure.

10 The present disclosure also concerns a kit for measuring ECL, which contains a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II and ii) at least one coreactant.

In another aspect, there is provided a kit for measuring ECL, which contains a reagent composition as defined above.

15 The disclosure, as well as additional objects, features and advantages thereof, will be understood more fully from the following detailed description of certain preferred embodiments.

Detailed Description

The practicing of the present invention will employ, unless otherwise indicated, conventional techniques of molecular biology (including recombinant techniques), microbiology, cell biology, biochemistry, and immunology, which are within the skill of the art. Such techniques are explained fully in the literature, such as, "Molecular Cloning: A Laboratory Manual", second edition (Sambrook et al., 1989); "Oligonucleotide Synthesis" (M. J. Gait, ed., 1984); "Animal Cell Culture" (R. I. Freshney, ed., 1987); "Methods in Enzymology" (Academic Press, Inc.); "Current Protocols in Molecular Biology" (F. M. Ausubel et al., eds., 1987, and periodic updates); "PCR: The Polymerase Chain Reaction", (Mullis et al., eds., 1994).

Unless defined otherwise, technical and scientific terms used herein have the same meaning as commonly understood by one of ordinary skill in the art to which this invention belongs. Singleton et al., Dictionary of Microbiology and Molecular Biology, 2nd ed., J. Wiley & Sons, New York (1994); March, Advanced Organic Chemistry Reactions, Mechanisms and Structure, 4th ed., John Wiley & Sons, New York (1992); Lewin, B., Genes V, published by Oxford University Press (1994), ISBN 0-19-854287 9); Kendrew, J. et al. (eds.), The Encyclopedia of Molecular Biology, published by Blackwell Science Ltd. (1994), ISBN 0-632-02182-9); and Meyers, R.A. (ed.), Molecular Biology and Biotechnology: a Comprehensive Desk Reference, published by VCH Publishers, Inc. (1995), ISBN 1-56081-569 8) provide one skilled in the art with a general guidance to many of the terms used in the present application.

All references cited herein, including patent applications and publications, are incorporated by reference in their entirety.

Definitions:

As used herein, each of the following terms has the meaning associated with it in this section.


The articles "a" and "an" are used herein to refer to one or to more than one (i.e. to at least one) of the grammatical object of the article. By way of example, "an analyte" means one analyte or more than one analyte. The term "at least" is used to indicate that optionally one or more further objects may be present. By way of

example, an array comprising at least two discrete areas may optionally comprise two or more discrete test areas.



The expression “one or more” denotes 1 to 50, preferably 1 to 20 also preferred 2, 3, 4, 5, 6, 7, 8, 9, 10, 12, or 15.

- 5 Examples for “carbonic acid amides” and their chemical structures are listed in Table 1.


Table 1: The carbonic acid amides have the following common structure, unless otherwise stated:

No.	CA Index name	CAS-No.	structure	Residues
1	formamide	75-122-7		R ₁ = H R ₂ = H R ₃ = H
2	acetamide	60-35-5	see above	R ₁ = CH ₃ R ₂ = H R ₃ = H
3	2-fluoroacetamide	640-19-7	see above	R ₁ = CH ₂ F R ₂ = H R ₃ = H
4	2,2,2-trifluoroacetamide	354-38-1	see above	R ₁ = CF ₃ R ₂ = H R ₃ = H
5	2-chloroacetamide	79-07-2	see above	R ₁ = CH ₂ Cl R ₂ = H R ₃ = H
6	2,2-dichloroacetamide	68372-7	see above	R ₁ = CHCl ₂ R ₂ = H R ₃ = H
7	2-bromoacetamide	683-57-8	see above	R ₁ = CH ₂ Br R ₂ = H R ₃ = H
8	2-jodoacetamide	144-48-9	see above	R ₁ = CH ₂ J R ₂ = H R ₃ = H
9	2-hydroxyacetamide	598-42-5	see above	R ₁ = CH ₂ OH R ₂ = H R ₃ = H
10	acetoacetamide	5977-14-0	see above	R ₁ = CH ₂ COCH ₃ R ₂ = H R ₃ = H
11	2-chloro-N,N-dimethylacetamide	2675-89-0	see above	R ₁ = CH ₂ Cl R ₂ = CH ₃ R ₃ = CH ₃
12	2-chloro-N-hydroxy-methylacetamide	2832-19-1	see above	R ₁ = CH ₂ Cl R ₂ = CH ₂ OH R ₃ = H

- 6 -

No.	CA Index name	CAS-No.	structure	Residues
13	2-chloro-N-methoxy-N-methylacetamide	67442-07-3	see above	$R_1 = \text{CH}_2\text{Cl}$ $R_2 = \text{CH}_3$ $R_3 = \text{OCH}_3$
14	propanamide (propionamide)	79-05-0	see above	$R_1 = \text{CH}_2\text{CH}_3$ $R_2 = \text{H}$ $R_3 = \text{H}$
15	2-chloropropanamide	27816-36-0	see above	$R_1 = \text{CHClCH}_3$ $R_2 = \text{H}$ $R_3 = \text{H}$
16	3-chloropropanamide	5875-24-1	see above	$R_1 = \text{CH}_2\text{CH}_2\text{Cl}$ $R_2 = \text{H}$ $R_3 = \text{H}$
17	N-methylpropanamide	1187-58-2	see above	$R_1 = \text{CH}_2\text{CH}_3$ $R_2 = \text{CH}_3$ $R_3 = \text{H}$
18	2,2-dimethyl-propanamide	754-10-9	see above	$R_1 = \text{C}(\text{CH}_3)_2\text{CH}_3$ $R_2 = \text{H}$ $R_3 = \text{H}$
19	propanediamide	108-13-4	see above	$R_1 = \text{CH}_2\text{CONH}_2$ $R_2 = \text{H}$ $R_3 = \text{H}$
20	Butanamide	541-35-5	see above	$R_1 = \text{CH}_2\text{CH}_2\text{CH}_3$ $R_2 = \text{H}$ $R_3 = \text{H}$
21	2-chlorobutanamide	2455-04-1	see above	$R_1 = \text{CClHCH}_2\text{CH}_3$ $R_2 = \text{H}$ $R_3 = \text{H}$
22	Pentanamide	626-97-1	see above	$R_1 = \text{CH}_2\text{CH}_2\text{CH}_2\text{CH}_3$ $R_2 = \text{H}$ $R_3 = \text{H}$
23	Hexanamide	628-02-4	see above	$R_1 = \text{CH}_2\text{CH}_2\text{CH}_2\text{CH}_2\text{CH}_3$ $R_2 = \text{H}$ $R_3 = \text{H}$
24	2-pyrrolidinone	616-45-5		
25	2,5-butanamide	123-56-8		


Formula I as used herein denotes



- 7 -

with $R_1 = \text{CH}_3, \text{CH}_2\text{F}, \text{CH}_2\text{Cl}, \text{CH}_2\text{CH}_3, \text{CHClCH}_3, \text{CH}_2\text{CH}_2\text{Cl}, \text{C}(\text{CH}_3)_2\text{CH}_3, \text{CH}_2\text{CH}_2\text{CH}_3, \text{CClHCH}_2\text{CH}_3$ or $\text{CH}_2\text{CH}_2\text{CH}_2\text{CH}_3$, with $R_2 = \text{H}$, and with $R_3 = \text{H}$.

5 Formula II as used herein denotes to the following structure named 2-pyrrolidinone, represented in Table 1 as No. 24.



The embodiments of the invention can be used to test a variety of samples which may contain an analyte or activity of interest. Such samples may be in solid, emulsion, suspension, liquid, or gas form. They may be, but are not limited to, samples containing or derived from human or animals, for example, cells (live or dead) and cell-derived products, immortalized cells, cell fragments, cell fractions, cell lysates, organelles, cell membranes, hybridoma, cell culture supernatants (including supernatants from antibody producing organisms such as hybridomas), waste or drinking water, food, beverages, pharmaceutical compositions, blood, serum, plasma, hair, sweat, urine, feces, stool, saliva, tissue, biopsies, effluent, separated and/or fractionated samples, separated and/or fractionated liquids, organs, plants, plant parts, plant byproducts, soil, water, water supply, water sources, filtered residue from fluids (gas and liquid), swipes, absorbent materials, gels, cytoskeleton, unfractionated samples, unfractionated cell lysates, cell nucleus/nuclei, nuclear fractions, chemicals, chemical solutions, structural biological components, skeletal (ligaments, tendons) components, separated and/or fractionated skeletal components, hair fractions and/or separations, skin, skin samples, skin fractions, dermis, endodermis, eukaryotic cells, prokaryotic cells, fungus, yeast, immunological cells, drugs, therapeutic drugs, oils, extracts, mucous, sewage, environmental samples, organic solvents or air. In an embodiment the sample can further comprise, for example, water, alcohols, acetonitrile, dimethyl sulfoxide, dimethyl formamide, n-methyl-pyrrolidone, methanol or other organic solvents.

A "sample" as used herein is obtained for the purpose of an evaluation in vitro. As the skilled artisan will appreciate, any such assessment is made in vitro. If the sample is a patient sample, it is discarded afterwards. The patient sample is solely

used for the in vitro diagnostic method of the invention and the material of the patient sample is not transferred back into the patient's body.

Analytes that may be measured include, but are not limited to, whole cells, cell surface antigens, protein complexes, cell signaling factors and/or components, second messengers, second messenger signaling factors and/or components, subcellular particles (e.g., organelles or membrane fragments), viruses, prions, dust mites or fragments thereof, viroids, immunological factors, antibodies, antibody fragments, antigens, haptens, fatty acids, nucleic acids (and synthetic analogs), ribosomes, proteins (and synthetic analogs), lipoproteins, polysaccharides, inhibitors, cofactors, haptens, cell receptors, receptor ligands, lipopolysaccharides, glycoproteins, peptides, polypeptides, enzymes, enzyme substrates, enzyme products, nucleic acid processing enzymes (e.g., polymerases, nucleases, integrases, ligases, helicases, telomerases, etc.), protein processing enzymes (e.g., proteases, kinases, protein phosphatases, ubiquitin-protein ligases, etc.), cellular metabolites, endocrine factors, paracrine factors, autocrine factors, cytokines, hormones, pharmacological agents, drugs, therapeutic drugs, synthetic organic molecules, organometallic molecules, tranquilizers, barbiturates, alkaloids, steroids, vitamins, amino acids, sugars, lectins, recombinant or derived proteins, biotin, avidin, streptavidin, or inorganic molecules present in the sample.

Whole cells may be animal, plant, or bacteria, and may be viable or dead. Examples include plant pathogens such as fungi and nematodes. The term "subcellular particles" is meant to encompass, for example, subcellular organelles, membrane particles as from disrupted cells, fragments of cell walls, ribosomes, multi-enzyme complexes, and other particles which can be derived from living organisms. Nucleic acids include, for example, chromosomal DNA, plasmid DNA, viral DNA, and recombinant DNA derived from multiple sources. Nucleic acids also include RNA's, for example messenger RNA's, ribosomal RNA's and transfer RNA's. Polypeptides include, for example, enzymes, transport proteins, receptor proteins, and structural proteins such as viral coat proteins. Preferred polypeptides are enzymes and antibodies. Particularly preferred polypeptides are monoclonal antibodies. Hormones include, for example, insulin and T4 thyroid hormone. Pharmacological agents include, for example, cardiac glycosides. It is of course within the scope of this invention to include synthetic substances which chemically resemble biological materials, such as synthetic polypeptides, synthetic nucleic acids, and synthetic membranes, vesicles and liposomes. The foregoing is not

intended to be a comprehensive list of the biological substances suitable for use in this invention, but is meant only to illustrate the wide scope of the invention.

Also, typically, the analyte of interest is present at a concentration of 10^{-3} molar or less, for example, at least as low as 10^{-18} molar.

5 The term "analyte specific reagent" (ASR) according to the present invention has to be understood as a molecule or biomolecule (e.g. a protein or antibody) with the capability to specifically bind the analyte. "Analyte specific reagents" (ASRs) are a class of biological molecules which can be used to identify and measure the amount of an individual chemical substance in biological specimens. ASRs are for
10 example antibodies, both polyclonal and monoclonal, specific receptor proteins, ligands, nucleic acid sequences, and similar reagents which, through specific binding or chemical reaction with substances in a specimen, are intended to use in a diagnostic application for identification and quantification of an individual chemical substance or ligand in biological specimens. In simple terms an analyte
15 specific reagent is the active ingredient of an assay. An ASR will fulfill both, the criteria for affinity as well as for specificity of binding the analyte.

The term "antibody" is used in the broadest sense and specifically covers monoclonal antibodies (including full length monoclonal antibodies), polyclonal antibodies, multispecific antibodies (e.g., bispecific antibodies), and antibody
20 fragments. The term "antibody" encompasses the various forms of antibody structures including whole antibodies and antibody fragments. The antibody according to the invention is in one embodiment a human antibody, a humanized antibody, a chimeric antibody, an antibody derived from other animal species like mouse, goat or sheep, a monoclonal or polyclonal antibody, or a T cell antigen
25 depleted antibody. Genetic engineering of antibodies is e.g. described in Morrison, S.L., et al., Proc. Natl. Acad. Sci. USA 81 (1984) 6851-6855; US 5,202,238 and US 5,204,244; Riechmann, L., et al., Nature 332 (1988) 323-327; Neuberger, M.S., et al., Nature 314 (1985) 268-270; Lonberg, N., Nat. Biotechnol. 23 (2005) 1117-1125.

30 Any antibody fragment retaining the above criteria of a analyte specific reagent can be used. Antibodies are generated by state of the art procedures, e.g., as described in Tijssen (Tijssen, P., Practice and theory of enzyme immunoassays, 11, Elsevier Science Publishers B.V., Amsterdam, the whole book, especially pages 43-78). In addition, the skilled artisan is well aware of methods based on immunosorbents that

can be used for the specific isolation of antibodies. By these means the quality of antibodies and hence their performance in immunoassays can be enhanced (Tijssen, P., *supra*, pages 108-115).

5 A "detection reagent" according to the present invention comprises an analyte specific reagent (ASR) labeled with an electrochemiluminescent group, or an analyte homolog labeled with an electrochemiluminescent group. According to the test format it is known to the skilled artisan, which detection reagent has to be selected for the various assay formats (e.g. sandwich assay, double antigen bridging assay (DAGS), competitive assay, homogeneous assay, heterogeneous
10 assay). A detection reagent in a heterogeneous immunoassay might be for example an antibody. It is known to a person skilled in the art that the detection reagent can be immobilized on a solid phase. In an embodiment the method for measuring an analyte in a sample via electrochemiluminescent detection can be performed as a homogeneous assay. In an embodiment the method can be performed as a heterogeneous assay. In an embodiment the method can be performed in a sandwich assay format. In an embodiment the method can be performed in a competitive assay format. Also in an embodiment the method can be performed in a double antigen bridging assay format (DAGS). Known immunoassay formats are described in detail in the book of Price, C.P. and Newman, D.J., *Principles and Practice of Immunoassay*, 2nd ed. (1997).
20

The term "label" as used herein refers to any substance that is capable of producing a detectable signal, whether visibly or by using suitable instrumentation. Various labels suitable for use in the present invention include, but are not limited to, chromogens, fluorescent, chemiluminescent or electrochemiluminescent
25 compounds, catalysts, enzymes, enzymatic substrates, dyes, colloidal metallic and nonmetallic particles, and organic polymer latex particles.

The term "luminescence" refers to any emission of light that does not derive energy from the temperature of an energy source (for example, a source of electromagnetic radiation, a chemical reaction, mechanical energy). In general, the source causes an
30 electron of an atom to move from a lower energy state into an "excited" higher energy state; then the electron releases that energy in the form of emitted light when it falls back to a lower energy state. Such emission of light usually occurs in the visible or near-visible range of the electromagnetic spectrum. The term "luminescence" includes, but is not limited to, such light emission phenomena such

as phosphorescence, fluorescence, bioluminescence, radioluminescence, electroluminescence, electrochemiluminescence and thermo-luminescence.

The term "luminescent label" refers to a label that generates a luminescent signal, e.g. an emission of light that does not derive energy from the temperature of the emitting source. The luminescent label may be, for example, a fluorescent molecule, a phosphorescent molecule, a radioluminescent molecule, a luminescent chelate, a phosphor or phosphor-containing compound, or a quantum dot.

An "electrochemiluminescence assay" or "ECLA" is an electrochemical assay in which bound analyte molecule is detected by a label linked to a detecting agent (target molecule). An electrode electrochemically initiates luminescence of a chemical label linked to a detecting agent. Light emitted by the label is measured by a photodetector and indicates the presence or quantity of bound analyte molecule/target molecule complexes. ECLA methods are described, for example, in U.S. Patent Nos. 5,543,112; 5,935,779; and 6,316,607. Signal modulation can be maximized for different analyte molecule concentrations for precise and sensitive measurements.

In an ECLA procedure microparticles can be suspended in the sample to efficiently bind to the analyte. For example, the particles can have a diameter of 0.05 μm to 200 μm , 0.1 μm to 100 μm , or 0.5 μm to 10 μm , and a surface component capable of binding an analyte molecule. In one frequently used ECLA-system (Elecsys®, Roche Dagnsotics, Germany), the microparticles have a diameter of about 3 μm . The microparticles can be formed of crosslinked starch, dextran, cellulose, protein, organic polymers, styrene copolymer such as styrene/butadiene copolymer, acrylonitrile/butadiene/styrene copolymer, vinylacetyl acrylate copolymer, vinyl chloride/acrylate copolymer, inert inorganic particles, chromium dioxide, oxides of iron, silica, silica mixtures, proteinaceous matter, or mixtures thereof, including but not limited to sepharose beads, latex beads, shell-core particles, and the like. The microparticles are preferably monodisperse, and can be magnetic, such as paramagnetic beads. See, for example, U.S. Patent Nos. 4,628,037; 4,965,392; 4,695,393; 4,698,302; and 4,554,088. Microparticles can be used in an amount ranging from about 1 to 10,000 $\mu\text{g/ml}$, preferably 5 to 1,000 $\mu\text{g/ml}$.

The expression "of interest" denotes an analyte or substance of possible relevance that shall be analyzed or determined.

"Detection" includes any means of detecting, including direct and indirect detection. The term "detection" is used in the broadest sense to include both qualitative and quantitative measurements of an analyte, herein measurements of an analyte. In one aspect, a detection method as described herein is used to identify the mere presence of an analyte of interest in a sample. In another aspect, the method can be used to quantify an amount of analyte in a sample.

To "reduce" or "inhibit" is to decrease or reduce an activity, function, and/or amount as compared to a reference. By reduce or inhibit is meant the ability to cause an overall decrease preferably of 10% or greater, more preferably of 25% or greater, and most preferably of 50%, 75%, 90%, 95%, or greater.

To "enhance", e.g. to "enhance specific signals" or "the enhancement of ECL signals", is to increase or rise an activity, function, and/or amount as compared to a reference. By increase or rise is meant the ability to cause an overall increase preferably of 10% or greater, more preferably of 25% or greater, and most preferably of 50% or greater.

The term "determining" is used here for both qualitative and quantitative detection of an analyte, and can include determination of the amount and/or concentration of the analyte.

The term "measuring"/"measurement" in science is the process of estimating or determining the magnitude of a quantity, such as length or mass, relative to a unit of measurement, such as a meter or a kilogram. Measuring/measurement uses a reference point against which other things can be evaluated. The term measurement can also be used to refer to a specific result (determined values) obtained from a measurement process. It is a basis for comparison. The skilled artisan is aware of materials and methods to correlate measured signals or determined values to concentrations.

A "reagent composition" or "ECL-reagent composition" according to the present invention comprises reagents supporting ECL-signal generation, e.g. a coreactant, a buffering agent for pH control, and optionally other components. The skilled artisan is aware of components to be present in a reagent composition which are required for ECL signal generation in electrochemiluminescent detection methods.

An "aqueous solution" as used herein is a homogeneous solution of particles, substances or liquid compounds dissolved in water. An aqueous solution may also

comprise organic solvents. Organic solvents are known to the person skilled in the art, e.g. methanol, ethanol or dimethylsulfoxid. As used herein it is also to be understood that an aqueous solution can comprise at most 50% organic solvents.

5 A species that participates with the ECL label in the ECL process is referred to herein as ECL "coreactant". Commonly used coreactants for ECL include tertiary amines (e.g. tripropylamine (TPA)), oxalate, and persulfate. The skilled artisan is aware of available coreactants useful for electrochemiluminescent detection methods.

10 A "solid phase", also known as "solid support", is insoluble, functionalized, polymeric material to which library members or reagents may be attached or covalently bound (often via a linker) to be immobilized or allowing them to be readily separated (by filtration, centrifugation, washing etc.) from excess reagents, soluble reaction by-products, or solvents. Solid phases for the method according to the invention are widely described in the state of the art (see, e.g., Butler, J. E.,
15 Methods 22 (2000) 4-23). The term "solid phase" means a non-fluid substance, and includes particles (including microparticles and beads) made from materials such as polymer, metal (paramagnetic, ferromagnetic particles), glass, and ceramic; gel substances such as silica, alumina, and polymer gels; capillaries, which may be made of polymer, metal, glass, and/or ceramic; zeolites and other porous
20 substances; electrodes; microtiter plates; solid strips; and cuvettes, tubes, chips or other spectrometer sample containers. A solid phase component of an assay is distinguished from inert solid surfaces with which the assay may be in contact in that a "solid phase" contains at least one moiety on its surface, which is intended to interact with the capture antibody or capture molecule. A solid phase may be a
25 stationary component, such as a tube, strip, cuvette, chip or microtiter plate, or may be non-stationary components, such as beads and microparticles. Microparticles can also be used as a solid phase for homogeneous assay formats. A variety of microparticles that allow either non-covalent or covalent attachment of proteins and other substances may be used. Such particles include polymer particles such as
30 polystyrene and poly(methylmethacrylate); gold particles such as gold nanoparticles and gold colloids; and ceramic particles such as silica, glass, and metal oxide particles. See for example Martin, C.R., et al., Analytical Chemistry-News & Features (1998) 322A-327A, which is incorporated herein by reference.

35 The terms "chip", "bio-chip", "polymer-chip" or "protein-chip" are used interchangeably and refer to a collection of a large number of probes, markers or

biochemical markers arranged on a shared substrate (e.g. a solid phase) which could be a portion of a silicon wafer, a nylon strip, a plastic strip, or a glass slide.

Methods:


In an embodiment the present invention concerns a method for measuring an analyte in a sample via electrochemiluminescent detection, comprising the steps of

5 a) incubating the sample with a detection reagent labeled with an electrochemiluminescent group, b) separating analyte-bound and analyte-free labeled detection reagent, c) incubating the separated labeled detection reagent with a reagent composition comprising i) at least one coreactant, and ii) at least one

10 compound selected from the group of carbonic acid amides of Formula I and Formula II,

Formula I

15




20

with $R_1 =$ with $R_1 = \text{CH}_3, \text{CH}_2\text{F}, \text{CH}_2\text{Cl}, \text{CH}_2\text{CH}_3, \text{CHClCH}_3, \text{CH}_2\text{CH}_2\text{Cl}, \text{C}(\text{CH}_3)_2\text{CH}_3, \text{CH}_2\text{CH}_2\text{CH}_3, \text{CClHCH}_2\text{CH}_3$ or $\text{CH}_2\text{CH}_2\text{CH}_2\text{CH}_3$, with $R_2 = \text{H}$, and with $R_3 = \text{H}$,

Formula II,

25



30

d) electrochemically triggering the release of luminescence, and e) determining the electrochemiluminescence (ECL) signal thereby measuring the analyte.

35

An aspect of the invention relates to improved ECL methods based on the reagent compositions of the present invention, particularly ECL methods featuring low detection limits. The reagent compositions surprisingly enhance specific signals and reduce background signals. More specifically, the methods of the invention

provide improved sensitivity at low detection levels by reducing the background electrochemiluminescence in the absence of ECL labels.

5 The inventors have surprisingly discovered that the use of certain compounds from the group of carbonic acid amides provide a number of advantages including improved signal generation in ECL detection methods and thus improved ECL assay performance.

10 A feature of the invention are methods for the determination of an analyte in a sample to be investigated using an electrochemiluminescent label, wherein one of the following listed methods for measuring electrochemiluminescent phenomena is employed.

15 Surprisingly the methods using compounds selected from the group of carbonic acid amides emit less background luminescence than conventional test reagents without these compounds. This is particularly an advantage at low detection levels where increasing the signal to background ratio (= signal to noise ratio) greatly improves the sensitivity. Surprisingly the inventors have found that performing an electrochemiluminescent detection using a method according to the present invention results in a 10% to 60% improved signal to noise ratio of ECL detection.

20 The method for measuring an analyte in a sample via electrochemiluminescent detection according to the present invention can be performed in an embodiment in an aqueous solution.

In an embodiment the carbonic acid amide used in the method is selected from the group consisting of acetamide, 2-fluoroacetamide, 2-chloroacetamide, propanamide, 2-chloropropanamide, 3-chloropropanamide, butanamide and 2-chlorobutanamide.

25 In a preferred embodiment the carbonic acid amide used in the method is selected from the group consisting of acetamid, 2-chloroacetamide, propanamide and butanamide.

In a preferred embodiment the carbonic acid amide used in the method is selected from the group consisting of acetamid, propanamide and butanamide.

30 In a preferred embodiment the carbonic acid amide is used in the method in a concentration of 0.01 M to 0.25 M. In a further preferred embodiment the carbonic

acid amide is used in a concentration of 0.01 M to 0.2 M. In a further preferred embodiment the carbonic acid amide is used in a concentration of 0.01 M to 0.1 M.

5 In an embodiment the method according to the present invention is particularly well suited to detect biomolecules, such as proteins, polypeptides, peptides, peptidic fragments, hormones, petid hormones, vitamins, provitamins, vitamin metabolites and amino acids in a sample of interest.

10 The sample used in the methods according to the present invention is in an embodiment a liquid sample, e.g., whole blood, serum or plasma. The sample, or more specific the sample of interest, in an embodiment may comprise any body fluid and stool. In an embodiment the sample will be a liquid sample like saliva, stool extracts, urine, whole blood, plasma or serum. In an embodiment the sample will be whole blood, plasma or serum.

15 It is known to a person skilled in the art that steps “a) incubating the sample with a detection reagent labeled with an electrochemiluminescent group” and “b) separating analyte-bound and analyte-free labeled detection reagent” in the method might be performed in the same location, e.g. in the same reaction vessel. Said steps (a) and (b) might be performed in an automatic process controlled by a control means.

20 Unspecific sample components and analyte-free labeled detection reagent can be removed in step (b) according to the method in a separation process. For example that analyte-bound and analyte-free labeled detection reagent can be separated using a washing step.

Also other test components supporting the electrochemiluminscent detection of an analyte may be used in the methods according to the present invention.

25 An aspect of the invention covers the need for effective preservation, e.g. for long term storage of reagent mixtures and reagent compositions. Suitable preservatives should have no effect on ECL signal generation or in an ideal case a positive influence on ECL signal generation.

30 As suitable preservative compounds boric acid and/or borate were identified that effectively control bacterial and fungal growth and suprisingly increase the specific ECL signals. An ECL detection method using a reagent composition comprising

boric acid and/or borate as preservative has the positive surprising effect of an increase in the specific ECL signal generated.

In one embodiment the present invention concerns a method for measuring an analyte in a sample via electrochemiluminescent detection, comprising the steps of

5 a) incubating the sample with a detection reagent labeled with an electrochemiluminescent group, b) separating analyte-bound and analyte-free labeled detection reagent, c) incubating the separated labeled detection reagent with a reagent composition comprising i) at least one coreactant, and ii) a preservative selected from the group consisting of boric acid and borate, d) electrochemically

10 triggering the release of luminescence, and e) determining the electrochemiluminescence (ECL) signal thereby measuring the analyte.

In an embodiment the method for measuring an analyte in a sample via electrochemiluminescent detection is characterized in that the reagent composition for ECL signal generation comprises a preservative selected from the group

15 consisting of boric acid and borate at a concentration of 0.1% to 5%, preferably at a concentrations of 0.5% to 4%, and more preferably at a concentration of 0.5% to 2%.

In an embodiment the method for measuring an analyte in a sample via electrochemiluminescent detection is characterized in that the reagent composition for ECL signal generation comprises boric acid as preservative at concentrations of

20 0.1% to 5%, preferably at concentrations of 0.5% to 4%, and more preferably at concentrations of 0.5% to 2%.

In an embodiment the method for measuring an analyte in a sample via electrochemiluminescent detection is characterized in that the reagent composition for ECL signal generation comprises borate as preservative at concentrations of

25 0.1% to 5%, preferably at concentrations of 0.5% to 4%, and more preferably at concentrations of 0.5% to 2%.

It has been found by the inventors, that a method for measuring an analyte in a sample via electrochemiluminescent detection combining the effect of carbonic acid amides and boric acid and/or borate in one reagent composition can result in a

30 further improved signal to noise ratio in ECL detection. The accumulated effect of carbonic acid amides and boric acid and/or borate in one reagent composition leads to at least 10%, 25% or 50% improved signal generation in ECL detection.

In an embodiment the present invention concerns a method for measuring an analyte in a sample via electrochemiluminescent detection, comprising the steps of a) incubating the sample with a detection reagent labeled with an electrochemiluminescent group, b) separating analyte-bound and analyte-free labeled detection reagent, c) incubating the separated labeled detection reagent with a reagent composition comprising i) at least one coreactant, ii) at least one compound selected from the group of carbonic acid amides of Formula I and Formula II and iii) at least one preservative selected from the group of boric acid and borate, d) electrochemically triggering the release of luminescence, and e) determining the electrochemiluminescence (ECL) signal thereby measuring the analyte.

In an embodiment the method for measuring an analyte in a sample via electrochemiluminescent detection is characterized in that the reagent composition comprises in addition a detergent and a buffering agent.

In an embodiment the method for measuring an analyte in a sample via electrochemiluminescent detection is characterized in that the reagent composition comprises in addition a salt and/or an anti-foam agent.

In an embodiment the invention relates to a method for conducting an electrochemiluminescence assay wherein electrochemiluminescence is induced in the presence of a reagent composition according to the present invention.

A typical ECL measurement process for an ECL immunoassay comprises multiple exchanges of liquids and/or mixtures in the ECL measurement cell (e.g. a flow cell). A typical ECL measurement process consists of several steps explained below.

The skilled artisan is aware that an ECL measurement cell has to be conditioned or regenerated before the ECL detection step takes place by rinsing said ECL measurement cell with a reagent composition according to the present invention and additionally the application of an electric potential. This step is one part of the process of determining analytes using ECL. It has been described in EP 1 051 621 that during this conditioning step a layer is formed on the surface of the measurement electrode(s) supporting the signal generation during the measurement of an analyte in an ECL measurement cell.

For a typical ECL measurement process, a reagent mixture is induced into the cleaned and conditioned ECL measurement cell through the fluid inlet channel into the ECL measurement cell cavity. This mixture is an incubate of the sample, reagents and magnetic particles. Said mixture induced into the measurement cell
5 may be surrounded by a reagent composition according to the present invention flowing in front and after said mixture.

In such an ECL immunoassay a detection reagent comprising complex-molecules which are labeled with an electrochemiluminescent group and which are characteristic for the analysis, are bound to these magnetic particles by a pair of
10 specific biochemical binding partners, e.g. streptavidin and biotin. The magnetic particles are for example coated with streptavidin-polymer, whereas biotin is bound to the complex-molecules.

In the ECL measurement cell the magnetic particles are trapped to the surface of an electrode together with the labeled complex-molecules bound thereto in the
15 magnetic field of a magnet arranged close to said electrode. The magnetic field is applied during a continuous flow of the mixture, whereby incubate and/or reagent composition discharges from the ECL measurement cell cavity through the fluid outlet channel.

After trapping the magnetic particles, a reagent composition according to the present invention containing an ECL coreactant is induced into the ECL
20 measurement cell in a next step, whereby the magnetic particles are washed by said reagent composition. This step of washing is to remove unbound components of said incubate from the electrode which potentially interfere with the electrochemical reaction.

Thereafter the release of the electrochemiluminescence (ECL) signal is electrochemically triggered by application of an electric potential, whereby the
25 intensity of the luminescence light is detected by means of a photosensor and may be evaluated as a measure for the concentration of the labeled complex-molecules on the magnetic particles located at the surface of the electrode, whereby this
30 concentration again serves as a measure for the concentration of the analyte in the sample.

After the electrochemiluminescence detection the ECL measurement cell usually is rinsed with a cleaning fluid.

An apparatus for carrying out detection methods by means of electrochemiluminescence is mentioned in the example section (Example 1, 2 or 3) or described in EP 1 892 524 (A1). Moreover, such an apparatus can comprise means for controlling the temperature of the measuring unit and/or a liquid vessel.

5 The measuring unit is understood to be a cell in which the electrochemiluminescence is measured. The liquid vessel can be a storage container, but also a feeding device; for example, a tube for the reagent solution, contained in the measuring unit during the measurement.

Compositions:

10 An aspect of the invention relates to an improved reagent composition for ECL-signal generation, which leads to enhanced signal to noise ratios. More specifically, the reagent composition of the invention provides improved sensitivity at low detection levels by reducing the background electrochemiluminescence in the absence of ECL labels. Surprisingly a reagent composition comprising

15 compounds like carbonic acid amides emit less background luminescence than conventional test reagents without these compounds. This is particularly an advantage at low detection levels where increasing the signal to background ratio (= signal to noise ratio) greatly improves the sensitivity. This improved reagent composition contains a compound from the group of carbonic acid amides as well

20 as other compounds supporting the method for generating ECL. Surprisingly the inventors have found that performing an electrochemiluminescent detection using a reagent composition according to the present invention results in a 10% to 60% improved signal to noise ratio of ECL detection.

25 An aspect of the invention relates to a reagent composition that gives high signal to background ratios in electrochemiluminescence assays. The signal difference between specific signals and background signals is increased. Such improved properties have been achieved through the identification of advantageous combinations of ECL coreactant, pH buffering agents, detergent and pH and, in particular, through the use of compounds selected from group consisting of

30 carbonic acid amides.

The reagent composition provides a suitable environment for efficiently inducing ECL labels to emit ECL and for sensitively measuring ECL labels via the measurement of ECL. The reagent composition of the invention may optionally comprise additional components including preservatives, detergents, anti-foaming

agents, ECL active species, salts, acidic and basic compounds for pH control (buffering agents), metal ions and/or metal chelating agents. The reagent composition of the invention may also include components of a biological assay, which in some cases may be labeled with an ECL label, including binding reagents, enzymes, enzyme substrates, cofactors and/or enzyme inhibitors. The invention also includes assay reagents, compositions, kits, systems and system components that comprise the reagent composition of the invention and, optionally, additional assay components. The invention also includes methods for conducting ECL assays using the reagent composition of the invention.

10 In an embodiment the current invention relates to a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, and ii) at least one coreactant.

In an embodiment the carbonic acid amide of the reagent composition is selected from the group consisting of acetamide, 2-fluoroacetamide, 2-chloroacetamide, propanamide, 2-chloropropanamide, 3-chloropropanamide, butanamide and 2-chlorobutanamide.

In a preferred embodiment the carbonic acid amide is selected from the group consisting of acetamid, 2-chloroacetamide, propanamide and butanamide. In a further embodiment the carbonic acid amide is selected from the group consisting of acetamid, propanamide and butanamide. Carbonic acid amides have individual concentration optima for the ECL enhancing effect. As shown in the experiments (especially table 2, 3 and 4) the skilled artisan is aware to select the appropriate concentration for the selected carbonic acid amide in the reagent composition. Methods to determine the optimal concentration for a carbonic acid amide in the reagent composition is known to the skilled artisan.

In an embodiment the reagent composition comprises the carbonic acid amides in a concentration of 0.01 M to 0.25 M. In a further embodiment the reagent composition comprises the carbonic acid amides in a concentration of 0.01 M to 0.2 M. In a further embodiment the reagent composition comprises the carbonic acid amides in a concentration of 0.01 M to 0.1 M.

The coreactant of the reagent composition in an embodiment is selected from the group of tertiary amines (e.g. tripropylamine (TPA)), oxalate, and persulfate. In a preferred embodiment the coreactant is TPA.

It may be beneficial when storing a reagent composition to include a preservative that prevents microbial growth. Additionally, suitable preservatives are identified to control bacterial and fungal growth to enable long term storage and use of the reagent composition. The reagent composition according to the present invention may additionally contain one or more preservatives. In an embodiment of the present invention the reagent composition comprises a preservative (preservative agent).

In an embodiment the current invention relates to a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant, and iii) at least one preservative.

Preferably, the preservative has no or a positive effect on ECL signal generation. It is known to a person skilled in the art, that oxazolidines (e.g., Oxaban A or 4,4 dimethyl oxazolidine), azide and related preservatives are compatible with ECL. Oxazolidines at concentrations of 0.01% to 1% are normally used in test reagents. In an embodiment the reagent composition comprises preservatives selected from the group of Oxazolidines, preferably Oxaban A. In an embodiment the reagent composition comprises preservatives in a concentration of 0.01% to 1%, in another embodiment the reagent composition comprises preservatives in a concentration of 0.1% to 1%. It might also be beneficial to use a mixture of two or more preservatives.

The carbonic acid amide 2-chloroacetamide (CAA), as already mentioned above, has in addition to its ECL signal enhancing effect also a preservative function.

As mentioned above, an aspect of the invention covers the need for adding an effective preservative that has no or a positive influence on ECL signal generation. As suitable inorganic compounds boric acid and/or borate were identified that effectively control bacterial and fungal growth. Surprisingly the inventors have found that boric acid and/or borate present in a reagent composition for determining ECL have no negative influence in ECL signal generation. It has surprisingly been found that a reagent composition comprising boric acid or borate as preservative has a positive effect on the ECL signal generation process, namely an increase of the specific signal. Additionally their high activity and low degree of problems associated with safety hazards or environmental concerns are advantageous. Boric acid or borate, is in contrary to some other commonly used

preservatives in reagent compositions, halogen free and does not release formaldehyde. Results with boric acid present in ECL signal generation are shown in the example section, e.g. Example 2.

- 5 In an embodiment the current invention relates to a reagent composition for determining ECL, comprising i) at least one coreactant, and iii) at least one preservative selected from the group consisting of boric acid and borate.

In an embodiment the current invention relates to a reagent composition for determining ECL, comprising i) at least one coreactant, and iii) the preservative boric acid.

- 10 In an embodiment the current invention relates to a reagent composition for determining ECL, comprising i) at least one coreactant, and iii) the preservative borate.

- 15 In an embodiment the current invention relates to a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant, iii) at least one preservative selected from the group consisting of boric acid and borate.

- 20 In an embodiment the reagent composition according to the present invention comprises boric acid or borate as preservative in a concentration of 0.1% to 5%, preferably in a concentration of 0.5% to 4%, and particularly preferred in a concentration of 0.5% to 2%.

In a preferred embodiment the reagent composition according to the present invention comprises boric acid as preservative at concentrations of 0.1% to 5%, preferably in a concentration of 0.5% to 4%, and particularly preferred in a concentration of 0.5% to 2%.

- 25 In a preferred embodiment the reagent composition according to the present invention comprises borate as preservative at concentrations of 0.1% to 5%, preferably in a concentration of 0.5% to 4%, and particularly preferred in a concentration of 0.5% to 2%.

- 30 The reagent composition according to the present invention optionally comprises in addition other test components. Other test components are selected from the group consisting of at least one detergent, at least one signal enhancing compound, buffering agents comprising acidic and basic agents for pH control, and water.

In an embodiment the current invention relates to a reagent composition for determining ECL comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant, iii) at least one preservative, iv) buffering agents, v) at least one detergent, vi) a salt and/or anti-foam agent, and vii) optionally other test components.

Suitable detergents for a reagent composition according to the present invention are those from the group consisting of fatty acid alcohol ethoxylates, including poly(ethylene glycol)ethers, for example polidocanol or other poly(ethylene glycol)ethers with the formula C_XEO_Y with $X = 8 - 18$ and $Y = 2 - 9$, genapol (isotridecylpoly((ethylene glycol ether)_n), Plantaren® (alkylpolyglucoside), octylglucoside (octyl-beta-D-glucopyranoside) as well as zwitterionic detergents like Zwittergent 3-12 or a mixture thereof. The detergents are used in concentrations ranging between 0.01% and 2%. The optimal concentration can be easily determined for each detergent. The most suitable concentrations are those ranging between 0.05% and 1%.

In an embodiment the reagent composition according to the present invention comprises detergents selected from the group consisting of polidocanol or other poly(ethylene glycol)ethers with the formula C_XEO_Y with $X = 8 - 18$ and $Y = 2 - 9$, octylglucoside (octyl-beta-D-glucopyranoside) or zwitterionic detergents like Zwittergent 3-12 or a mixture thereof. In an preferred embodiment the reagent composition comprises detergents selected from the group consisting of polydocolanol, octylglucoside (octyl-beta-D-glucopyranoside) and Zwittergent 3-12, or a mixture thereof.

Further the electrochemiluminescent signal can also be increased by adjusting the pH to a value between 6.0 and 8.0, preferably between 6.0 and 7.5, particularly preferred between 6.2 and 6.9. This can be done conventionally by using a pH buffering agent suitable for this range, known to a person skilled in the art. In an embodiment the buffering agent suitable for the reagent composition comprises KOH and phosphoric acid (H_3PO_4).

Further more, the signal can be increased by adding salts, including inorganic salts like, for example NaBr, NaCl, NaI. The salts, especially NaCl, are added in concentrations ranging between 1 mM and 1 M, preferably between 10 mM and 100 mM and most preferably between 10 mM and 50 mM.

It may be beneficial, especially in HTS applications, to avoid the production of bubbles or foam. For this reason it may be desirable to add anti-foaming agents to a reagent composition. Many commercial antifoaming agents (including Antifoams o-30, Antifoam 204, Antifoam A, Antifoam SE-15, Antifoam SO-25 and Antifoam 289) may be added to the reagent composition according to the present invention.

The reagent composition of the invention may include ECL labels. The ECL labels may be conventional ECL labels. Examples of ECL labels include tris-bipyridyl-ruthenium (RuBpy) and other organometallic compounds where the metal is from, for example, the metals of group VII and VIII, including Re, Ru, Ir and Os. These ECL labels are used by a person skilled in the art to label an analyte specific reagent with an electrochemiluminescent group, or to label the analyte itself with an electrochemiluminescent group. In an embodiment the reagent composition of the invention contains a labeled analyte and/or a labeled analyte specific reagent, wherein the ECL label is selected from the group consisting of ECL labels disclosed in US 5,310,687 (A) (BPRu = Ru(bpy)₂-bpyCO-OSu), US 2003/0124572 (A1) (Sulfo-BPRu NHS Ester), EP720614(A1) (Bpy₂-Ru-bpy-CO-UEEK-korks.-OSu) and WO 2002/027317 (A2) (BPRu-(UE)-25-K and BPRu₂-SK4), respectively.

The reagents and mixtures thereof used in the reagent composition might be provided either in liquid, frozen, deep frozen, vaporize frozen, lyophilized, gas, solid or dried form before usage. At least before usage of the reagent composition the reagents are solved in a solvent. The reagent compositions of the present invention will be an aqueous solution. In a preferred embodiment the reagents are solved in water.

These improved formulations are of particular value in high sensitivity assays. In some embodiments of the invention, the performance of ECL assays is improved even further through optimal combinations of reagent composition with electrode composition. Such suitable ECL electrode compositions comprise electrodes of Ir, Pt or Carbon.

These advantageous combinations including the afore mentioned ECL enhancing carbonic acid amides and suitable preservatives selected from the group consisting of boric acid and borate, which both have improved properties. These include a higher dynamic range and an improved ratio of ECL signal from bound label to ECL background signal using the disclosed reagent composition according to the

present invention. This increased sensitivity is important, for example, in assays that benefit from a lower detection limit (e.g. TroponinT assay (TNThs; Order-No.:05092744), Hepatitis-B envelope antigen assay (HBeAg; Order-No.: 11820583), anti-Thyrotropin receptor assay (anti-TSHR; Order-No.:04388780) –
5 for details see the example section).

These improved formulations of reagent compositions can give a better precision that may result in a lower detection limit in ECL assays.

Another aspect of the invention relates to a reduction of costs due to a reduction of the required volumes of sample, test-specific reagents and/or test reagent. The
10 signal loss for lower reagent volumes can be compensated by using the advantageous reagent composition according to the present invention.

Yet another aspect of the invention relates to improved systems and apparatus containing the reagent composition of the invention and/or improved systems and apparatus adapted to perform the improved methods of the invention.

15 The ECL signal generation can also be improved when the above findings are used either alone or in combination with each other.

Reagent mixture:

For the determination of ECL the reagent composition according to the present invention may be mixed with additional compounds forming a reagent mixture. In
20 an embodiment the current invention relates to a reagent mixture for determining ECL, comprising a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant, iii) a sample to be investigated and iv) at least one detection reagent labeled with an electrochemiluminescent group.

25 In an embodiment the current invention relates to a reagent mixture for determining ECL, comprising a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant, iii) at least one preservative iv) a sample to be investigated and v) at least one detection reagent labeled with an
30 electrochemiluminescent group.

In an embodiment the current invention relates to a reagent mixture for determining ECL, comprising a reagent composition for determining ECL, comprising i) a

compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant, iii) a preservative selected from the group consisting of boric acid and borate iv) a sample to be investigated and v) at least one detection reagent labeled with an electrochemiluminescent group.

- 5 In a further embodiment the current invention relates to a reagent mixture for determining ECL, comprising a reagent composition for determining ECL, comprising i) a preservative selected from the group consisting of boric acid and borate, ii) at least one coreactant, iii) a sample to be investigated and iv) at least one detection reagent labeled with an electrochemiluminescent group. In a preferred
10 embodiment the reagent mixture comprises the preservative boric acid. In a preferred embodiment the reagent mixture comprises the preservative borate.

- The current invention relates in an embodiment also to a reagent mixture for determining ECL, comprising a reagent composition for determining ECL, comprising i) a preservative selected from the group consisting of boric acid and
15 borate, ii) at least one coreactant, iii) a sample to be investigated, iv) a detergent, v) a buffering agent, vi) at least one detection reagent labeled with an electrochemiluminescent group, and vii) comprising a salt and/or an anti-foam agent.

- In a further preferred embodiment the current invention relates to a reagent mixture
20 for determining ECL, comprising a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant, iii) a preservative selected from the group consisting of boric acid and borate, iv) a sample to be investigated, v) a detergent, vi) a buffering agent and vii) at least one detection reagent labeled
25 with an electrochemiluminescent group.

The reagent mixture in addition may comprise at least one detergent and a buffering agent for pH control. Optionally the reagent mixture may comprise a salt and/or an anti-foam agent.

- Other test components in the reagent mixture are selected from the group consisting
30 of non labeled analyte specific reagents, analyte homologs, solid phase coatings and substances that reduce interference.

Use:

An aspect of the present invention relates to the use of an improved reagent composition and/or an improved reagent mixture of the invention for performing an electrochemiluminescent detection method.

5 In an embodiment the invention relates to the use of a carbonic acid amide selected from the group of Formula I and Formula II for performing an electrochemiluminescent detection. In an embodiment the current invention relates to the use of a carbonic acid amide selected from the group of Formula I and Formula II for performing an electrochemiluminescent detection method
10 procedure.

In an preferred embodiment the invention relates to the use of carbonic acid amides selected from the group consisting of acetamide, 2-fluoroacetamide, 2-chloroacetamide, propanamide, 2-chloropropanamide, 3-chloropropanamide, butanamide and 2-chlorobutanamide for performing an electrochemiluminescent
15 detection.

In another preferred embodiment the invention relates to the use of a carbonic acid amide selected from the group consisting of acetamid, 2-chloroacetamide, propanamide and butanamide for performing an electrochemiluminescent detection. In another embodiment the invention relates to the use of a carbonic acid
20 amide selected from the group consisting of acetamid, propanamide and butanamide for performing an electrochemiluminescent detection.

In an embodiment the current invention relates to the use of a reagent composition comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant, and iii) at least one
25 preservative for determination of ECL.

In an embodiment the current invention relates to the use of a reagent composition comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant, and iii) a preservative selected from the group consisting of boric acid and borate for determination of ECL.

30 In an embodiment the current invention relates to the use of a reagent composition comprising i) a compound selected from the group of carbonic acid amide selected from the group consisting of acetamide, 2-fluoroacetamide, 2-chloroacetamide,

propanamide, 2-chloropropanamide, 3-chloropropanamide, butanamide and 2-chlorobutanamide ii) at least one coreactant, and iii) a preservative for determining ECL.

5 The reagent composition according to the present invention is in an embodiment appropriate for conditioning or regeneration of an ECL measurement cell and for determining an ECL signal. In an embodiment said reagent composition is used as a conditioning solution. In an embodiment the reagent composition according to the present invention is used for conditioning or regeneration of an ECL measurement cell. In an embodiment said reagent composition comprising a
10 compound selected from the group of carbonic acid amide selected from the group consisting of acetamide, 2-fluoroacetamide, 2-chloroacetamide, propanamide, 2-chloropropanamide, 3-chloropropanamide, butanamide and 2-chlorobutanamide is used for the conditioning or regeneration of ECL measurement cells. In another embodiment said reagent composition comprising a compound selected from the
15 group of carbonic acid amide selected from the group consisting of acetamid, propanamide and butanamide is used for the conditioning or regeneration of ECL measurement cells.

For the use for performing an electrochemiluminescent detection method the reagent composition can be mixed with additional compounds, e.g. a sample to be
20 investigated, at least one detection reagent with an electrochemiluminescent group as well as other components mentioned below supporting the method forming a reagent mixture.

In an embodiment the current invention relates to the use of a reagent mixture comprising a reagent composition, a) comprising i) a compound selected from the
25 group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant, and iii) a preservative, b) a sample to be investigated, and at c) at least one detection reagent labeled with an electrochemiluminescent group in the determination of ECL.

The current invention relates to the use of a reagent mixture comprising a reagent
30 composition for determining ECL, a) comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant, and iii) a preservative selected from the group consisting of boric acid and borate, b) a sample to be investigated, and c) at least one detection reagent labeled with an electrochemiluminescent group in the determination of ECL.

5 In a further embodiment the invention relates to the use of boric acid or borate for performing an electrochemiluminescent detection. Also in an embodiment the current invention relates to the use of a preservative selected from the group consisting of boric acid and borate for performing an electrochemiluminescent detection method procedure.

10 In an embodiment the current invention relates to the use of a reagent mixture comprising a) a reagent composition for determining ECL, comprising i) a preservative selected from the group of boric acid and borate and ii) at least one coreactant, b) a sample to be investigated, and c) at least one detection reagent labeled with an electrochemiluminescent group in the determination of ECL.

15 In addition the reagent mixture used for determining ECL may comprise components selected from the group consisting of a detergent and a buffering agent for pH control. Optionally the used reagent mixture may comprise a salt and/or an anti-foam agent. Other test components in the reagent mixture are selected from the group consisting of non labeled analyte specific reagents, analyte homologs, solid phase coatings and substances that reduce interference.

Kit:

20 One aspect of the invention relates to kits comprising, in one or more containers, one or more components of the reagent composition of the invention. These components may be combined, optionally with additional reagents, to form the reagent composition of the invention. The kits may also comprise in an embodiment additional assay related components such as ECL labels, ECL labeled assay reagents, diluents, washing solutions, protein denaturing reagents, enzymes, binding reagents, assay plates, disposables, etc.

25 In an embodiment the reagent composition is contained in one or more glass or plastic containers, appropriately labeled with information regarding the reagent composition contents and instructions regarding proper storage and use. Information regarding the reagent composition contents, lot number, production date, best before date, instructions regarding proper storage and use may be also
30 stored on a RFID chip placed on the glass oder plastic containers. The information stored on such RFID chip can be read by an antenna connected to a RFID reader device and further processed in a control means.

In an embodiment some or all of the components of the reagent composition may be stored in an embodiment in a liquid or dry state.

5 In an embodiment the present invention concerns a kit for measuring ECL, which contains a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II and ii) at least one coreactant.

10 In a preferred embodiment the present invention concerns a kit for measuring ECL, which contains a reagent composition for determining ECL, comprising i) a carbonic acid amide selected from the group consisting of acetamide, 2-fluoroacetamide, 2-chloroacetamide, propanamide, 2-chloropropanamide, 3-chloropropanamide, butanamide and 2-chlorobutanamide and ii) at least one coreactant.

15 In another preferred embodiment the present invention concerns a kit for measuring ECL, which contains a reagent composition for determining ECL, comprising i) a carbonic acid amide selected from the group consisting of acetamide, 2-chloroacetamide, propanamide and butanamide and ii) at least one coreactant.

20 In an embodiment the present invention concerns a kit for measuring ECL, which contains a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant and iii) a preservative.

25 In a preferred embodiment the present invention concerns a kit for measuring ECL, which contains a reagent composition for determining ECL, comprising i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, ii) at least one coreactant and iii) a preservative selected from the group consisting of boric acid and borate.

30 In another preferred embodiment the present invention concerns a kit for measuring ECL, which contains a reagent composition for determining ECL, comprising i) a carbonic acid amide selected from the group consisting of acetamide, 2-fluoroacetamide, 2-chloroacetamide, propanamide, 2-chloropropanamide, 3-chloropropanamide, butanamide and 2-chlorobutanamide, ii) at least one coreactant and iii) a preservative selected from the group consisting of boric acid and borate.

In another preferred embodiment the present invention concerns a kit for measuring ECL, which contains a reagent composition for determining ECL, comprising i) a carbonic acid amide selected from the group consisting of acetamide, 2-chloroacetamide, propanamide and butanamide, ii) at least one coreactant and iii) a preservative selected from the group consisting of boric acid and borate.

In an embodiment the present invention concerns a kit for measuring ECL, which contains a reagent composition for determining ECL, comprising at least i) a preservative selected from the group consisting of boric acid and borate and ii) at least one coreactant.

The aforementioned measures per se already significantly improve the known procedures. Moreover, it is possible to further significantly increase the sensitivity and/or the dynamic measuring range of an analyte detection assay by combining these measures.

The following examples and figures are provided to aid the understanding of the present invention, the true scope of which is set forth in the appended claims. It is understood that modifications can be made in the procedures set forth without departing from the spirit of the invention.

Description of the Figures

Figure 1 Measurement results with propanamide concentrations of 0.001 M to 0.25 M (X-axis); relative recovery rate (% of Reference) of the measurement of Δ Artificial assay (artificial assay - assay buffer background), assay buffer background and free label assay are shown (Y-axis). See Example 1 for details.

Figure 2 Measurement results with 2-chloroacetamide concentrations of 0.001 M to 1 M (X-axis); relative recovery rate (% of Reference) of the measurement of Δ Artificial assay (artificial assay - assay buffer background), assay buffer background and a free label assay are shown (Y-axis). See example 1 for details.

Figure 3 Measurement results with butanamide concentrations of 0.001 M to 1 M (X-axis); relative recovery rate (% of Reference) of the measurement of Δ Artificial assay (artificial assay - assay buffer background), assay buffer background and free label assay are shown (Y-axis). See example 1 for details.

5 **Figure 4** Measurement results with acetamide concentrations of 0.001 M to 1 M (X-axis); relative recovery rate (% of Reference) of the measurement of Δ Artificial assay (artificial assay - assay buffer background), assay buffer background and free label assay are shown (Y-axis). See example 1 for details.

10 **Figure 5** Measurement results with boric acid concentrations of 0 to 5 % (X-axis); The artificial assay was used as an example for a high specific signal; the TSH calibrator 1 as a low level calibrator gives a background signal (TSH Cal 1); the TSH calibrator 2 gives a signal at a high detection level (TSH Cal 2). The results are plotted as % of the reference reagent composition without addition of boric acid. See example 2 for details.

Example 1

15 **ECL Measurements using assay buffers (reagent compositions) with carbonic acid amides**

ECL measurements were carried out using the Roche Elecsys® 2010 device using protocols available for the assays mentioned below.

20 Various concentrations of compounds selected from the group of carbonic acid amides as indicated in tables 2, 3 and 4 were added to the following assay buffer:

180 mM tripropylamine (TPA)
0.1% polidocanol
300 mM phosphate buffer

25 The final pH was adjusted to pH 6.8 using KOH/H₃PO₄. The assay buffer was also used as blank value.

30 The compounds selected from the group of carbonic acid amides (chemical formulas of carbonic acid amides are shown in Table 1) were added to the assay buffer (reagent composition) at the indicated concentrations. Results are reported as signal recovery relative to measurements using an assay buffer lacking these compounds.

Assay buffer background measurements with an assay buffer containing the carbonic acid amides at the concentration shown in Table 2 were performed. Values below 100% indicate a reduced ECL background signal by addition of the

selected carbonic acid amide at the indicated concentrations. Reducing the background electrochemiluminescence in the absence of ECL labels is particularly an advantage at low detection levels, where increasing the signal to background ratio (= signal to noise ratio) greatly improves the sensitivity of an assay.

5 **Table 2:**

Concentration	Assay Buffer Background				
	0.25 M	0.1 M	0.01 M	0.001 M	0.0001 M
2,2 dichloro-acetamide	30%	40%	70%	84%	93%
2-chloroacetamide	66%	69%	80%	94%	98%
2-chlorobutanamide	91%	79%	80%	92%	98%
2-chlor-N-hydroxymethylacetamide	0%	46%	97%	91%	97%
2-chlor-N,N-dimethylacetamide	64%	75%	92%	99%	98%
2-chlor-N-methoxy-N-methylacetamide	0%	0%	88%	99%	102%
2-chloropropanamide	70%	68%	78%	93%	97%
3-chloropropanamide	n.d.	n.d.	79%	91%	96%
Acetamide	79%	80%	90%	95%	97%
Acetoacetamide	n.d.	55%	85%	94%	98%
2-bromoacetamide	n.d.	n.d.	39%	72%	93%
Butanamide	68%	68%	81%	108%	108%
Formamide	74%	75%	81%	94%	98%
2-fluoroacetamide	73%	77%	91%	101%	102%
2-hydroxy-acetamide	93%	90%	96%	100%	101%
Hexanamide	56%	53%	67%	83%	90%
2-jodoacetamide	0%	0%	0%	0%	34%
Propanediamide	79%	84%	92%	91%	99%
N-methylpropanamide	88%	90%	94%	98%	100%
2,2 dimethylpropanamide	69%	68%	88%	98%	99%
propanamide (propionamide)	77%	74%	86%	95%	88%
2-pyrrolidone	88%	84%	96%	101%	100%
2,5-butanamide	135%	103%	105%	103%	98%
2,2,2-trifluoro-acetamide	98%	85%	87%	95%	99%
Pentanamide	75%	63%	73%	91%	97%

n.d. = not determined

In an analogous experiment the signals of free label were determined. The free label value represents the signal generated by a solution containing a free ECL label in the absence of microparticles (10 nM RuBpy in the assay buffer). This value is stated in Table 3 relative to the assay buffer without any additional compound in %. This assay format is also known as a homogenous measurement or homogeneous assay format. Values above 100% indicate an enhanced ECL signal

by addition of the selected carbonic acid amide at the selected concentration. Results are shown in Table 3.

Table 3:

Concentration	Free label assay				
	0.25 M	0.1 M	0.01 M	0.001 M	0.0001 M
2,2 dichloro-acetamide	33%	66%	145%	120%	104%
2-chloroacetamide	155%	151%	131%	109%	101%
2-chlorobutanamide	157%	147%	126%	108%	100%
2-chlor-N-hydroxymethylacetamide	0%	6%	103%	125%	99%
2-chlor-N,N-dimethylacetamide	125%	116%	103%	100%	100%
2-chlor-N-methoxy-N-methylacetamide	2%	5%	124%	0%	0%
2-chloropropanamide	113%	155%	132%	110%	102%
3-chloropropanamide	n.d.	n.d.	132%	107%	100%
acetamide	141%	132%	110%	102%	100%
acetoacetamide	6%	21%	130%	104%	97%
2-bromoacetamide	n.d.	n.d.	89%	139%	110%
butanamide	145%	136%	114%	103%	99%
formamide	134%	145%	127%	111%	103%
2-fluoroacetamide	144%	134%	111%	102%	100%
2-hydroxy-acetamide	130%	137%	109%	101%	99%
hexanamide	118%	105,7%	120%	111%	103%
2-jodoacetamide	0%	0%	2%	3%	127%
propanediamde	123%	135%	115%	115%	100%
N-methylpropanamide	106%	99%	97%	98%	98%
2,2 dimethylpropanamide	125%	120%	106%	99%	97%
propanamide (propionamide)	144%	135%	113%	101%	99%
2-pyrrolidone	142%	130%	107%	100%	99%
2,5-butanamide	102%	136%	107%	100%	99%
2,2,2-trifluoro-acetamide	130%	150%	116%	103%	100%
pentanamide	151%	140%	122%	108%	103%

n.d. = not determined

- 5 Additionally, the values using the a simplified assay including beads were determined. This artificial assay is an assay including RuBpy labeled microparticles for a high specific signal. This assay format is also known as a heterogeneous measurement or heterogeneous assay format. The difference between the specific artificial assay-signal and the background signal (assay buffer background) using the assay buffer as described above with the additional compounds as indicated as ΔArtificial assay in % relative to assays buffer without any additional compound. Values above 100% indicate an enhanced ECL signal by
- 10

addition of the selected carbonic acid amide at the selected concentration. Results are shown in Table 4.

Table 4:

	ΔArtificial assay = (artificial assay – assay buffer background)				
Concentration	0.25 M	0.1 M	0.01 M	0.001 M	0.0001 M
2,2 dichloro-acetamide	73%	67%	92%	122%	109%
2-chloroacetamide	160%	165%	148%	113%	104%
2-chlorobutanamide	119%	152%	149%	119%	102%
2-chlor-N- hydroxymethylacetamide	100%	54%	109%	121%	107%
2-chlor-N,N-dimethylacetamide	97%	115%	113%	103%	104%
2-chlor-N-methoxy-N- methylacetamide	100%	100%	108%	100%	95%
2-chloropropanamide	63%	109%	145%	119%	104%
3-chloropropanamide	n.d.	n.d.	148%	111%	104%
acetamide	138%	144%	123%	110%	106%
acetoacetamide	n.d.	61%	125%	107%	98%
2-bromoacetamide	n.d.	n.d.	68%	129%	114%
butanamide	159%	161%	131%	95%	91%
formamide	51%	76%	116%	110%	105%
2-fluoroacetamide	148%	148%	118%	100%	97%
2-hydroxy-acetamide	32%	85%	108%	100%	99%
hexanamide	52%	59%	81%	113%	109%
2-jodoacetamide	0%	0%	0%	0%	125%
propanediamde	50%	84%	112%	108%	100%
N-methylpropanamide	104%	106%	104%	102%	100%
2,2 dimethylpropanamide	147%	149%	119%	103%	99%
propanamide (propionamide)	157%	163%	133%	111%	116%
2-pyrrolidone	117%	144%	116%	100%	99%
2,5-butanamide	-11%	84%	101%	98%	103%
2,2,2-triflouro-acetamide	47%	130%	133%	106%	100%
pentanamide	129%	147%	136%	119%	107%

n.d. = not determined

- 5 In Figure 1 a graph containing results for propanamide as shown in Tables 2, 3 and 4 is presented. In Figure 2 a graph containing results for 2-chloroacetamide as shown in Tables 2, 3 and 4 is presented. In Figure 3 a graph containing results for butanamide as shown in Tables 2, 3 and 4 is presented. In Figure 4 a graph containing results for acetamide as shown in Tables 2, 3 and 4 is presented.

Example 2**Boric acid as a signal enhancing preservative**

ECL measurements were carried out using the Roche Elecsys® 2010 device using the recommended protocols for the assays mentioned below.

5 The following ECL assay buffer was used to determine the blank value:

180 mM tripropylamine (TPA)

0.1% polidocanol

0.1% Oxaban A

300 mM phosphate buffer

10

To this assay buffer increasing amounts of boric acid were added as indicated. The final pH was adjusted to pH 6.8 using KOH/H₃PO₄.

15

Assay buffer background measurements with an assay buffer containing boric acid at the concentrations shown in Table 5 were performed. The free label value represents the signal generated by a solution containing a free ECL label in the absence of microparticles (10 nM RuBpy in the assay buffer, homogenous measurement) relative to the assay buffer without any additional compound in %. The artificial assay is an assay including RuBpy labeled microparticles for a high specific signal. As a commercial in vitro diagnostic assay, the Elecsys® TSH assay (Thyrotropin assay for Elecsys®; Order-No.: 11731459) was used to determine Δ TSH. The TSH calibrator 1 (TSH Cal set; Order-No.: 04738551) as a low level calibrator (no analyte present) was used in the TSH assay giving a background signal (TSH Cal 1); the TSH calibrator 2 was used in the TSH assay to give a high signal value (TSH Cal 2).

20

25

The results for the artificial assay, TSH Cal1 and TSH Cal 2 are plotted in Figure 5 as the relative recovery in % of the reference assay buffer without addition of boric acid.

In particular the following measurements were performed.

Table 5:

Relative recovery (% of Reference) - Comparision to an assay buffer without boric acid						
Boric acid conc.	0%	0.1%	0.5%	1.0%	2.0%	5.0%
Assay buffer background	100%	103%	102%	98%	103%	103%
Free label assay	100%	100%	99%	97%	93%	93%
Artificial assay	100%	104%	108%	107%	108%	108%
TSH Cal2	100%	105%	108%	108%	109%	112%
TSH Cal1	100%	100%	98%	95%	98%	94%

- 5 Addition of boric acid as a preservative in an assay buffer improves the specific heterogenous signal, especially in the artificial assay and in the determination of TSH Cal2.

Example 3

Effect of assay buffers containing propanamide and boric acid on the lower detection limit of Elecsys® assays.

- 10 The lower detection limit with several commercial in vitro diagnostic assays (HBeAg: Roche Order-No.: 11820583, Anti-TSHR: Roche Order-No.: 04388780, TNThs: Roche Order-No.: 05092744) was determined to compare two assay buffer preparations.

Assay buffer A:

- 15 180 mM TPA, 0.1 % polidocanol, 300 mM phosphate buffer, 0.1% Oxaban A

Assay buffer B:

180 mM TPA, 0,1% polidocanol, 50 mM propanamide, 300 mM phosphate buffer, 1% boric acid

20

The final pH of both assay buffers A and B was adjusted to pH 6.8 using KOH/H₃PO₄.

The three commercially available assays mentioned above have been analyzed to show the effect of the carbonic acid amide propanamide and the preservative boric acid on the assay performance detecting very low analyte concentrations.

The assays were measured on a Roche Elecsys® analyzer and calibrated as described in their package inserts. To calculate the lower detection limit the signals of a sample without analyte (HBeAg, anti-TSHR) or with a very low analyte concentration (TNThs) were determined. The standard deviation of the 21-fold determination was calculated, multiplied by 2 (2SD) or 3 (3SD) and added (HBeAg, TNThs), or subtracted (antiTSHR, competitive assay) to the mean of the signal. The corresponding concentration of the calculated signals was then determined using the calibration curve for each assay. For samples with a low analyte concentration (TNThs) the analyte concentration of the sample was subtracted from these calculated concentrations.

The three assays benefit from the improved reagent composition containing a carbonic acid amide according to the present invention as well as containing boric acid, which has also a preservative function. The results for HBeAg, anti-TSHR and TNThs assays are shown in Tables 6, 7, and 8, respectively.

Table 6:

	Lower detection limit [u/ml]	
HBeAg	2 SD	3 SD
Assay buffer A	0.0030	0.0044
Assay buffer B	0.0018	0.0030

Table 7:

	Lower detection limit [u/ml]	
anti-TSHR	2 SD	3 SD
Assay buffer A	0.324	0.500
Assay buffer B	0.218	0.332


Table 8:

	Lower detection limit [pg/ml]	
TNThs	(2SD) – Conc. of sample	(3SD) – Conc. of sample
Assay buffer A	1.193	1.832
Assay buffer B	0.782	1.140

Claims


1. A method for measuring an analyte in a sample via electrochemiluminescent detection, comprising the steps of:
 - a) incubating the sample with a detection reagent labeled with an electrochemiluminescent group,
 - b) separating analyte-bound and analyte-free labeled detection reagent,
 - c) incubating the separated labeled detection reagent with a reagent composition comprising
 - i) at least one coreactant, and
 - ii) at least one compound selected from the group of carbonic acid amides of Formula I and Formula II,

Formula I



with R₁ = with R₁ = CH₃, CH₂F, CH₂Cl, CH₂CH₃, CHClCH₃, CH₂CH₂Cl, C(CH₃)₂CH₃, CH₂CH₂CH₃, CClHCH₂CH₃ or CH₂CH₂CH₂CH₃, with R₂ = H, and with R₃ = H,

Formula II,



- d) electrochemically triggering the release of luminescence, and
 - e) determining the electrochemiluminescence (ECL) signal thereby measuring the analyte.
2. The method according to claim 1, characterized in that the measurement of an analyte in a sample using ECL is performed in an aqueous solution.

3. The method according to claims 1 or 2, characterized in that the carbonic acid amide is selected from the group consisting of acetamide, 2-fluoroacetamide, 2-chloroacetamide, propanamide, 2-chloropropanamide, 3-chloropropanamide, butanamide and 2-chlorobutanamide.
4. The method according to any of claims 1 to 3, characterized in that the reagent composition comprises a carbonic acid amide in a concentration of 0.01 M to 0.25 M.
5. The method according to any of claims 1 to 4, characterized in that the reagent composition of step 1c) comprises a preservative.
6. The method according to claim 5, characterized in that the reagent composition of step 1c) comprises a preservative in a concentration of 0.1% to 5%.
7. The method according to claims 5 or 6, characterized in that the reagent composition of step 1c) comprises a preservative selected from the group consisting of boric acid and borate.
8. The method according to any of claims 1 to 7, characterized in that the reagent composition of step 1c) comprises a detergent and a buffering agent.
9. The method according to claim 8, characterized in that the reagent composition of step 1c) comprises in addition a salt and/or an anti-foam agent.
10. A reagent composition for determining ECL, comprising
 - i) a compound selected from the group of carbonic acid amides of Formula I and Formula II, and
 - ii) at least one coreactant,wherein the coreactant is selected from the group consisting of tertiary amines (e.g. tripropylamine (TPA)), oxalate, and persulfate.
11. The reagent composition according to claim 10, characterized in that the reagent composition contains in addition a preservative selected from the group consisting of boric acid and borate.

12. A reagent mixture for determining ECL, comprising a reagent composition according to claims 10 or 11, a sample to be investigated and at least one detection reagent labeled with an electrochemiluminescent group.
13. Use of a reagent composition according to claims 10 or 11 in the determination of ECL.
14. Use of a carbonic acid amide selected from the group of Formula I and Formula II for performing an electrochemiluminescent detection procedure.
15. A kit for measuring ECL, which contains a reagent composition according to claims 10 or 11.