

1 -

HAEMOSTATIC DEVICE, HAEMOSTATIC COATING DISPERSION AND HYDROPHOBIC SURFACE

Technical Field

5

The present invention relates to a haemostatic coating dispersion, a hydrophobic surface formed using the haemostatic coating dispersion and a haemostatic device. The present invention relates to, but is not limited to, haemostatic patches for wound treatment.

10

Background

Uncontrolled haemorrhage and wound infection are leading causes of death in the medical field of wound care. Improperly dressed wounds will prolong healing
15 time and impose a high infection risk, leading to significantly increased mortality and economic burden. Haemostatic materials are consequently of great importance in medicine, in the successful repair and treatment of wounds.

Their successful implementation depends on two generally competing
20 attributes: a) achieving blood coagulation rapidly, before significant blood loss, and b) enabling subsequent facile wound-dressing removal, without clot tears and secondary bleeding. The conventional method to deal with bleeding is mechanically pressing the wound with a cotton gauze. Cotton gauze unavoidably absorbs blood and causes unnecessary blood loss and gauze adhesion onto the
25 wound. Blood absorbed in the gauze forms a solid clot-gauze composite, forced peeling of which often tears the wound and causes secondary bleeding and pain. This makes it difficult to replace the old wound dressing without causing secondary infections or haemorrhage, in procedures ranging from common wounds to surgery, and to the extreme case of haemophilic patients, where
30 excessive bleeding will occur before coagulation.

2 -

To deal with these problems, active clotting agents (chitosan or kaolin) have been adopted into haemostatic materials, to reduce bleeding by expediting the coagulation process. However, such agents employ free micro-particles that may cause micro-thrombosis if they enter the vascular system, and can thus become
5 a safety threat.

Recently, superhydrophobic (SHP) or superhydrophilic materials have been proposed for haemostatic purposes. A super-hydrophilic material (graphene sponge) is reported to absorb water from the blood quickly, forming a dense
10 layer of blood cells and platelets, thus promoting coagulation. Hydrophilic haemostatic material can also be prepared by spray coating β -chitosan on a porous nanofiber mat, and the hydrophilic β -chitosan coating can increase blood wettability and thus enhance clotting. Alternatively, a SHP coating can be applied on the back of the normal superhydrophilic gauze as an impervious layer
15 to prevent blood loss through the gauze. However, the core functionality of these approaches is still either based on a blood-absorbing haemostatic material (superhydrophilic) that does not minimize blood loss and secondary bleeding or a blood-repelling (superhaemophobic) material that simply repels blood but does not actively trigger clotting. Therefore, the aforementioned two key
20 challenges on wound management still remain poorly addressed.

It is desirable that a device or material be provided to overcome or ameliorate at least one of the above-described problems, or at least to provide a useful
25 alternative.

Summary

Described herein is a strategy for achieving haemostasis by designing a SHP and blood-repelling surface that simultaneously achieves fast clotting with no blood
30 loss, while also providing an anti-bacterial property and facilitating clot self-detachment.

3 -

The non-wetting feature of the SHP haemostatic surface can withstand substantial blood pressure and help reduce blood loss and bacteria attachment. As shown in investigations described with reference to the figures, carbon
5 nanofibers (CNFs) immobilized on this surface can promote fast fibrin growth and thus clotting. Due to the presence of micro-air pockets within the blood-substrate contact area, there is minimal contact between the clot and the SHP CNF patch, leading to natural clot detachment after clot maturation and shrinkage. This reduces the peeling tension required to peel off the patch by
10 about 1~2 orders of magnitude compared with a normal hydrophilic gauze or commercial haemostatic products. These features have been verified *in vitro* and *in vivo*, demonstrating the effectiveness of this strategy for designing haemostatic patch materials.

15 Described herein is a haemostatic device comprising:
a substrate; and
a surface formed on the substrate, the surface comprising at least one of micro- and nano-sized materials, the materials being partially embedded in a base, the surface substantially preventing wetting of the substrate. The surface
20 may also be referred to as a coating. In some embodiments, the surface may be composed of the materials.

The materials may be hydrophobic. The materials may comprise hydrophobic nanofibers immobilized in the base.

25 The surface may have random surface morphology to obtain a water contact angle of at least 130°.

The surface morphology may entrain air pockets between the surface and a
30 liquid coming into contact with the surface.

4 -

The haemostatic device may be one of a wound dressing, catheter-based stent, coil or graft – e.g. a medical patch with the surface formed on a cotton, synthetic fabric or polymeric membrane.

5 Also disclosed herein is a haemostatic coating dispersion comprising at least one of micro- and nano-sized materials, and a base (e.g. organic matrix) in a dispersion. The coating dispersion is for depositing on a substrate to form a hydrophobic surface comprising the micro/nano-materials partially embedded in the base. The obtained surface is substantially preventing wetting of the
10 substrate.

The micro/nano-materials may be hydrophobic when at least partially embedded. These materials may comprise nanofibers immobilized in the base on formation of the surface. The nanofibers may be carbon nanofibers. The
15 surface morphology may entrain air pockets between the surface and a liquid coming into contact with the surface.

The surface, once formed, may have a random surface morphology comprising micro-roughness and/or nano-roughness. The materials may comprise
20 nanofibers having a diameter ranging from about 5nm to 1 micron or microstructures having a diameter from about 1 μm to 200 μm . The materials may have a length of about 5 to 500 μm .

The base may be hydrophobic. The base may be an organic or polymeric matrix.
25 The matrix may be at least one of polytetrafluoroethylene (PTFE), beeswax and polydimethylsiloxane (PDMS), or other biocompatible polymers capable of immobilizing the nanofibers.

Also disclosed herein is a hydrophobic surface formed by applying the
30 haemostatic coating dispersion described above.

5 -

Advantageously, embodiments of the present invention overcome issues in the prior art relating to excessive blood loss during the period that the clot is forming and/or strong clot adhesion on the haemostatic dressing that causes pain, secondary bleeding, and possible infection during the wound-dressing removal.

5

Advantageously, embodiments of the present invention provide a haemostatic coating and device that can achieve (1) fast clotting, (2) clotting without blood loss, (3) clot self-detachment after clot maturation or solidification, and (4) minimize the adhesion of bacteria.

10

Advantageously, the haemostatic coating / device is superhydrophobic, providing enhanced repelling of water-based fluids such as blood. The present coating / device will thus reduce blood loss, as it repels blood to contain or restrict it within the wound. Moreover, the coating provides a non-wetting feature that can eliminate blood infiltration into the haemostatic material, thus preventing unnecessary blood loss. The non-wetting feature remains under high pressure of up to 300 mmHg, and thus will make the superhydrophobic haemostatic coating capable of dealing with artery bleeding. Moreover, the non-wetting property of the superhydrophobic CNF coating can prevent blood loss at the wound site by repelling blood and keeping blood within the wound. Blood will not penetrate through the material. This reduces blood loss. The coating is also capable of triggering fast clotting which can narrow the time window for bleeding, thus reducing blood loss during coagulation.

15

20

25

Advantageously, bacteria cannot attach to the coating easily. The anti-bacterial property can help to maintain the sterility of the haemostatic bandage and prevent wound infection.

30

Advantageously, the superhydrophobic haemostatic bandage or dressing or gauze can be easily peeled off from the clotted wound, without tearing the wound and causing secondary bleeding. In some cases, it facilitates self-

6 -

detachment of the clot – e.g. during maturation and shrinkage.

Advantageously, the device is safe. Carbon nanofibers are immobilized onto the dressing or gauze by a biocompatible polymer. The immobilized carbon
5 nanofibers would thus not enter the blood stream.

Brief description of the drawings

Embodiments of the present invention will now be described, by way of non-
10 limiting example, with reference to the drawings in which:

Figure 1 illustrates a device in accordance with present teachings;

Figure 2 schematically represents a method for forming the device of Figure 1;
15

Figure 3 shows the effect of different CNF nano-composite coatings on hydrophobicity and surface morphology, in which:

Figure 3a shows the effect of concentrations on water contact angle, roll-off angle, and blood contact angle;

20 Figure 3b shows SEM images of the hydrophobic surfaces HP #1 and HP #2; and

Figure 3c shows SEM images of superhydrophobic nano-composite CNF surfaces;

25 Figure 4 shows photographs of fibrin fibre generation on SHP CNF surfaces, in which:

Figure 4a is a scanning electron microscope (SEM) image of the SHP CNF/PTFE Ti surface and a water droplet sitting on the surface, demonstrating excellent superhydrophobicity – scale is 10 μm ;

30 Figure 4b shows long fibrin fibres generated at the receding side of a 10 μl EDTA PPP droplet;

7 -

Figure 4c illustrates a touch-lift test of a PPP droplet on the CNF surface, showing fibre generation upon EDTA PPP-substrate separation; and

Figures 4d to 4g show close-up sequential frames of fibre generation and fracture during rolling down motion of the EDTA PPP droplet (20 μ l), with Figure 4e showing visible fibrin fibre footprints left on the CNF surface after the PPP sliding test;

Figure 5 schematically illustrates the setup used to visualize fibrin fibre generation;

Figure 6 provides a schematic and photos of fibrin fibres generating on the superhydrophobic CNF/PTFE Ti surface after EDTA blood/plasma sliding, in which:

Figure 6a illustrates blood/plasma sliding;

Figures 6b and 6c are optical images showing fibrin fibres left on the surface after PPP or blood sliding, and visible fibrin fibre footprints left on the CNF surface after sliding tests;

Figure 6d provides SEM images of the CNF surface after blood sliding;

Figure 6e provides a SEM image of the long, straight and aligned fibrin fibres left on the CNF surface after the PPP droplet rolled down the superhydrophobic CNF surface; and

Figure 6f is a SEM image of the long and straight fibrin fibres left on the CNF surface after the PPP droplet touch-lift test;

Figure 7 comprises Figures 7a to 7h that show progressive stages in the development and severing of fibrin at the receding surface of a blood droplet during a roll-off test;

Figure 8 shows citrated blood generated fibrin fibres on the superhydrophobic CNF/PTFE and CNF/PDMS surfaces, in which:

Figure 8a shows a 20 μ l water droplet sitting on the superhydrophobic

8 -

CNF/PTFE Ti mesh;

Figure 8b shows fibrin fibres generating at the receding side of a 40 μ l citrated blood droplet on the superhydrophobic CNF Ti mesh:

Figure 8c shows a 20 μ l water droplet sitting on the superhydrophobic
5 CNF/PDMS Ti surface; and

Figure 8d shows fibrin fibres generating at the receding side of a 20 μ l citrated blood droplet on the superhydrophobic CNF/PDMS Ti surface;

Figure 9 shows an absence of fibrin fibres generating on the hydrophobic CNF
10 surfaces, and particularly:

Figure 9a is an image showing no fibre observed for 20 μ l EDTA blood or PPP droplets on the HP #1 sample (spray-coated with only PTFE);

Figure 9b is an image showing no fibre observed for 20 μ l blood or PPP droplets on the HP #2 sample (spray-coated with 0.2 wt% CNF in PTFE);

15 Figure 9c shows SEM images of the HP #1 surface before and after PPP sliding, showing no straight fibrin fibres on the surface – scale bar is 10 μ m;

Figure 9d shows SEM images of the HP #2 surface before and after PPP sliding – scale bar is 10 μ m;

20 Figure 10 shows fibrin confirmation, in which:

Figure 10a shows a fibrin ELISA test confirming the existence of fibrin on the superhydrophobic CNF surfaces after a sliding test by plasma (n, being in each case the number of measurements or repetitions, is 3);

Figure 10b shows blood and PPP droplets with anti-thrombin (20 μ l)
25 quickly rolled down the superhydrophobic CNF/PTFE Ti surface at a small tilt angle without generating fibrin fibres at the receding side;

Figure 10c illustrates the actual and nominal receding angles for blood or PPP droplets sliding down the superhydrophobic CNF surface with and without fibrin fibre generation; and

30 Figure 10d shows the dynamic contact angles for water droplets and blood or PPP droplets with and without anti-thrombin;

Figure 11 shows that no fibrin fibre was generated on the superhydrophobic CNF/PTFE Ti surface, for blood/PPP with anti-thrombin, and particularly:

Figure 11a schematically illustrates fibrin inhibition by anti-thrombin,
5 argatroban;

Figures 11b and c are SEM images of the CNF surface after sliding by anti-thrombin blood and PPP, respectively, showing no aligned/straight fibrin fibre;

Figure 11d shows no fibrin fibre was detected when a blood drop with
10 anti-thrombin came into contact and was subsequently lifted from the superhydrophobic CNF surface;

Figure 11e shows that swiping the CNF surface by PPP with anti-thrombin generated no fibrin fibres; and

Figure 11f is a SEM image of the CNF surface after PPP swiping test,
15 showing no aligned/straight fibrin fibre;

Figure 12 shows angle measurements during testing, schematically and photographically, and the related results. In particular:

Figure 12a shows that, due to the droplet protrusion, the nominal
20 receding angle θ_{r_nom} is different from the actual receding angle θ_r for blood or PPP droplets on the superhydrophobic CNF surfaces;

Figure 12b shows θ_{r_nom} , measured at the moment before catastrophic fibrin fibre fracture and droplet rolling down;

Figure 12c shows the liquid-air interface;

Figure 12d shows the smoothed liquid-air or liquid-substrate interface;
25 and

Figures 12d to 12f show the position of D confirmed at the point on the circle where the pixel grey value had an abrupt change;

30 Figure 13 illustrates anti-bacterial, rapid coagulation and non-wetting features, and includes:

10 -

Figure 13a shows a fibrous fibrin meshwork formed on the superhydrophobic CNF/PTFE Ti surface;

Figure 13b illustrates an anti-bacterial capability, by comparing a substrate coated with the present coating side-by-side with an uncoated
5 substrate;

Figure 13c provides SEM images of the normal gauze and the CNF gauze;

Figure 13d shows a relative haemoglobin absorbance $RHA(t)$ plot, showing the fast clotting performance of the CNF gauze; and

Figure 13e schematically illustrates clotting without blood loss ($n = 3$);
10

Figure 14 shows a normal cotton gauze and, in particular:

Figure 14a provides optical images of the uncoated gauze;

Figure 14b provides SEM images of the uncoated gauze; and

Figure 14c provides SEM images of the gauze with clot, the clot and gauze
15 being a solid piece;

Figure 15 illustrates *in vitro* clotting tests. In particular:

Figure 15a shows a clotting test by sandwiching 20 μ l blood between two gauze samples (size: 15 mm by 15 mm);

Figure 15b shows the nominal blood contact area of 20 μ l blood on the uncoated white cotton gauze of Figure 14a and a superhydrophobic CNF-coated gauze;
20

Figure 15c shows clotting without blood loss;

Figure 15d is a measurement of the maximum hydrostatic pressure that one layer of the CNF gauze (without an impervious membrane) can withstand before blood leakage; and
25

Figure 15e shows non-wetting of the CNF gauze by blood under high pressure;

30 Figure 16 illustrates facile clot detachment, in which:

Figure 16a – illustrates of fibrin fibres formation from blood on a superhydrophobic CNF surface;

Figure 16b shows progress during clot maturation;

Figure 16c is a SEM image of the micro-fibrin fibres adhered on the CNFs-
5 coated cotton fibre after clot shrinkage (see also Figures 19a and 19b) – scale
bar 25 μm ;

Figure 16d provides SEM images showing CNFs transferred onto clot
after clot detachment, resulting in a smooth cotton fibre, and a hairy clot
surface (see also Figure 19c) – scale bar 50 μm ;

10 Figure 16e shows a clot self-peeling from a stiff CNF Ti mesh surface
(see also Figures 18d to 18f); and

Figure 16f shows facile clot detachment;

Figure 17 show facile clot detachment, in which:

15 Figure 17a is a SEM image showing the aftermath of the clot contraction
on the CNF gauze – scale bars are 50 μm ;

Figure 17b show some micro fibrin fibres remaining adhered on the CNF-
coated cotton fibre after clot shrinkage – scale bars are 50 μm ; and

Figure 17c provides SEM images of the clot surface in contact with the
20 CNF gauze after clot detachment – scale bars are 50 μm ;

Figure 18 shows clot self-detachment on the stiff CNF/PTFE Ti mesh, in which:

Figure 18a is a SEM of the pristine Ti mesh #60 before CNF coating;

25 Figures 18b and 18c are SEM images of the Ti mesh after coating with
superhydrophobic CNF (thickness of the CNF layer was approximately 20 μm);

Figure 18d shows clot contraction on the CNF Ti mesh;

Figure 18e shows clot self-detached from the CNF Ti mesh; and

Figure 18f is a photograph of the hairy area on the clot, which was in
contact with CNF;

30

Figure 19 illustrates *in vivo* animal experiments, in which:

Figure 19a shows plaster-like gauzes were patched onto incisions on the back of a rat (see also Figure 21a to 21c);

Figure 19b shows peeling of the gauze at 3 min to measure the blood loss;

5 Figure 19c shows peeling of the gauze at about 2 h to measure the peeling force (Figures 21a to 21i);

Figure 19d shows that the CNF gauze minimized blood loss ($n = 6$) – shown as mean $\pm$ SD, the error bar representing SD and individual data points in d and e are represented by black stars;

10 Figure 19e confirms the peeling force for the CNF gauze was significantly smaller than that for the normal gauze ($n = 5$) – shown as mean $\pm$ SD, the error bar representing SD and individual data points in d and e are represented by black indicia;

Figure 19f provides SEM images of the area in contact with blood on the CNF gauze in Figure 19c – CNF residuals after clot detachment were observed on cotton fibres - scale bar is 1119 μm ;

Figure 19g provides SEM images of the area not in contact with blood on the CNF gauze in Figure 19c, where cotton fibres were densely coated with CNFs - scale bar is 15 μm ;

20 Figure 19h provides SEM images of the peeled normal gauze in Figure 19c - scale bar is 100 μm ; and

Figure 19i is a schematic of the haemostatic CNF gauze/plaster for wound treatment;

25 Figure 20 shows the comparison of the clot peeling tension of the CNF gauze with commercial products, in which:

Figure 20a shows the commercial products used for comparison;

Figure 20b shows the procedures used to measure the clot peeling force; and

30 Figure 20c summarizes the water contact angle and the clot peeling tension for different samples;

Figure 21 relates to *in vivo* experimentation, in which:

Figure 21a shows the design of the plaster-like gauze;

Figure 21b is a photograph of the prepared plaster-like CNF gauze and
5 the control normal gauze;

Figure 21c is a schematic of the application of gauze onto the wound
made on the back of an anaesthetised and shaved rat;

Figure 21d shows the CNF gauze and the control gauze applied onto the
wounds on back of the rat;

10 Figure 21e provides a comparison between the CNF gauze and the normal
gauze peeled at 3 minutes;

Figure 21f shows the CNF gauze after peeling;

Figure 21g schematically illustrates the gauze and control after trimming
away the adhesive film around the gauze;

15 Figure 21h illustrates the peeling force measurement process; and

Figure 21i provides two typical peeling force versus time curves for the
normal gauze and the CNF gauze, and the maximum peeling force was used for
comparison ($n=5$);

20 Figure 22 provides measurement of the clot peeling tension, in which:

Figure 22a shows a clot formed by dispensing 20 μ l blood onto the CNF
gauze surface;

Figure 22b shows a clot formed between two normal gauzes by dispensing
20 μ l blood;

25 Figure 22c shows a typical clot peeling force curve; and

Figure 22d provides a measurement of the clot maximum width W for the
calculation of the clot peeling tension $F_{\max}/W$;

Figure 23 shows the effects of CNF weight percentage on hydrophobicity and
30 the fibrin fiber generation capability, in which:

Figure 23a shows the CNF/PDMS composite surface; and

14 -

Figure 23b shows a CNF/beeswax composite surface;

Figure 24 shows the comparison between the present material and the hydrophobically modified chitosan from four perspectives: (1) the time to stop
5 bleeding, (2) the time to form a clot, (3) blood loss and (4) the clot peeling force.

Figure 25 relates to *in vivo* skin compatibility testing, in which:

Figure 25a shows the design of the prepared sample for test;

Figure 25b shows four pieces of prepared samples were attached onto rat
10 skin with hairs shaved; and

Figure 25c shows the haemostatic surface is skin compatible and does not cause skin irritation after contact with skin for 12 hours; and

Detailed description

15

Described with reference to the drawings are a hydrophobic surface, a haemostatic dispersion and a haemostatic device for achieving haemostasis. Some embodiments of these surfaces, dispersions and devices are directed to achieving both rapid blood coagulation before significant blood loss, and
20 subsequent removal of facile wound-dressings without clot tears and secondary bleeding. These objectives are achieved via a superhydrophobic (SHP) surface with immobilized carbon nanofibers (CNFs).

CNFs are shown herein to promote quick fibrin growth and cause rapid clotting.
25 Moreover, due to their superhydrophobic nature they severely limit blood wetting to prevent blood loss and drastically reduce bacteria attachment. Furthermore, minimal contact between the clot and the superhydrophobic CNF surface yields an unforced clot detachment after clot shrinkage.

30 These important attributes are verified *in vitro* and *in vivo* with rat experiments. The work set out below thereby demonstrates that this

strategy for designing haemostatic patch materials has great potential.

A haemostatic device 100 in accordance with present teachings, is shown in Figure 1. The haemostatic device 100 includes a substrate 102 and a surface 104 formed on the substrate 102. The surface 104 includes materials 106 that are partially embedded in a base 108. The base 108 may be hydrophobic, though this is not essential since hydrophobic nanofibres will confer the necessary hydrophobicity on the device 100.

10 This partial embedding results in part of the materials 106 protruding from the base 108. The materials referred to herein may also be referred to as structures, if permitted by context. The protruding ends can give the surface 104 a random surface morphology. With appropriate formation of the surface 104 using micro-sized and/or nano-sized materials in accordance with the method below,

15 random morphology can be used so the surface 104 can obtain a high water contact angle (WCA) – for example, some embodiments may achieve a WCA of at least 130°, where other embodiments may achieve a WCA lower than 130°. The random surface morphology will typically be one of micro-roughness and nano-roughness. The roughness will largely depend on the size of the materials

20 106 embedded in the base 108.

The surface morphology results in the creation of air pockets that entrain air when a liquid comes into contact with the surface 104. This entrainment of air prevents the liquid from wetting down the materials 106, and onto the base 108 or substrate 102.

The materials 106 are each a micro-sized or nano-sized material, through which the surface 104 substantially prevents wetting of the substrate 102. The materials 106, presently nanofibers or predominantly nanofibers, are hydrophobic. The materials may be formed from carbon or a non-carbon material such as polymeric fibre, silicon fibre or coated cellulose fibre.

To avoid displacement of the materials, which might otherwise create areas on the substrate 102 to which liquids may attach, the hydrophobic nanofibers 106 are immobilized in the base 108. This also largely prevents the materials from
5 dislodging during use.

With reference to Figure 2, the surface 104 can be formed by depositing a haemostatic coating dispersion 200 on the substrate 102. The haemostatic coating dispersion 200 includes micro-sized or nano-sized materials 202 and a
10 base 204. The present surface 104 is a superhydrophobic carbon nanofiber (CNF) coating. In an experiment, CNFs of 98% purity, 100nm diameter, 20–200 μm length, were pre-mixed with a biocompatible polymer capable of immobilising the CNFs, presently being polytetrafluoroethylene (PTFE) powder (1 μm particle size). In some embodiments, the CNF diameter may vary
15 (e.g. between about 5nm to 1,000nm) and the length may similarly vary as needed (e.g. between about 5 μm to 500 μm). In other embodiments, a micro-sized material may be used having, for example, a diameter from about 1 μm to 200 μm . Polydimethylsiloxane (PDMS) was also trialled in place of PTFE, and other compounds such as beeswax, non-hydrophobic compounds and combinations
20 thereof can provide similar properties. In accordance with embodiments herein, the case may be any compound capable of fixing the materials (e.g. nanofibres) onto the substrate in a manner that creates air pockets when the materials come into contact with fluids such as blood and PPP. The dispersion – i.e. materials 202 and base 204 – were then applied on substrates via spray coating 206.

25

The substrates used were a flat Titanium (Ti) substrate (Ti6Al4V, 1 mm thick) though, in other embodiments, transparent glass slides (25 mm by 75 mm, 1.1 mm thick) may be appropriate. The Ti forms a mesh having a nominal aperture of 250 μm , and wire diameter of 125 μm . (see Figure 18a, showing a SEM image
30 of the pristine Ti mesh #60 before CNF coating, with a nominal aperture of about 250 μm and a wire diameter of about 125 μm). These substrates were

ultrasonically cleaned (10 min) with acetone and isopropanol, and were further cleaned with oxygen plasma. Commercially available cotton woven gauze 208 (see Figures 16a and 16b) was also used as substrate – other substrates, such as fabrics, may also be used to, for example, create a flexible patch for
5 conforming to various bodily surface.

Before spray coating, a CNF/PTFE or CNF/PDMS composite dispersion in dichloromethane was first prepared by mixing the CNF and dichloromethane dispersion with the PTFE and dichloromethane or PDMS and dichloromethane
10 dispersion. The CNF and dichloromethane dispersion was prepared by dispersing CNFs in dichloromethane with a probe ultrasonicator for 1 min. The PTFE and dichloromethane or PDMS and dichloromethane dispersion was prepared by dispersing the PTFE powder or PDMS (pre-polymer to cross linker weight ratio 9:1) in dichloromethane under ultrasonication for 20 min. The CNF
15 and dichloromethane dispersion was then mixed with the PTFE and dichloromethane or PDMS and dichloromethane dispersion, and sonicated for 10 min to prepare the CNF/PTFE or CNF/PDMS composite dispersion in dichloromethane. The composite dispersion was spray-coated onto sample substrates at a pressure of 430 kPa. The spray-coated CNF/PTFE sample
20 was baked for 30 min at 400 °C in a low oxygen environment to prevent oxidization. The CNF/PDMS sample was baked at 80 °C for 1 h. As the cotton could not withstand the high temperature (400 °C) for PTFE coating, the cotton gauze was coated with CNF/PDMS instead. The CNF/PTFE nanocomposite had a CNF to PTFE weight ratio of 1:9. The CNF/PDMS nanocomposite had a CNF to
25 PDMS weight ratio of 1:2, whereas a higher concentration of PDMS would cause CNF agglomeration and result in significant coverage of the exposed CNF surface by PDMS. This would lead to roll-off angles (RAs) larger than 10°.

The effect of CNF concentration on superhydrophobicity was also investigated
30 by spray coating low-concentration CNF on Ti substrates (HP #1: only PTFE; HP #2: 0.2 wt% CNF in PTFE) following the same protocol. This allowed a study

of the weight ratios of binding/polymer material and their effect on hydrophobicity and clot-promotion. Hydrophobicity is characterized by the water contact angle ($\geq 150^\circ$). Clot-promoting capability is tested by fiber fibrin generation at the receding side of the blood droplet as it is sliding/rolling on the
5 CNF composite surface.

This investigation was conducted using two binding polymers/materials, PDMS and beeswax – results are shown in Figure 23, comprising Figures 23a and 23b. Each graph is divided into three regions, (A), (B) and (C). Region (A) starts from
10 the left-hand side and continues to $WP_{\min}$, being the minimum desirable weight percentage ratio of PDMS or beeswax to CNFs. In Region (A) the CNFs were free or not sufficiently immobilized. CNFs were considered sufficiently immobilized for weight percentage ratios of PDMS or beeswax to CNFs, of at least $WP_{\min}$. In region (B), extending from $WP_{\min}$ to $WP_{\max}$, CNFs were sufficiently immobilized
15 and there was fibrin fibre generation. In Figure 23(a), region (C) showed substantially no fibrin fibre generation. In Figure 23(b), region (C) initially showed some fibrin fibre generation to point (D) but with free CNF and beeswax agglomerates, and thereafter fibrin fibre generation was inhibited. Accordingly, this investigation found the weight percentage of the binding polymer/material
20 should be in a given range (that is, not smaller than a minimal value $WP_{\min}$, and not larger than a maximum value $WP_{\max}$) to make the CNF composite surface both superhydrophobic and fibrin fibre generation-promoting. When the weight percentage is smaller than $WP_{\min}$, CNFs are not sufficiently immobilized on the surface. Free CNFs will be observed floating on the water droplet, if one droplet
25 is dispensed on the surface. When the weight percentage of the binding polymer/material is larger $WP_{\max}$, the CNF composite surface will either not be superhydrophobic, be unable to promote fibrin fiber generation, or have some loosely attached CNF and binding material/polymer agglomerates.

30 As for the CNF/PDMS composites, the PDMS to CNF weight ratio should be in the range of 4:1 to 30:1, as shown in Figure 23a. When the PDMS/CNF weight

ratio is smaller than 4:1, free CNFs will be observed on the surface. When the PDMS/CNF weight ratio is larger than 30:1, the surface will not be superhydrophobic (contact angle smaller than 150°), will have strong adhesion to blood, and no fibrin fibre will generated at the receding side of the blood
5 droplet.

As for the CNF/beeswax composites, the beeswax to CNF weight ratio should be in the range of 8:1 to 150:1. When the beeswax/CNF weight ratio is smaller than 8:1, free CNFs will be observed on the surface. When the beeswax/CNF
10 weight ratio is higher than 150:1, the surface is still superhydrophobic, partially due to self-agglomeration of beeswax in the spray coating process (micro-beeswax particles could be observed on the spray coated beeswax surface), and also promotes fibrin fibre generation (due to the presence of CNF fibre). However, at this high beeswax weight ratio, spray-coating will generate some
15 free CNF & beeswax agglomerate particles (that are loosely attached on the surface); these loose particles can be observed when a blood droplet is dispensed on the surface (Figure 23b). Thus a high concentration of beeswax for CNF immobilization is not beneficial.

20 In Figure 23 dichloromethane (CH_2Cl_2) was used as the carrier medium to dissolve/disperse PDMS, beeswax, and CNF. The carrier medium is used to evenly mix the binding polymer/material with the CNF. After spray coating, the carrier medium will evaporate. The carrier medium can be a single solvent, a mixture with a few different solvents, or a mixture with water. Solvents such as
25 acetone, ethanol, toluene, hexane, acid solution, alkaline solution, and water that can dissolve/disperse the binding polymer/material and the CNF can also be used.

Relatedly, the results of the investigation into the effect on hydrophobicity and
30 surface morphology of different carbon nanofibre (CNF) nano-composite coatings is shown in Figure 3 – particularly Figures 3a and 3b. Figure 3a shows

20 -

the water contact angle, roll-off angle, and blood contact angle on different superhydrophobic surfaces. Data in the bar graph are shown as mean $\pm$ SD, and the error bar represents SD (the number of measurements n is 5). Figure 3b shows SEM images of the hydrophobic surfaces HP #1 (301) and HP #2 (303). HP #1 refers to a hydrophobic Ti surface coated with only PTFE that has a water contact angle of $130.8 \pm 2.1^\circ$. HP #2 refers to a Ti surface coated with 0.2 wt% CNF in PTFE that has a water contact angle of $143.6 \pm 3.1^\circ$.

Figure 3c provides SEM images of superhydrophobic nano-composite CNF surfaces. The CNF/PTFE surface had a surface morphology similar to the CNF/PDMS surface. On the superhydrophobic CNF/PTFE and CNF/PDMS Ti surfaces, blood had a smaller contact angle than water (but still larger than 150°) due to the lower surface tension of blood. On the superhydrophobic CNF/PTFE Ti mesh surface, the contact angle could not be accurately measured due to the special mesh topography. On the hydrophobic surfaces (HP #1 and HP #2), the reduced surface roughness led to a smaller water contact angle.

To remove any loosely attached CNFs, samples were prepared and exposed to compressed N₂ gas from a spray gun (nozzle diameter: 0.8 mm; pressure: 430 kPa; sample to nozzle distance ~ 15 cm) before testing to make sure there was no free CNF on the present material.

The result is a hydrophobic surface formed using a haemostatic dispersion. That hydrophobic surface may be used in a device such as device 100 shown in Figure 1 – the device 100 being, for example, a wound dressing, catheter-based stent, coil or graft, formed from a haemostatic dispersion.

Devices, in the form of wound dressings, formed according to the method described above were tested for fibrin formation, anti-bacterial properties, clotting and facile clot detachment. Fibrin 1102 is converted from fibrinogen 1100. Fibrinogen is a glycoprotein complex, made in the liver that circulates in

21 -

the blood. During tissue and vascular injury, fibrinogen is converted enzymatically by thrombin 1104 to fibrin and then to a fibrin-based blood clot. Fibrin clots function primarily to occlude blood vessels to stop bleeding. Fibrin formation is therefore critical in wound healing and promoting its formation can
 5 reduce bleeding.

Referring firstly to results relating to fibrin fibre formation on superhydrophobic CNF surfaces: for effective haemostatic performance, the present SHP surface is designed first to be strongly blood-repellent and second to be capable of
 10 triggering fast coagulation upon blood contact. Being extremely blood repellent is achieved by spray coating a nanocomposite dispersion, such as that described above consisting of CNFs (diameter: 100 nm, length: 20–200 μ m), and polytetrafluoroethylene (PTFE) or polydimethylsiloxane (PDMS), onto a substrate as shown in Figure 3. Examples of the bases used are set out in Table 1.

15

Coating abbreviation	Substrate	Polymer matrix	CNF to polymer weight ratio
CNF/PTFE Ti	Ti plate	PTFE	1:9
CNF Ti mesh	Ti mesh #60	PTFE	1:9
CNF/PDMS Ti	Ti plate	PDMS	1:2
CNF gauze	Cotton gauze	PDMS	1:2

Table 1 Superhydrophobic nanocomposite coatings

The surface generated by the above process has a dense layer of CNF network with micro/nano-roughness that is partially embedded in a hydrophobic
 20 polymer matrix (PTFE or PDMS). A photograph of the surface 300 is shown in Figure 4a, with an inset photograph of a water droplet 302, showing a water contact angle of 162.5°, and air plastron (air pocket) region being indicated by reference numeral 304. Photographs of relevant actual experimental results are shown in Figure 3c.

25

The use of hydrophobic base components, the micro/nano-scale topography from spray coating, and the morphology of the nanofibers collaboratively result in superhydrophobicity. The CNF/PTFE Ti surface had a water contact angle (WCA) of $162.1 \pm 2.9^\circ$ (mean $\pm$ SD) and a water roll-off angle (WRA) of about
5 1° , and the CNF/PDMS Ti surface had a WCA of $154.9 \pm 0.6^\circ$ and a WRA of about 4° - see Figure 3a.

The effect of CNF concentration on surface wetting was also investigated as shown in Figure 3. Compared with water, blood has a smaller surface tension
10 of 58 mN m^{-1} (72 mN m^{-1} for water) and a complex composition. However, the present SHP CNF surfaces could still repel it, with blood contact angles (CAs) of $153.6 \pm 1.4^\circ$ for the CNF/PTFE Ti surface and $151.4 \pm 1.8^\circ$ for the CNF/PDMS Ti surface (see Figure 3a).

15 Very interestingly and unexpectedly, it was observed that when blood or platelet poor plasma (PPP) with anticoagulant ethylenediaminetetraacetic acid (EDTA) or sodium citrate was brought into contact with the present SHP CNF surfaces, long straight fibres (later confirmed to be fibrin) formed rapidly. This was tested using the setup shown in Figure 5, for visualizing fibrin fibres
20 generation at the receding side of the blood/plasma droplet sliding down a superhydrophobic CNF surface. In order to visualize fibrin fibres during blood or PPP sliding, backlighting in a typical goniometer was avoided due to over-exposure, and a light source was projected at a proper angle from the side to make fibrin fibres visible. The setup 500 was used in conjunction with a
25 blood/plasma droplet 502 on a SHP CNF surface 504 formed in accordance with present teachings. A light-emitting diode (LED) torch 506 emitted light 507 at a desired angle, onto the CNF surface 504. The CNF surface 504 was positioned on a stage assembly comprising a rotatable stage 508 and a translatable stage 510. Angles and fibrin formation, particularly during sliding tests, were
30 observed by photographs captured using camera 512.

In the sliding test shown in Figure 4b, abundant long fibres 306 were generating at the receding side of the plasma droplet 308 as it rolled in direction X down surface 310. These fibres 412 pulled the droplet 414 as shown in Figure 4d, retarding its sliding motion, until a critical angle was reached, causing catastrophic fibre fracture and allowing the droplet to roll off quickly. After the first one or two fibres started to rupture, the remaining fibres could not hold the droplet weight and their rupturing occurred in the form of a rapid domino effect. After the PPP droplet rolled off, visible fibre footprints 416 were left on the surface 418 (Figure 4e), in the form of straight and ordered fibres 600 on top of random CNFs 602 (Figure 6e), aligning in the droplet rolling direction 604. The same was observed after blood droplet sliding (Figure 6d). Similar observations were made in a slightly different test, the touch-lift test, where a PPP or blood droplet was brought onto the CNF surface for a brief contact and then retracted (contact duration ~ 3 s; Figure 4c), and fibrin fibre generation was also observed. Long and straight fibres (up to 300 μm) projecting outward from the contact centre were detected under scanning electron microscopy (see SEM image Figure 6f).

Figure 6 provides images of fibrin fibres generating on the superhydrophobic CNF/PTFE Ti surface after EDTA blood/plasma sliding tests. Figure 6a illustrates blood/plasma sliding. Figures 6b and 6c are optical images showing fibrin fibres left on the surface after PPP or blood sliding, and visible fibrin fibre footprints were left on the CNF surface after sliding tests. Figure 6d provides SEM images of the CNF surface after blood sliding. Long, straight and aligned fibrin fibres can be observed on top of random CNFs. Arrows are used to highlight the aligned fibrin fibres. Figure 6e is a SEM image of the long, straight and aligned fibrin fibres left on the CNF surface after the PPP droplet rolled down the superhydrophobic CNF surface. Figure 6f is a SEM image of the long and straight fibrin fibres left on the CNF surface after the PPP droplet touch-lift test.

Fibrin fibre generation was observed despite the use of EDTA or sodium citrate

(with anti-coagulation properties), and occurred on both the SHP CNF/PTFE and CNF/PDMS surfaces (Figures 4, 7 and 8), but it was not observed on hydrophobic surfaces with a low CNF concentration or no CNF (Figure 13). Figure 7, comprising Figures 7a to 7h, shows progressive fibrin fibre generation at the receding side of a 20 μ l EDTA blood droplet sliding down a superhydrophobic CNF/PTFE Ti surface. A similar fibrin fibre generation phenomenon was observed for EDTA or citrated blood/PPP droplets sliding on the CNF/PTFE surface. Figure 8 shows citrated blood generated fibrin fibres 800 on the superhydrophobic CNF/PTFE and CNF/PDMS surfaces 802. Figure 8a shows a 20 μ l water droplet 804 sitting on the superhydrophobic CNF/PTFE Ti mesh 806. Figure 8b shows fibrin fibres 800 generating at the receding side of a 40 μ l citrated blood droplet 808 on the superhydrophobic CNF Ti mesh. Figure 8c shows a 20 μ l water droplet 804 sitting on the superhydrophobic CNF/PDMS Ti surface 810, and Figure 8d shows fibrin fibres 812 generating at the receding side of a 20 μ l citrated blood droplet 814 on the superhydrophobic CNF/PDMS Ti surface 816. Fibrin fibres were also observed for PPP with EDTA on this CNF/PDMS Ti surface.

A similar fibrin fibre generation phenomenon was observed for all EDTA or citrated blood/PPP droplets sliding on the CNF/PTFE Ti mesh and the CNF/PDMS Ti surfaces. Figure 9 shows that no fibrin fibres were generated on the hydrophobic CNF surfaces. Figure 9a shows no fibre observed for 20 μ l EDTA blood or PPP droplets on the HP #1 sample, spray-coated with only PTFE. Figure 9b shows that no fibre was observed for 20 μ l blood or PPP droplets on the HP #2 sample, spray-coated with 0.2 wt% CNF in PTFE. Figure 9c provides SEM images of the HP #1 surface before and after PPP sliding, showing no straight fibrin fibres on the surface. Instead, the surface became bio-fouled by plasma. Similar results were observed on this surface after blood sliding. Figure 9d provides SEM images of the HP #2 surface before and after PPP sliding, showing no straight fibrin fibres on the surface. The surface also became bio-fouled by plasma. Similar results were observed on this surface after blood sliding.

25 -

This suggested that the nano-engineered SHP CNF surface promoted fibre formation and the choice of PTFE or PDMS for CNF immobilization would not affect the fibre generation.

5 Fibrin presence was then confirmed. To confirm that these fibres were indeed fibrin, verification tests were performed. Existence of fibrin, on both the SHP CNF/PTFE and CNF/PDMS surfaces after PPP sliding, was verified through porcine fibrin enzyme-linked immunosorbent assay (ELISA) tests. A standard dilute solution from the ELISA kit was flushed over the CNF surface with and
10 without PPP sliding to wash the generated fibrin fibres into the reaction well for a qualitative analysis. The positive results in Fig. 10a confirmed the fibres to be fibrin (data in Figures 10a and 10d are shown as mean $\pm$ standard deviation (SD), the error bar represents SD, and individual data points in Figure 10a are represented by black stars).

15

To further support this result, it was found that adding anti-thrombin to the blood/plasma could prevent fibre formation (see Figure 11a). As thrombin is a key factor for fibrin formation, thrombin inhibition is a potent way to prevent fibrin generation. A high dose (2 mg ml^{-1}) of a thrombin inhibitor, argatroban,
20 was used in EDTA or citrated blood and PPP, and it was found that both blood and PPP droplets ($20 \mu\text{l}$) quickly rolled off the SHP CNF surface at a very small angle without generating any fibres (Figure 10b). SEM imaging further confirmed the absence of aligned long fibrin fibres on the surface after sliding by anti-thrombin containing blood or plasma (Figures 11b and 11c). Similarly,
25 fibrin formation was not detected in the touch-lift test with such anti-thrombin treatment (Figures 11d to 11f). These findings collaboratively confirmed the identity of the fibres as fibrin.

With reference to the particular images in Figure 11, no fibrin fibre was observed
30 as generating on the superhydrophobic CNF/PTFE Ti surface, for blood/PPP with anti-thrombin. Figure 11a shows fibrin inhibition by anti-thrombin, argatroban.

Figures 11b and 11c are SEM images of the CNF surface after sliding by anti-thrombin blood and PPP, respectively, showing no aligned/straight fibrin fibre. Figure 11d shows that no fibrin fibre was detected when a blood drop with anti-thrombin came into contact and was subsequently lifted from the
 5 superhydrophobic CNF surface. Figure 11e shows swiping the CNF surface by PPP with anti-thrombin generated no fibrin fibres. Figure 11f is a SEM image of the CNF surface after PPP swiping test, showing no aligned/straight fibrin fibre.

Fibrin fibres generated at the receding side (Figure 4d) greatly affected the
 10 sliding dynamics of blood/plasma droplets on the present SHP CNF surfaces. First, fibrin fibres behaved similar to micro-strings pulling on the blood/plasma droplets, retarding the rolling-off motion and increasing the adhesion of blood/plasma droplets on the surface (Figure 10c, in which θ_a is the advancing angle, θ_r is the actual receding angle, and θ_{r_nom} is the nominal receding
 15 angle.

Measurement of θ_{r_nom} is explained with references to the images in Figure 12, showing the nominal receding angle. In Figure 12a, the nominal receding angle θ_{r_nom} is different from the actual receding angle θ_r for blood or PPP
 20 droplets on the superhydrophobic CNF surfaces, due to the droplet protrusion. In Figure 12b, θ_{r_nom} , measured at the moment before catastrophic fibrin fibre fracture and droplet rolling down, is shown as the angle between line AB and line CD. AB is at the droplet-substrate interface and CD is at the straight portion of the droplet-air interface above the droplet protrusion. Figure 12c shows the
 25 liquid-air interface 1200, that is blurry on the high-resolution (1920×1080 pixels) black and white image. Figure 12d shows the liquid-air or liquid-substrate interface smoothed via pixel intensity averaging and reducing pixel resolution image to 480×270 pixels. This smoothed image made it easier to localize lines AB and CD. A reference circle, with a radius r of 30 pixels, was
 30 drawn to intersect the straight portion of the liquid-air interface at two points C and D. Figures 12e and 12f show the position of D being further confirmed at

the point on the circle where the pixel grey value had an abrupt change. The position of C was confirmed similarly. Line AB on the liquid-solid interface 1202 was determined using the same method in Figures 12d to 12f. These procedures were performed manually and had sufficient accuracy, with errors being in the
 5 order of a pixel width, which translated to an angle error of about 0.95° (equal to $\arctan(1/60)$).

The roll-off angle (RA) for blood and PPP droplets (mean $\pm$ SD) were $28.1 \pm 11.6^\circ$ and $15.0 \pm 6.6^\circ$, respectively - Figure 10d, showing the dynamic contact
 10 angles for water droplets and blood or PPP droplets with and without anti-thrombin on the superhydrophobic CNF/PTFE Ti surface ($n = 4$), contact-angle hysteresis $CAH = \theta_a - \theta_r$, and nominal contact-angle hysteresis $CAH_{nom} = \theta_a - \theta_{r_nom}$. The droplet volume is 20 μ l, Blood-Anti or PPP-Anti, respectively, represents blood or PPP with anti-thrombin. When fibrin generation was
 15 inhibited by anti-thrombin, blood and PPP droplets could roll off at a smaller RA of $4.3 \pm 1.4^\circ$ and $2.6 \pm 1.3^\circ$, respectively. Due to the dose effect of anti-thrombin, some micro-fibrin fibres may still exist at the liquid-solid interface, affecting the receding CA and leading to a large CAH ($CAH > RA$) for blood/PPP droplets with anti-thrombin (see Figure 10d). Second, the blood or plasma droplet formed a
 20 protrusion at its receding region near the surface under the tugging force from fibrin fibres (see Figure 4d) as shown in progressive images Figures 7g to 7h, generating a nominal receding angle θ_{r_nom} that was smaller than the actual receding angle θ_r (Figure 10c). Consequently, the droplet demonstrated a large nominal CA hysteresis CAH_{nom} ($CAH_{nom} = \theta_a - \theta_{r_nom}$, actual $CAH = \theta_a - \theta_r$ and
 25 θ_a is advancing in angle). This is not customarily observed on SHP surfaces, which typically have low RAs corresponding to low contact-angle hysteresis.

Regarding the generation mechanism, fibrin fibres were initiated upon the
 30 blood/plasma contact with CNFs. Exposure to CNFs would trigger an extrinsic coagulation cascade reaction, causing the formation of thrombin from

prothrombin, and subsequently converting fibrinogen into fibrin monomer. Fibrin monomers would attach onto CNF and then self-polymerize into long insoluble fibrin fibres, which became visible upon blood/plasma-substrate separation (see Figures 4b and 4c). An interesting and unexpected observation
5 in this study is that fibrin fibres still generated despite the presence of anti-coagulant EDTA or 3.8% sodium citrate (see Figures 4b, 7 and 8), demonstrating its potency in expediting fibrin formation. In addition, CNFs were reported to activate platelets, bind serum albumin and fibrinogen, and activate the serum complement system, which would further trigger coagulation. Also,
10 electric charges concentrated on sharp geometries, such as on the tips of long CNFs, could attract fibrinogen adsorption and possibly promote coagulation.

The capability of CNFs to promote fibrin fibre formation was further demonstrated by cultivating EDTA PPP on the present CNF surface at 37°C for 4
15 minutes, following published methods. A fibrous fibrin meshwork was observed to form (see Figure 13a). Such a fibrous meshwork could be beneficial for hosting activated platelets and blood cells to accelerate coagulation, making the present superhydrophobic CNF surface promising for haemostatic applications.

20 An important aspect of many medical devices, and particularly wound dressings and the like, as that they have anti-bacterial properties or can have those properties imparted to them prior to use. The present SHP CNF surfaces demonstrated excellent anti-bacterial properties. A solution containing *Escherichia coli* (a major infection-causing bacteria) was flushed with green
25 fluorescence protein (GFP) expression plasmid over a glass slide that was half-coated with CNFs and nearly no bacteria was found on the SHP CNF surface under the confocal microscope with a 473 nm laser for GFP excitation. Figure 13b shows that after flushing with fluorescent bacteria solution, nearly no bacteria (dots) were observed to attach on the superhydrophobic CNF surface
30 under the microscope with green fluorescent protein excitation.

The low adhesion of bacteria on the present SHP CNF surface is attributed to the low surface energy hydrophobic materials and the micro/nano- roughness. This excellent anti-bacterial capability will be beneficial, as it helps keep the haemostatic patch sterile and prevent wound infections.

5

In addition to being anti-bacterial, a haemostatic material should promote quick coagulation to minimize blood loss. As a proof-of-concept prototype of using the present material as a wound patch, a normal cotton gauze was coated with SHP CNF as shown in Figure 13c. The normal gauze is
10 superhydrophilic and blood absorbing. The CNF gauze has a dense layer of CNF coating and repels blood. As cotton could not withstand the high annealing temperature (400 °C) for CNF/PTFE coating, CNF/PDMS was used for coating, taking advantage of the low polymerization temperature of PDMS. As verified previously, the CNF/PDMS surface can promote fibrin fibre generation just like
15 the CNF/PTFE surface (see Figure 8d). The cotton gauze, which was initially superhydrophilic and blood absorbing (see Figure 14), became SHP after the CNF/ PDMS coating (Figure 13c). Figure 14 shows normal cotton gauze with Figure 14a providing optical images of the uncoated gauze, Figure 14b being SEM images of the uncoated gauze, and Figure 14c being SEM images of the
20 gauze with clot, the clot and gauze being a solid piece.

The clotting performance of this superhydrophobic CNF gauze was then evaluated. Evaluation via *In vitro* clotting tests is shown in Figure 15. Figure 15a shows a schematic of a clotting test by sandwiching 20 µl blood between
25 1500 two gauze samples 1502, 1504 (size: 15 mm by 15 mm). For the superhydrophobic CNF gauze, blood was contacting the surface coated with CNF. Figure 15b shows the nominal blood contact area 1510 of 20 µl blood on the uncoated white cotton gauze 1506 (Figure 14a) and a superhydrophobic CNF-coated gauze 1508. The blood became sufficiently coagulated after 60
30 minutes, making it easier to separate the two gauzes and to measure the blood contact area. The blood contact contour is highlighted by the white line. Figure

30 -

15c shows clotting without blood loss. The superhydrophobic CNF gauze 1512 was used to seal an opening 1514 (8 mm by 5 mm) on the silicone tube 1516 by medical tape 1518, which was subsequently filled with blood, initiating coagulation by adding CaCl_2 . Figure 15d provides the measurement of the maximum hydrostatic pressure that one layer of the CNF gauze (without an impervious membrane) can withstand before blood leakage. The maximum blood column height h without blood leakage was used to calculate the blood leakage-free pressure. Figure 15e shows non-wetting of the CNF gauze 1520 by blood 1522 under a high pressure of 33 mmHg. The CNF gauze, the periphery of which (not the central region) was glued onto the petri dish 1524, was immersed in the citrated blood, in a pressure chamber with the CNF surface contacting blood. The petri dish acted as an impervious membrane to retain the air plastron across the CNF/blood interface. The hydrostatic pressure exerted onto the CNF surface in blood was controlled by pumping air into the pressure chamber and monitoring the air pressure 1526 using a pressure sensor 1528. A camera 1530 projecting upward was used to detect whether the central part of the CNF gauze was wetted by blood at a given pressure. In comparison, it was experimentally observed that the normal gauze instantly got wetted by blood at ambient pressure. The CNF coating was in contact with blood in these tests.

20

With reference to Figure 15a, 20 μl blood 1508, placed between two pieces of gauze 1510, 1512, was allowed to coagulate for a fixed time. Coagulation was terminated by adding 10 ml deionized (DI) water. Free haemoglobin from red blood cells, not trapped in the clot, was released into water and was assayed. A lower haemoglobin level indicates faster clotting. The CNF gauze was shown to have a lower haemoglobin level and thus faster clotting compared to normal gauze, at 3 minutes. Figure 13d shows the relative haemoglobin absorbance $\text{RHA}(t)$ plot, demonstrating the fast clotting performance of the CNF gauze. As the experiment was performed in a petri dish 1514, clotting in the petri dish 1514 without any gauze was used as control. The absolute haemoglobin absorbance at clotting time t , $\text{HA}(t)$, was measured by the spectrometer at

30

540 nm ($n = 3$), HA(0). The haemoglobin absorbance at $t = 0$ min was used as the reference, and the relative haemoglobin absorbance at clotting time t , RHA(t), equals HA(t)/HA(0). As the haemoglobin came from unclotted red blood cells, a lower haemoglobin absorbance value would mean faster clotting.

5

The non-wetting property of the present SHP CNF coating can prevent blood loss at the wound site, by keeping blood within the wound. This feature was demonstrated in vitro, with a silicone tube filled with blood that had a hole opened on its side to mimic a bleeding wound. Cotton gauzes, with and without SHP CNF coating, were used to cover the holes (see Figure 15c). The SHP CNF gauze achieved clotting without blood loss, whereas the normal cotton gauze experienced severe blood seepage. See Figure 13e, clotting without blood loss over three measurements using the CNF gauze 1300 to seal an opening 1302 on a silicone tube 1304, mimicking a skin wound covered by a gauze. Individual data points are represented by black indicia. The CNF coating was in contact with blood in Figure 13d and Figure 13e. The scale bars are 10 μm in Figure 13a, 50 μm in Figure 13b, and 20 μm in Figure 13c. Data in Figure 13d and Figure 13e are shown as mean $\pm$ SD, and the error bar represents SD. Therefore, owing to the CNF coating's synergetic capability of promoting fibrin formation and minimal wetting (superhydrophobicity), the present material design strategy can achieve fast clotting without blood loss. This performance can be especially beneficial for chronic bleeding disorders.

Furthermore, the air plastron trapped on the SHP CNF surface can be a functional component of the SHP wound patch, as it can help retain the non-wetting feature under high pressure. Without an impervious plastic membrane (see Figure 13e), a single layer of CNF gauze could withstand a pressure of 4.9 ± 0.3 mmHg (mean $\pm$ SD) without blood infiltration (see Figure 15d); with a impervious plastic membrane applied at the back of the gauze like the plaster, the trapped air plastron could prevent the CNF gauze immersed in blood from

30

32 -

wetting even at 300 mmHg (Figure 15e), which was significantly larger than a lotus leaf's non-wetting pressure (about 100 mmHg).

Another unexpected and rather inherent feature of the SHP CNF patch is that the formed clot can easily detach from the CNF gauze by itself upon clot maturation. This is in sharp contrast to existing hydrophilic haemostatic dressing materials – blood will soak and coagulate inside the pores of these materials and the generated clot-dressing composite mixture has strong adhesion to the wound and can be difficult to detach (see Figure 14c). Forced peeling of these hydrophilic haemostatic dressings will tear the wound and cause secondary bleeding, complicating subsequent wound care.

It was presently found that the driving force for easy clot detachment from the present CNF gauze was the contraction of clot as it matured. In the early stage of coagulation, fibrin fibres would form a fibrin meshwork for clot formation. This fibrin generation was initiated on CNFs as illustrated in Figure 6a in which fibrin fibres form from blood on a superhydrophobic CNF surface. Due to its non-wetting property and the presence of micro-air pockets, the CNF surface would only partially attach onto the clot. As such, the clot would have micro-fibrin fibres connected onto the CNFs as shown in Figure 16c and Figure 17. Therefore, a weak connection between the clot and the CNF surface was formed due to the presence of air pockets at the blood–substrate interface – this is called a "Cassie–Baxter" state. During clot maturation, filopodia from platelets would pull and bend fibrin fibres, generating a contractile stress and causing clot contraction – see Figures 18d and 17a (reference for CNF gauze). Figure 17d shows clot contraction on the CNF Ti mesh. The clot border contracted inwards from the initial blood-substrate contact line. The contractile stress in the clot would pull and remove micro-fibrin fibres adhered on CNFs. As shown in Figure 16b, during clot maturation the fibrin mesh would squeeze out the serum and shrink the clot into a smaller size. Contractile stress during clot contraction and

solidification would remove CNFs connected with micro-fibrin fibres, causing the clot to free itself from the superhydrophobic CNF surface after clot solidification and contraction.

- 5 Self-peeling behaviour of the clot was observed on the rigid Ti mesh substrate coated with CNF after clot contraction as shown in Figure 16e, along contraction directions X, X', and self-peel direction Y, and in Figures 18d to 18f, in which Figure 18d shows the initial clot position 1800, and final clot position 1802 after contraction in direction 1804, Figure 18e shows a clot self-detached
10 from the CNF Ti mesh, showing localized contact between blood and the Ti mesh, and in which Figure 18f shows the hairy area on the clot which was in contact with CNF. On the flexible CNF gauze, clot peeling was not visibly self-accomplished, as cotton fibres would be deformed by clot contraction. However, the clot could be easily picked off (see Figure 16f). The average
15 clot-peeling tension for the present SHP CNF gauze was about $1.7 \pm 1.5 \text{ mN mm}^{-1}$ (mean $\pm$ SD). This is about 54 times smaller than that for the normal hydrophilic gauze, which was about $91.3 \pm 19.4 \text{ mN mm}^{-1}$ as shown in Figure 16f over four measurements.
- 20 After clot detachment, CNFs that were initially immobilized on cotton fibres were transferred onto the clot, making the cotton fibres appear smooth – which was drastically different from their original appearance in Figure 5c – and the detached clot surface appear hairy - see also Figure 16d. Figure 17c includes a close-up view of Figure 16d and provides SEM images of the clot
25 surface in contact with the CNF gauze after clot detachment, with localized contact between blood and the CNF gauze. 1700 and 1702 show CNFs transferred onto the clot after clot detachment, resulting in a smooth cotton fibre, and a hairy clot surface. 1704 shows micro fibrin fibres initially attached onto CNFs. This intrinsic clot self-detachment mechanism thus greatly
30 facilitates wound dressing removal, avoiding wound tear and eliminating secondary bleeding.

The clot-peeling force of the present material was also compared with three representative commercial haemostatic products shown in Figure 20a. The Nexcare hydrophilic dressing material from 3 M was used as the substrate material for clotting to occur (Figure 23b). The Nexcare substrate material (25 mm wide and 100 mm long) was adhered onto the stainless base. Citrated blood mixed with 0.2 M CaCl_2 solution at a volume ratio of 10:1 (2000) was then dispensed onto the substrate 2002 to soak it with blood (b1). Haemostatic samples 2004 (30 mm long and 10 mm wide) were immediately placed onto the substrate material 2002, allowing the clot to form between the haemostatic sample 2004 and the substrate material 2002 (b2). Clotting was first allowed to occur at 37 °C for 1 h. Clot solidification was subsequently accelerated by exposing the clotted samples to warm air flow from a hair dryer for 30 min (b3). After that, samples were peeled from one side to measure the peeling force (b4) – see also Figure 23b. Peeling tension was calculated as the maximum peeling force divided by the sample width (repeated for three times; $n = 3$). Besides, WCA was measured with 10 μl DI water ($n = 5$) to provide information on hydrophobicity for different samples (Figure 23c). These three commercial products are marketed as low-adherence or pain-free on removal, but the clot-peeling tension of the present material is 24–52 times smaller than the clot-peeling tension of the commercial haemostatic products, as shown in Figure 20c. Specifically, the clot-peeling tension on the present material is about 52 times smaller than that of one of the representative commercial haemostatic products and about 24 times smaller than that of the other two representative commercial haemostatic products. Therefore, the present material design strategy has brought the clotting peeling force of haemostatic materials to a very low level.

To verify the aforementioned features of the present haemostatic material, *in vivo* experiments were performed on rats with the back-bleeding model – incisions made on the back of rats for gauze application – as shown in Figures

35 -

21c and 21d. The normal gauze was blood absorbing, leaving behind an open wound as referenced by numerals 1900, 1902 in Figures 19a and 19b. For Figure 19a, the plaster-like gauzes were patched onto incisions on the back of a rat – the control cotton gauze got wet quickly, while the CNF gauze prevented
5 blood loss. For Figure 19b the gauze was peeled at 3 min to measure the blood loss. The CNF gauze helped form a gel-like clot, which properly sealed the wound. Under the control gauze, an open wound was observed. In contrast, no blood was observed to seep through the CNF gauze 1904, as shown in Figure 19a. This demonstrates the excellent blood-repelling property
10 of the present devices.

In addition, a darkened gel-like clot was observed under the CNF gauze 706 at 3 min as shown in Figure 19b. This clot properly sealed the wound whereas the wound remained open under the control gauze, demonstrating the ability
15 of the CNF gauze to promote coagulation *in vivo*. The average blood loss for the CNF gauze was 0.3 ± 0.7 mg as shown in Figure 19d (reference 1916), which was about 1.5% of that for the normal gauze at 19.8 ± 9.0 mg as shown in Figure 19d (reference 1918). Blood loss was characterized by the weight increase in gauze at 3 min – see photo in Figure 21e for comparison of the CNF
20 gauze and control gauze at 3 min. This confirms the ability of the CNF gauze to minimize blood loss. Thus, *in vivo* work corroborated the present *in vitro* findings – the blood-repelling CNF gauze could promote coagulation, minimize blood loss, and help achieve a good clot-sealed wound.

25 The force required for gauze removal was also measured *in vivo*. As the wound under the normal gauze 1908 was torn seriously during peeling, it was difficult to accurately measure the gauze-wound contact width. See Figure 19c, showing peeling of the gauze at 3 min to measure the blood loss. The CNF gauze helped form a gel-like clot, which properly sealed the wound.
30 By comparison, under the control gauze, an open wound was observed. The maximum peeling force was used to qualitatively evaluate the performance

of the present CNF gauze (Figure 22i). The average maximum peeling force (mean $\pm$ SD) for the CNF gauze was 7.2 ± 8.6 mN as indicated by reference 1920 in Figure 19e. This measurement was about 43 times smaller than that for the normal gauze, at 315.2 ± 61.3 mN - see Figure 19e, reference 1922.

5 As shown in Figure 19c, reference 1910, the present, easy-to-peel CNF gauze did not tear the wound due to CNF detachment. Many CNFs that were initially coated on cotton fibres as shown in Figure 19g (reference 1924) were removed during clot peeling just like in *in vitro* tests - Figure 19h shows cotton fibres 1926 embedded in a clot. Moreover, gentle stretching did not

10 cause wound tearing (see Figure 19c, reference 1912) or secondary bleeding, whereas peeling the normal gauze tore the wound and caused bleeding - see Figure 21f. In the control group, the clot strongly bound the wound and even skin hairs (see Figure 19c, 1914) to the gauze, forming a stiff clot-gauze-hair concrete. Figure 7h shows SEM images of the peeled normal gauze in Figure

15 19c, showing a clot-gauze-hair concrete, with rat skin hairs imbedded in clot. A hair root is shown in the inset, implying that skin hairs that were stuck in the clot were pulled out from skin during gauze peeling. Forced peeling would tear the wound and cause secondary bleeding, increasing the chance of infection as shown in the top right of Figure 19c and right-hand side of Figure 11f. *In*

20 *vivo* findings therefore confirm the remarkable features of the present wound-dressing materials -. Figure 19i shows nanofiber structured surface 1928, antibacterial surface properties at 1930, blood loss prevented at the wound site 1932, easy peeling in the direction of arrow 1934 without tearing the clot, and fast clotting at the wound 1936 under patch 1938. This can

25 successfully address the serious problems - blood loss and strong clot adhesion - plaguing the application of conventional hydrophilic haemostatic materials.

Studies have shown that multi-wall CNFs and the CNF/ PDMS composite were

30 non-toxic, with the 1-week cell viability exceeding 95%. The present material is designed for haemostatic purposes and will be in contact with skin for a short

- time. *In vivo* skin compatibility tests were therefore conducted by attaching a 10 mm by 10 mm patch of the present material onto a rat's skin (hair shaved) with clinically approved tape for 12 h (see Figures 24a and 24b). Compared with the skin under the pristine gauze (control), the skin under the present CNF gauze appeared normal and no itching or erythema was observed after 12 h (see Figure 24c). Thus, based on previous reports on cell non-toxicity and this skin compatibility test, the present CNF/PDMS material would be safe for haemostatic use.
- 10 As described above, a strategy for the design of wound-dressing materials featuring both rapid coagulation and facile clot removal has been developed and demonstrated, based on superhydrophobicity imparted by surface-immobilized CNFs. The developed SHP haemostatic CNF gauzes are shown to simultaneously combine a host of therapeutic functionalities. First: they
- 15 achieve fast clotting by expediting the formation of micro-fibrin fibres. Second: clotting without blood loss can be achieved, owing to the pressure-resistant non-wetting properties of the SHP surface. Third: there is minimal contact between the clot and the present SHP CNF surface. Driven by the contractile stress in the clot contraction phase, the SHP CNF haemostatic surface will
- 20 inherently tend to detach from the clot upon clot maturation, thus allowing a natural and unforced removal of the haemostatic dressing without causing secondary bleeding. Fourth: the SHP CNF patch can significantly reduce bacterial attachment, reducing infection risk. Finally: the CNFs are immobilized in a polymeric matrix on the gauze substrate, preventing free
- 25 micro/nanoparticles or fibres from entering the vascular stream. Moreover, skin compatibility tests show that the present materials can be safe for haemostatic use.

- Different from existing haemostatic products/materials, the present material
- 30 works in a distinctive manner. It stops bleeding first, reinforces clotting subsequently, and enables unforced natural detachment in the end. This offers

the following advantages. First: the SHP feature helps stop bleeding immediately upon application. Bleeding therefore ceases before the formation of a strong clot. This can be life-saving for haemorrhage cases such as in severe accidents and military combat, as it takes time for clotting to occur even with the most effective haemostatic materials. Second: rapid clotting is achieved by the nano-engineered surface structure, without the help of active clotting agents/chemicals. The present nano-engineered fibrous structure promotes quick growth of fibrin network to seal the wound. Clotting agents are routinely used in haemostatic products, such as the hydrophobically modified chitosan, which may provide shorter clotting time than the present material. However, the present material brings it with a distinct strategy of hastening coagulation using a nano-structuring approach also ensuring facile removal, a critical feature to avoid secondary bleeding – a quantitative comparison between the present material and the hydrophobically modified chitosan is provided in Figure 24).

The present material yields a reduction of the clot peeling force, by a good order of magnitude. Compared with commercial haemostatic products, the clot-peeling tension of the present material is about 24–52 times smaller (Figure 23c). This natural and facile clot detachment feature is unique to SHP haemostatic materials disclosed herein.

20

The present methods therefore pioneer a strategy for designing more efficient haemostatic material using surface nano-engineering. It has been demonstrated that a nano-structured surface can achieve fast clotting with unique blood loss-free, unforced detachment from the wound site, and also reduced bacteria adhesion. The multifunctional material concept presented herein shows clear potential to significantly advance the state-of-the-art of wound dressings, and medical devices, bringing benefits to common wounds, surgery, and even haemophilia. In the current study, non-biodegradable carbon nanofibres were used to design the nano-engineered surface. Biodegradable nanofibres could be developed for internal use and are intended to be encompassed by the present disclosure.

30

For completeness, fibrin fibre generation tests (Figures 4 and 6 to 11) were performed using blood and platelet poor plasma (PPP) with the presence of EDTA or sodium citrate. Fresh EDTA porcine blood was collected into sterile vacutainer tubes with K2 EDTA (3.0 ml, purple) and transported in a sealed foam container with ice. PPP was obtained by centrifuging the blood in EDTA tubes first at 1800 G for 10 min (Sorvall ST 8 Centrifuge, at room temperature – 20 °C) and then at 3000 G for another 10 min. As EDTA would prevent blood coagulation by irreversibly chelating calcium ions, EDTA blood was not suitable for coagulation testing. For *in vitro* clotting tests (see Figures 13 and 16), 3.8% sodium citrate was used as the anticoagulant. Citrated porcine blood from the same vendor was collected into polypropylene tubes by mixing blood with 3.8% sodium citrate solution at a volume ratio of 9:1 and was transported following the same protocol.

CAs and RAs were measured using a custom-built device. CA was measured using the sessile method by dispensing 5 µl liquids (deionized (DI) water, blood and PPP) on flat substrates (the CNF/ PTFE Ti and CNF/PDMS Ti surfaces) and 20 µl liquids on the CNF Ti mesh. Dynamic CAs, including advancing CA θ_a , receding CA θ_r , CA hysteresis CAH ($CAH = \theta_a - \theta_r$), nominal receding CA θ_{r_nom} , nominal CA hysteresis CAH_{nom} ($CAH_{nom} = \theta_a - \theta_{r_nom}$), and RA θ were measured using the tilting method, by placing a 20 µl droplet (for water, blood, and PPP with or without anti-thrombin) on the sample surface and tilting the sample till droplet roll-off. For PPP or blood on the SHP CNF surface, due to the existence of fibrin fibres, RA θ was defined as the tilt angle when fibrin fibres fractured and the droplet rolled off quickly. Static and dynamic CAs were averaged over four repetitions ($n = 4$). Surface morphology of the prepared samples and the surfaces after blood/plasma tests were characterized by SEM (SEC, SNE-4500M) after gold coating.

40 -

The anti-thrombin, argatroban (Purity: > 98%) was first dissolved in 0.9% NaCl solution and then added into EDTA blood or PPP to achieve a final dose of 2 mg ml⁻¹, which was higher than the value used in animal studies to ensure sufficient thrombin inhibition.

5

The porcine fibrin ELISA kit was ordered from MyBioSource, Inc. (Catalogue #MBS261977). Sliding tests with EDTA PPP droplets were respectively performed on three SHP CNF/PTFE and CNF/PDMS Ti surfaces (15 mm by 7 mm) and fibrin fibre generation during the sliding test was confirmed using the setup in Figure 5. After sliding test, the standard diluent from the kit was flushed over the surfaces using a pipette into reaction wells. As a control, the same procedures were performed on three pristine SHP CNF/PTFE Ti surfaces. Then, an ELISA test was performed following standard instructions. Optical density reading (450 nm) was used to calculate the fibrin concentration washed into the reaction wells from the tested SHP surface, with reference to the standard curve (fibrin concentration to optical density) acquired with porcine fibrin standard samples, to verify the existence of fibrin on the CNF surfaces after PPP sliding tests.

For testing static fibrin growth, forty microliters of EDTA PPP was dispensed onto a SHP CNF/PTFE Ti surface placed in a plastic petri dish and cultivated for 4 min at 37 °C. The reaction was terminated by slowly adding sufficient DI water into the petri dish. The samples were then carefully rinsed by dipping into DI water three times. Samples were air-dried and fibrin structures grown on the surface were observed under SEM after gold coating.

For convenient observation of bacteria attached on the surface, a SHP CNF/PTFE coating was half-coated on a glass slide following the same procedure for the CNF/PTFE Ti surface. The SHP CNF/PTFE glass surface and the CNF/PTFE Ti surface were shown to have the same performance through water, blood, and PPP testing. Ten microliters of glycerol stock (50% glycerol, 50% cell

41 -

culture in Luria–Bertani (LB)) of *E. coli* harbouring constitutive GFP expression plasmid stored at $-80\text{ }^{\circ}\text{C}$, was added into 3 ml fresh LB broth, supplemented with Kanamycin (Km) ($50\text{ }\mu\text{g ml}^{-1}$). The cells were incubated in a shaking incubator at $37\text{ }^{\circ}\text{C}$ with a shaking speed of 225 r.p.m. The culture was diluted
5 to OD_{600} of 0.5 with fresh LB ($50\text{ }\mu\text{g ml}^{-1}$ Km). The CNF/PTFE glass slide sample was first sterilized with ultraviolet, then $40\text{ }\mu\text{l}$ cell culture was flushed over the sample surface across areas with and without CNF coating. The sample was subsequently air-dried in a biosafety cabinet for 20 min. Bacteria attached on the CNF surface were observed under a confocal laser scanning microscope,
10 with laser wavelength of 473 nm for exciting the GFP.

To compare the clotting performance of the present CNF gauze, chitosan gauze was prepared by spray coating the chitosan and PDMS composite dispersion onto the pristine gauze. For fair comparison, weight ratio of chitosan to PDMS
15 was 1:2, the same as CNF to PDMS weight ratio for the present CNF gauze. Twenty-five milligrams of chitosan (medium molecular weight, Sigma Aldrich) was first dissolved in 5 ml 0.15 M acetic acid with an ultrasonic probe; 50 mg PDMS (pre-polymer to cross linker weight ratio 9:1) was dispersed in 10 ml acetone by the ultrasonic probe. After mixing the two dispersions under
20 ultrasonication for 5 min, the composite dispersion was spray-coated onto a pristine gauze (size: 10 cm by 10 cm) following the same protocol for preparing the CNF gauze. After coating, the chitosan gauze was baked at $80\text{ }^{\circ}\text{C}$ for 1 h to evaporate the solvent.

25 The clotting test was performed using the citrated blood. Different types of gauzes were used – the SHP CNF gauze and the normal superhydrophilic gauze, in patches of 15 mm by 15 mm. These were pre-warmed in 20 ml polystyrene (PS) plastic petri dishes ($37\text{ }^{\circ}\text{C}$ water bath). After mixing the citrated blood with 0.2 M CaCl_2 at a volume ratio of 10:1 to initiate coagulation, $20\text{ }\mu\text{l}$ blood was
30 immediately dispensed and sandwiched between two gauzes in the petri dish as shown in Figure 15a - on the CNF gauze, blood was in contact with the surface

coated with CNF. Blood dispensed in an empty petri dish without gauze was used as the control case, denoted as Control in Figure 13d. Blood dispensed on different samples was allowed to coagulate at 37 °C for 0, 3, and 5 min, terminating coagulation by adding 10 ml DI water into the petri dish without
5 disturbing the clot. Non-clotted red blood cells (free blood cells not trapped in the clot) would haemolyse and release haemoglobin into the water. Optical absorbance of the resulting haemoglobin solution at different clotting time t , $HA(t)$, measured by the spectrophotometer at 540 nm, would represent the amount of haemoglobin from unclotted red blood cells. $HA(0)$. The absolute
10 haemoglobin absorbance of 20 μ l blood in 10 ml DI water at $t = 0$ min, was used as reference. $RHA(t)$, the relative haemoglobin absorbance at clotting time t , was calculated as $HA(t)/HA(0)$. This was averaged over three repetitions, $n = 3$, using the same batch of blood. As the haemoglobin came from unclotted red blood cells, a smaller $RHA(t)$ would mean faster clotting.

15

As for the nominal blood contact area (NCA) during the clotting test, following the same protocol, 20 μ l blood was placed between two normal gauzes or two CNF gauzes for coagulation to occur. After 1 h, the two pieces of gauze were separated and a picture taken from the top was processed to calculate the NCA
20 (averaged over three repetitions, $n = 3$). NCA on the normal gauze was considered as the darkened area, which was soaked by blood, whereas NCA on the CNF gauze was considered as the clot-gauze contact area (Figure 15b). In this test, the uncoated gauze absorbed blood quickly, leading to a large blood contact area, whereas the SHP CNF gauze repelled blood and its blood contact
25 area was only $14.2 \pm 0.7\%$ (mean $\pm$ SD, $n = 3$) of that on the uncoated pristine gauze (Figure 15b). Following the same protocol, clotting performance of the present CNF gauze was further compared with the chitosan gauze with 50 μ l blood ($n = 3$), the results (Figure 21) showing that the present CNF gauze can have better clotting performance than the chitosan gauze.

30

For determining clotting without blood loss: mimicking a wound dressed with a medical gauze, an opening representing the wound (8 mm by 5 mm) was made on a silicone tube (inner diameter: 7.8 mm) and was covered with the SHP CNF gauze by medical tape (Figure 15c). A silicone tube dressed with the normal gauze was used as control. The citrated blood was mixed with CaCl_2 to initiate coagulation. Blood (1.5 ml) was quickly filled into the silicone tube and placed in a petri dish. After 10 min at room temperature, the tube was removed from the petri dish and the weight increase in petri dish due to blood leakage was measured as the blood loss. The test was repeated three times ($n = 3$) to get the average blood loss.

To test the maximum pressure that the CNF gauze (without a back-supporting impervious membrane) could withstand before blood leakage, citrated blood was slowly filled into a long tube with a hole sealed by the CNF gauze (Figure 15d). The height of blood column h , which was measured at the moment of blood droplets leaking through the CNF gauze, was used to calculate the maximum anti-leakage pressure P for a single layer of CNF gauze without the back-supporting impervious membrane ($P = \rho gh$, where ρ is the blood density and g is the gravitational constant, averaged over three repetitions; $n = 3$).

The anti-wetting property of the SHP CNF gauze with a back-supporting impervious membrane was further tested using the setup in Figure 15e. A transparent PS petri dish was glued (epoxy glue) onto the back of the SHP CNF gauze as the impervious membrane. Glue was not applied at the central part of the gauze. Subsequently, the petri dish with the attached CNF gauze was placed in a transparent pressure chamber and filled with blood to completely immerse the CNF gauze (Figure 15e). Citrated blood was used to ensure blood fluidity. Hydrostatic pressure applied on the CNF gauze surface was controlled by pumping air into the sealed pressure chamber with the assistance of a check valve and a pressure sensor. As the CNF gauze was placed 2–3 mm deep in the blood, the air pressure in the pressure chamber was taken as the hydrostatic

pressure exerted onto the CNF gauze. A camera projecting upward was used to detect whether the central part of the CNF gauze was wetted by blood at a given pressure.

- 5 To test clot detachment, the SHP CNF Ti mesh and the SHP CNF gauze were placed in a 20 ml PS plastic petri dish pre-warmed at 37 °C (Figure 15a). After mixing citrated blood with 0.2 M CaCl₂ at a volume ratio of 10:1 to initiate coagulation, 100 µl blood was immediately dispensed onto the CNF Ti mesh or the CNF gauze for the clot to form. After natural solidification and drying
10 overnight, clots on the SHP CNF surfaces, before and after detachment, were coated with gold for SEM observation.

To detect clot peeling force, citrated blood was mixed with 0.2 M CaCl₂ at a volume ratio of 10:1 to trigger the coagulation. The blood clot was allowed to
15 form and solidify overnight at 37 °C on the CNF gauze surface or between two pieces of uncoated gauze by dispensing 20 µl blood onto the CNF gauze surface or onto two stacked normal gauzes (Figure 22). Using the setup in Figures 22a and 22b, the force required for peeling the clot from the CNF-coated gauze or the uncoated gauze was measured. Figure 22a shows a clot 2200 formed by
20 dispensing 20 µl blood onto the CNF gauze surface 2202. The gauze 2202 was mounted onto a strain gauge force transducer 2204 (capacity: 980 mN, resolution: 0.1 mN) by 3M high-strength tape 2206. A thin cotton 2208 wire was glued onto the clot 2200 by epoxy 2210. The force for peeling the clot 2200 was recorded by the force sensor 2204. Figure 22b shows a clot 2212
25 formed between two normal gauzes 2214, 2216 by dispensing 20 µl blood. The force required to peel the two normal gauzes 2214, 2216 was measured while pulling the top gauze 2214. Pulling motion was performed with a homemade translation stage at 0.5 mm/s and the force data were recorded by a National Instruments multifunction data-acquisition device (NI USB-6218) controlled by
30 Labview. Figure 22c shows a typical clot peeling force curve, showing the maximum peeling force $F_{\max}$. Figure 22d provides the measurement of the clot

maximum width W for the calculation of the clot peeling tension $F_{\max}/W$. As the peak peeling force $F_{\max}$ (Figure 22c) would occur at the maximum clot width W , the normalized clot peeling tension $F_{\max}/W$ (averaged over three repetitions) was used to compare the clot peeling force for the CNF gauze and the uncoated
5 normal gauze.

For the *in vivo* test, CNF gauze and the normal uncoated gauze (used as control) were prepared to be plaster-like – Figures 21a and 21b (CNF gauze 2106 and normal gauze 2108). The gauze was cut to be 35 mm long and 15
10 mm wide, and was mounted onto a transparent adhesive tape (50 mm long and 25 mm wide). An opening (10 mm long and 5 mm wide) was made in the centre of adhesive tape, allowing blood to seep through the gauze in case of excessive bleeding. In Figure 21a, for the design of the plaster-like gauze, the CNF gauze or the normal gauze 2100 (35 mm by 15 mm) was applied onto the
15 adhesive film 2102 (50 mm by 25 mm), and the opening 2104 was made on the adhesive film 2102 to allow blood to seep through in case of excessive bleeding. Prepared CNF gauze and normal gauze were sterilized under UV for 30 min before experiment.

20 For the *in vivo* experiments, female rats 2110 (Sprague–Dawley, 11–13 weeks old, average body weight: 255.6 ± 19.7 g, mean $\pm$ SD) were used, complying with all relevant ethical regulations for animal testing and research. Rats were anaesthetized by isoflurane (4–5% in 100% oxygen for induction and 1–3% for maintenance) with a thermal pad to prevent hypothermia. Hair on the back
25 was first shaved with an electric razor. Residual hair was removed with hair removal cream. Surgical instruments were autoclaved. The experiment was performed in the biological safety cabinet using aseptic techniques. The operative site on back of the rat was disinfected with iodine and 70% v/v ethanol successively for three times. Two incisions 2112 (about 1 cm long; cut
30 down to muscle) were made on the back of the rat, one on the left and one on

the right at the same position. Upon incision, the plaster-like gauze was applied onto the wound instantly (Figures 21c and 21d).

Blood loss was characterized by measuring the weight increase in gauze.

- 5 Weight of the gauze peeled at 3 min (Figure 21e) minus its initial weight was taken as the blood loss. Gauze weight was measured by a high-precision weighing balance (resolution: 1 mg). The data were averaged over six repetitions.
- 10 The peeling force was measured by peeling the gauze along the wound after about 2 h (132 ± 10 min and 133 ± 10 min for the control and the CNF gauze, respectively; mean $\pm$ SD, $n = 5$), allowing the clot to sufficiently mature and solidify. Before gauze peeling, the adhesive film around the gauze was carefully trimmed away, otherwise the adhesive film taped on skin would interfere with
- 15 the peeling force. With reference to Figure 21g, in the absence of trimming the adhesive film (2114) taped onto skin would interfere with the peeling force. Care was taken not to stretch or tear the gauze during the trimming. After adhesive film trimming, one end of the gauze was adhered onto a metal hook 2116 by high-strength tape 2118, which was attached onto a force sensor
- 20 (capacity: 980 mN; resolution: 0.1 mN) by a thin cotton wire 2120 – see Figure 21h. Peeling was assisted by a vertical numerical translation stage and the peeling force was recorded. The maximum force during peeling, averaged over five measurements see Figure 21i, was used for a quantitative comparison.
- 25 The skin compatibility test was also conducted *in vivo* by attaching the present material onto rat skin for 12 h (see Figure 25). Preparation of the samples used for the skin compatibility test is illustrated in Figure 25a, with a 10 mm by 10 mm CNF sample 2500 attached onto a 20 mm by 20 mm pristine gauze 2502.
- 30 This design permitted comparison of two adjacent skin areas contacted with different gauze materials. Before testing, rats were anesthetized and shaved.

47 -

Prepared samples were then applied on rat skin of three rats, with transparent clinical tape for 12 h (see Figure 25b). To prevent samples from being scratched off by rats, rat jackets were used to protect the gauze samples. After 12 h, rats were anesthetized to remove the gauze. The tested skin area was gently
5 cleaned with 70% v/v ethanol to examine any difference between the areas under the present CNF gauze and the pristine gauze (see Figure 25c) – no difference was observed.

It will be appreciated that many further modifications and permutations of
10 various aspects of the described embodiments are possible. Accordingly, the described aspects are intended to embrace all such alterations, modifications, and variations that fall within the spirit and scope of the appended claims.

Throughout this specification and the claims which follow, unless the context
15 requires otherwise, the word “comprise”, and variations such as “comprises” and “comprising”, will be understood to imply the inclusion of a stated integer or step or group of integers or steps but not the exclusion of any other integer or step or group of integers or steps.

20 The reference in this specification to any prior publication (or information derived from it), or to any matter which is known, is not, and should not be taken as an acknowledgment or admission or any form of suggestion that that prior publication (or information derived from it) or known matter forms part of the common general knowledge in the field of endeavour to which this
25 specification relates.