

ASEPTIC CONNECTORS FOR BIO-PROCESSING CONTAINERS

CROSS-REFERENCE TO RELATED APPLICATION

[0001] This application is a continuation-in-part of U.S. Patent Application No. 14/292,637, filed May 30, 2014, and titled "ASEPTIC CONNECTORS FOR BIO-PROCESSING CONTAINERS," and U.S. Patent Application No. 14/724,659, filed May 28, 2015, and titled "ASEPTIC CONNECTORS FOR BIO-PROCESSING CONTAINERS," which are herein incorporated by reference in their entireties and for all purposes.

10

BACKGROUND

[0002] Over the last few decades there has been a progressive shift to single-use or disposable components in the worlds of medicine and pharmaceutical development/bioprocessing. Much of this shift has been pushed by sterility and health requirements but also by issues such as cost-per-use as mediated by convenience and minimization of labor and overhead considerations. In this context, a sensing apparatus for bioprocessing must not interfere with the benefits of single-use systems.

SUMMARY

[0003] Provided herein are apparatuses and methods for installing a sterilized peripheral in a bio-processing vessel or component. One aspect is an aseptic peripheral connection assembly for installing a sterilized peripheral in a bio-processing vessel or component via an aseptic connector affixed to the vessel. The bio-processing component may be a self-contained container (e.g. bio-reactor) or a flow path. In some embodiments, the bio-processing component may be a bioreactor or a filter. The aseptic peripheral connection assembly includes a carrier, an applicator, a plunger, and a removable hermetic sealing tab.

[0004] The carrier includes the sterilized peripheral and a sealing member configured to form a leak-tight seal with an aseptic connector or aseptic vessel connector on the bio-processing vessel or component at a location where the carrier is

to be installed. The sealing member may include a clip or a ledge. In some embodiments, the sealing member includes an o-ring.

5 [0005] The carrier may be disk shaped with the peripheral having an exposed sensing surface on a flat side of the carrier. In some embodiments, the carrier includes two or more peripherals having exposed sensing surfaces on a flat side of the carrier. In various embodiments, the carrier may be generally sheath shaped with the peripheral having an exposed sensing surface on an end of the carrier.

10 [0006] The peripheral may be configured or designed to detect an optical response. In some embodiments, the peripheral is an electrochemical peripheral, or a temperature peripheral, or a pH peripheral, or an oxygen peripheral. In some embodiments, the carrier includes an oxygen peripheral, a pH peripheral, and a temperature peripheral. In various embodiments, the peripheral is a single use peripheral. The carrier may include two or more peripherals.

15 [0007] The applicator includes a sleeve and vessel-facing opening adjacent to an applicator connector configured to temporarily connect to the aseptic vessel connector on the bio-processing vessel. In some embodiments, the applicator includes a sleeve and component-facing opening adjacent to an applicator connector configured to temporarily connect to the aseptic connector on the bio-processing component. The applicator may be configured to be removed from the carrier and the bio-processing vessel or component after the carrier is installed in the bio-processing vessel. The applicator sleeve may be constructed of a rigid material. In some embodiments, the applicator sleeve has a tubular shape and the plunger has a substantially circular shape that forms a seal with an interior surface of the applicator.

25 [0008] The applicator connector may include a clip or a ledge. In some embodiments, the applicator connector may include an o-ring. The applicator connector may be configured to provide a hermetic seal with the aseptic connector or aseptic vessel connector while the plunger inserts the carrier into the aseptic connector or aseptic vessel connector. In some embodiments, the applicator is composed of polycarbonates, polysulfone, polyvinylidene fluoride, or co-polyester. In various embodiments, the applicator is composed of USP Class VI material that is animal derived component free, latex free, phthalate free, and capable of being sterilized by

ionizing radiation. In some embodiments, sterilizing by ionizing radiation may involve e-beam or gamma sources.

5 [0009] The plunger is within the applicator sleeve and be configured to plunge the carrier from a position within the applicator sleeve to a position engaged with the aseptic connector or aseptic vessel connector to form the leak-tight seal. In some embodiments, the plunger includes an o-ring for forming a hermetic seal with the applicator sleeve while plunging the carrier into position with the aseptic connector or aseptic vessel connector.

10 [0010] The removable hermetic sealing tab covers the vessel-facing or component-facing opening of the applicator sleeve to maintain the sterilized peripheral in aseptic condition prior to installation in the bio-processing vessel or component. The removable hermetic sealing tab may be configured to be removed from the aseptic peripheral connection assembly after connecting the applicator to the aseptic vessel connector on the bio-processing vessel at a location where the sterilized peripheral is
15 to be installed, and before plunging the carrier from a position within the applicator sleeve to a position engaged with the aseptic vessel connector. In some embodiments, the removable hermetic sealing tab includes a film or sheet having a thickness of between about 1 mil and about 20 mil (thousandths of one inch). The removable hermetic sealing tab may be composed of a simple film if the bio-processing vessel is
20 to be used such that the internal pressure is less than about 1 psig. In some embodiments, the removal hermetic sealing tab may include a USP Class VI, latex free, phthalate free animal derived component free polymeric plate. In some embodiments, the removal hermetic sealing tab is coated with an adhesive.

25 [0011] Another aspect is a bio-processing vessel or component kit including the above aseptic peripheral connection assembly and a bio-processing vessel or component housing with the aseptic vessel connector or aseptic connector affixed. In various embodiments, the aseptic vessel connector is affixed to the bio-processing vessel housing. The bio-processing component housing may be a single use bio-processing vessel housing or a self-contained container or a flow path. In some
30 embodiments, the bio-processing component housing is configured or designed as a bioreactor or a filter. In various embodiments, the bio-processing component housing

is a filter with a flow path. The bio-processing component housing may be a filter including a container packed with material for product separation.

[0012] Yet another aspect is a method of fabricating an aseptic peripheral connection assembly for installing a sterilized peripheral in a bio-processing vessel via an aseptic connector affixed to the vessel by (a) placing the carrier in an applicator; (b) sterilizing the aseptic peripheral connection assembly such that the sterilizing does not employ exposure to radiation at a level of greater than about 15 kGy; and (c) packaging the aseptic peripheral connection assembly in a hermetically sealed package. The aseptic peripheral connection assembly includes (i) a carrier including the peripheral and a sealing member configured to form a leak-tight seal with the aseptic vessel connector on the bio-processing vessel at a location where the carrier is to be installed; and (ii) the applicator including a sleeve and vessel-facing opening adjacent to an applicator connector configured to temporarily connect to the aseptic vessel connector on the bio-processing vessel. The method may further include sending the packaged aseptic peripheral connection assembly to a site for installation in the bio-processing vessel. In alternative embodiments, the bio-processing vessel described with respect to the disclosed embodiments may instead be a flow path or filter path.

[0013] In some embodiments, the method may include, prior to (a), applying the peripheral to a carrier structure. The method may include, prior to (a): treating a carrier structure, which does not include the complete peripheral, by a process that reduces colony forming units (CFUs) on the carrier structure; and subsequently applying the peripheral to a carrier structure. In some embodiments, the method may further include calibrating the peripheral in the carrier; and storing information from the calibrating.

[0014] The sterilizing further includes exposing the aseptic peripheral connection assembly and packaging to gamma, beta, and/or x-ray radiation. In some embodiments, the sterilizing includes plasma cleaning the aseptic peripheral connection assembly. In some embodiments, the plasma cleaning is performed prior to (c). In some embodiments, plasma cleaning performed prior to (c) may involve packaging prior to sterilizing with ethylene oxide. In some embodiments, the plasma cleaning includes sterilizing using atmospheric plasma at a temperature less than

about 40°C. In some embodiments, the plasma cleaning is performed at room temperature. The plasma cleaning may include sterilizing using exposure to non-toxic gases in a plasma. Examples of toxic gases that are not used include formaldehyde and ethylene oxide. In some embodiments, ethylene oxide may be used if adopted for
5 a flow-through system. An example of a non-toxic gas used in plasma cleaning is air.

[0015] The sterilizing may further include exposing the aseptic peripheral connection assembly to gamma, beta, and/or x-ray radiation after the packaging in (c). The packaging may include a vacuum packaging procedure.

[0016] The aseptic peripheral connection assembly may further include: (iii) a
10 plunger within the applicator sleeve and configured to plunge the carrier from a position within the applicator sleeve to a position engaged with the aseptic vessel connector to form the leak-tight seal; and (iv) a removable hermetic sealing tab covering the vessel-facing opening of the applicator sleeve to maintain the sterilized peripheral in aseptic condition prior to installation in the bio-processing vessel.

15 [0017] Another aspect is a method of using an aseptic peripheral connection assembly for installing a sterilized peripheral in a bio-processing vessel via an aseptic connector affixed to said vessel by (a) connecting an applicator connector of an applicator to the aseptic connector on the bio-processing vessel; (b) removing a hermetic sealing tab covering the vessel-facing opening of the applicator sleeve; and
20 (c) plunging the carrier from a position within the applicator sleeve to a position engaged with the aseptic vessel connector, and forming the leak-tight seal. The aseptic peripheral connection assembly may include: (i) a carrier including the peripheral and a sealing member configured to form a leak-tight seal with the aseptic vessel connector on the bio-processing vessel at a location where the carrier is to be
25 installed; and (ii) the applicator including a sleeve and vessel-facing opening adjacent to the applicator connector. In alternative embodiments, the bio-processing vessel described with respect to some embodiments may instead be a flow path or filter path.

[0018] Connecting the applicator connector to the aseptic vessel connector in (a) may include providing a hermetic seal with the aseptic vessel connector while
30 plunging the carrier into the aseptic vessel connector. In various embodiments, removing the hermetic sealing tab is performed after connecting the applicator

connector to the aseptic vessel connector on the bio-processing vessel, and before plunging the carrier from a position within the applicator sleeve to a position engaged with the aseptic vessel connector.

5 [0019] In some embodiments, the method further includes removing the applicator from the carrier and the bio-processing vessel after the carrier forms said leak-tight seal with the bio-processing vessel. Connecting the applicator connector to the aseptic vessel connector may include providing a hermetic seal with the aseptic vessel connector that is maintained while plunging the carrier into the aseptic vessel connector.

10 [0020] The bio-processing vessel may be a self-contained container. In some embodiments, the bio-processing vessel may be a bioreactor. The bio-processing vessel may be a single use bio-processing vessel. In alternative embodiments, the bio-processing vessel described with respect to the above embodiments may instead be a flow path or filter path.

15 [0021] In various embodiments, the peripheral is configured or designed to detect an optical response. The peripheral may be a single use peripheral. The carrier may include two or more peripherals. In some embodiments, the carrier includes an oxygen peripheral, a pH peripheral, and a temperature peripheral. The carrier may be disk shaped with the peripheral having an exposed sensing surface on a flat side of the carrier. In various embodiments, the carrier may be generally sheath shaped with the peripheral having an exposed sensing surface on an end of the carrier. The sealing member may include a clip or a ledge. In some embodiments, the applicator sleeve is constructed of a rigid material. The applicator sleeve may have a tubular shape and the plunger has a substantially circular shape that forms a seal with an interior surface
20 of the applicator. The applicator connector may also include a clip or a ledge.
25

[0022] Another aspect involves an aseptic peripheral connection assembly for installing a sterilized peripheral in a bio-processing component via an aseptic connector affixed to the component, the peripheral connection assembly including: a. a carrier disposed within an applicator sleeve, the carrier including the sterilized peripheral and a sealing member configured to form a leak-tight seal with the aseptic connector on the bio-processing component at a location where the carrier is to be
30

installed, the carrier configured to be inserted from a position within the applicator sleeve to a position engaged with the aseptic connector to form the leak-tight seal; b. an applicator including the applicator sleeve and an applicator connector, the applicator connector including a component-facing opening configured to connect to
5 the aseptic connector on the bio processing component; and c. a removable hermetic sealing tab covering the component-facing opening of the applicator sleeve to maintain the sterilized peripheral in aseptic condition prior to installation in the bio processing component.

[0023] The bio-processing component may be a self-contained container or a flow
10 path. In some embodiments, the bio-processing component is a bioreactor or a filter flow path. In various embodiments, the peripheral is a single use sensor.

[0024] The carrier may include two or more sensors. In some embodiments, the carrier includes one or more sensors having exposed sensing surfaces on a side of the carrier. The carrier may be generally sheath shaped and includes a sensor having an
15 exposed sensing surface on an end of the carrier.

[0025] The applicator may include a clip or a ledge configured to engage with an aperture on the applicator sleeve. In some embodiments, the sealing member includes an o-ring. The applicator sleeve may be constructed of a rigid material. In some
20 embodiments, the applicator sleeve has a tubular shape and has a substantially circular shape that forms a seal with an interior surface of the applicator.

[0026] In some embodiments, the aseptic peripheral connection assembly further includes a clamp configured to clamp the applicator connector and the aseptic connector. The applicator may include one or more apertures and the carrier may include one or more clips configured to be inserted into the one or more apertures
25 when the carrier is engaged with the aseptic connector to form the leak-tight seal. In some embodiments, the peripheral is installed in the bio-processing component such that the carrier in its installed position has one end extending through the aseptic connector and within the bio-processing component. In some embodiments, the applicator connector is configured to provide a hermetic seal with the aseptic
30 connector while the carrier is inserted into the aseptic connector.

[0027] In various embodiments, the applicator is made from polycarbonate, polysulfone, polyvinylidene fluoride, co-polyester, or any combination thereof. In various embodiments, the applicator is composed of USP Class VI material that is animal derived component free, latex free, phthalate free, and capable of being
5 sterilized by ionizing radiation. In some embodiments, sterilizing by ionizing radiation may involve e-beam or gamma sources. The carrier may include an o-ring for forming a hermetic seal with the applicator sleeve when the carrier is inserted into position with the aseptic connector.

[0028] The removable hermetic sealing tab may be configured to be removed from
10 the aseptic peripheral connection assembly after connecting the applicator to the aseptic connector on the bio processing component at a location where the sterilized peripheral is to be installed, and before inserting the carrier from a position within the applicator to a position engaged with the aseptic connector. In some embodiments, the removable hermetic sealing tab includes a film or sheet having a thickness of
15 between about 1 mil and 20 mil. The removable hermetic sealing tab may include a USP Class VI, latex free, phthalate free animal derived component free polymeric plate. In some embodiments, the removal hermetic sealing tab is coated with an adhesive.

[0029] Another aspect involves a bio-processing component kit including: a. the
20 aseptic peripheral connection assembly as described above; and b. a bio-processing component housing with the aseptic connector affixed. In some embodiments, the bio-processing component housing is a single use bio-processing vessel housing or a flow path. The bio-processing component housing may be configured or designed as a bioreactor or a filter with a flow path. In some embodiments, the bio-processing
25 component housing includes a container packed with material for product separation.

[0030] Another aspect involves a method of fabricating an aseptic peripheral connection assembly for installing a sterilized peripheral in a bio-processing component via an aseptic connector affixed to the component, whereby the aseptic peripheral connection assembly includes (i) a carrier disposed within the applicator
30 sleeve, the carrier including the sterilized peripheral and a sealing member configured to form a leak-tight seal with the aseptic connector on the bio-processing component at a location where the carrier is to be installed; and (ii) an applicator including an

applicator sleeve and an applicator connector, the applicator connector including a component-facing opening configured to connect to the aseptic connector on the bio-processing component, the method including: a. placing the carrier in the applicator; b. packaging the aseptic peripheral connection assembly in a hermetically sealed package; and c. sterilizing the aseptic peripheral connection assembly, whereby the
5 sterilizing does not employ exposure to radiation at a level of greater than about 15 kGy.

[0031] In some embodiments, the method further includes, prior to packaging, plasma cleaning the carrier in the applicator. In some embodiments, the method
10 further includes adhering a hermetic sealing tab on the component-facing opening of the applicator connector.

[0032] Another aspect involves a method of using an aseptic peripheral connection assembly for installing a sterilized peripheral in a bio-processing component via an aseptic connector affixed to the component, whereby the aseptic peripheral
15 connection assembly includes (i) a carrier including the peripheral and a sealing member configured to form a leak-tight seal with the aseptic connector on the bio-processing component at a location where the carrier is to be installed; and (ii) an applicator including a sleeve and an applicator connector including a component-facing opening, the method including: a. connecting the applicator connector of the
20 applicator to the aseptic connector on the bio-processing component; b. removing a hermetic sealing tab covering the component-facing opening of the applicator connector; and c. plunging the carrier from a position within the applicator sleeve to a position engaged with the aseptic connector, and forming the leak-tight seal therewith. The method may further include clamping the applicator connector to the aseptic
25 connector.

BRIEF DESCRIPTION OF THE FIGURES

[0033] Figure 1 shows the layer by layer breakdown of a Thermo Fisher bag film.

[0034] Figure 2 shows a port design with a seal based on a friction fit.

[0035] Figure 3 shows relative phase difference of the excitation and fluorescent
30 signal.

- [0036] Figure 4 shows a fiber-optic phase fluorimetry system.
- [0037] Figure 5 shows a free space optical phase fluorimetry system.
- [0038] Figure 6 shows a free space optical carrier.
- [0039] Figure 7 shows the carrier of Figure 6 in a mounting flange for use in
5 flexible bio-processing system containers.
- [0040] Figure 8 shows a free space carrier (sheath) used in single-use stirred tanks systems.
- [0041] Figure 9 shows one side of an aseptic connector.
- [0042] Figure 10 shows both sides of an aseptic connector system.
- 10 [0043] Figure 11 shows inactivation rates for bacteria as a function of UV dosage.
- [0044] Figure 12 shows reduction in CFU levels as a function of exposure to cold atmospheric plasma treatment.
- [0045] Figure 13 shows both halves of an aseptic sensor connection assembly prior to connecting them.
- 15 [0046] Figure 14 shows the detailed cross-sectional view of the sensor connector assembly shown in Figure 13.
- [0047] Figure 15 shows the step-by-step deployment of the aseptic sensor connector assembly.
- [0048] Figure 16 shows both sides of a sheath based sensor aseptic connector
20 system.
- [0049] Figure 17 shows the detailed cross-sectional view of the sheath based sensor aseptic connector system of Figure 16.
- [0050] Figure 18 shows the first three steps in the deployment of the sheath based sensor aseptic connector system.

- [0051] Figure 19 shows the last two steps in the deployment of the sheath based sensor aseptic connector system.
- [0052] Figures 20A-20C show detailed cross-sectional views of the sheath based sensor aseptic connector system.
- 5 [0053] Figure 21 shows a detailed cross-sectional view of the sheath based sensor aseptic connector system with clamp.
- [0054] Figure 22 shows a connected sheath based sensor aseptic connector system with clamp.
- [0055] Figure 23 shows an alternative method of capturing the sensor carrier and
10 flange in the aseptic vessel connector using threads.
- [0056] Figure 24 shows an alternative method of capturing the sensor carrier and flange in the aseptic vessel connector using a press fit.
- [0057] Figure 25 shows an alternative method of capturing the sensor carrier and flange in aseptic vessel connector using a bayonet mount.
- 15 [0058] Figure 26 shows a cross-sectional view of components of an assembly before the sensor carrier assembly is retained in the aseptic vessel connector assembly.
- [0059] Figure 27 shows a flow chart of a method currently employed for inserting single-use sensors into single-use bioprocess vessels. It details the construction
20 process, the calibration process, the sterilization process, and the transportation required between parties involved. The parties are the single-use sensor manufacturer, the single-use bioprocess vessel manufacturer, and the end-user.
- [0060] Figure 28 shows a flow chart detailing the method for inserting the sheaths for inserting single-use sensors into single-use bioprocess vessels utilizing the aseptic
25 connector described here.

DESCRIPTION OF CERTAIN EMBODIMENTS

[0061] The bio-processing market has moved rapidly to adopt single-use technology. This move towards the implementation of single-use systems for bio- can be readily understood by considering a typical biotech manufacturing facility. The infrastructure required to implement a facility using traditional glass/steel bioreactors, mixers, and purification systems is substantial, as is the time and expense required to construct the aforementioned facility. The requirement that both the equipment itself and also the ingress and egress tubing utilize inert materials such as 316L electro-polished stainless steel requires a large initial capital investment. Additionally, the bioreactors, mixers (i.e. bio-process vessels) and down-stream processing equipment (e.g. chromatography skids, filtration systems) all have fairly large footprints vis-à-vis available clean-room space and once installed tend to remain in fixed configurations. In contrast, the size and inherent nature of single-use platforms generally permits easier storage and re-configurability when compared to traditional, rigid glass/steel solutions. Other advantages of single-use systems include lower requirements for support infrastructure and time savings over traditional designs. Specifically there is a reduction in preparation and sterilization time, a reduction in the need for purified water, water-for-injection, steam generation and a significantly reduced post-growth-run maintenance time. Additionally, single use systems and their associated plastic tubing lend themselves to being re-configured and validated quickly and efficiently as manufacturing or process requirements change.

[0062] As bio-processing becomes more sophisticated in order to reduce manufacturing costs and gain market expansion to geographies with large populations that require significant price reductions, automation will gain in importance. In order to enable cost-effective local production where educated and trained resources as scarce, not only will automation need to reduce operator error and increase batch to batch reproducibility, but it will also need to minimize consumption of raw materials such as media, feed formulations and buffers. Production-on-demand of these materials will in turn limit manual unit operations and increase the level of automation required for the supporting infrastructure in the manufacturing plant. Furthermore, regulatory demands will drive more comprehensive data collection for

each batch produced, leading to more complex measurements strategies for each production process step.

[0063] All of these drivers will increase the number measurement points to enable better process control as well as more detailed batch records with automated software (rather than operator) driven process alarming, loop correcting, and deviation reporting in the manufacturing execution system layer, and subsequently in automatically generated electronic batch records. Therefore, the insertion of more and diverse sensors into upstream bio-process vessels will be become widespread.

[0064] Examples of such sensors in upstream and infrastructure vessels will include:

- pH, dissolved oxygen, temperature, headspace pressure, metabolic parameters such as glucose, cell density and cell viability for bioreactors
- pH, conductivity, temperature, and osmolality for media or buffer preparation mixers
- pH, conductivity and temperature for product holding, mixing, and freeze/thaw vessels

[0065] Similarly, single-use purification and product isolation skids also require more measurements points. These downstream process units, however, focus more on flow paths in and out of filters or chromatography columns to process the biological liquids in a continuous stream unlike upstream or infrastructure vessels that retain, recirculate, and/or mix the liquids in a fixed volume. Examples of such “flow path” sensors include:

- pH, conductivity, temperature, tubing pressure, and liquid flow for filtration skids used for harvest, clarification, virus removal, and ultra/dia-filtration of final product
- pH, conductivity, temperature, ultraviolet response (including spectra) for proteins or anti-bodies and other electronic-chip-based analytics for chromatography steps including product capture with Protein A and product separation using anion and/or cation exchange.

[0066] Sensors could also be used in single-use fill-and-finish skids for measurement of liquid flow, syringe filling, and final product purity verification (e.g. spectroscopy such as Raman). Thus, the ability to insert single-use sensors into a single-use process unit operation will be a ubiquitous requirement for the successful implementation of any single-use production facility. As the number of sensors multiplies, the ability to connect them to the process in a robust manner that does not increase the risk of contamination will become paramount.

[0067] Disclosed embodiments may be used for bio-processing components such as vessels or flow paths. The term “vessel” generally refers to a self-contained and sterilized liquid container, of whatever shape or configuration, e.g., a cylindrically shaped bioreactor or mixer. Some embodiments involve a flow path with multiple junctions and made of pressure-resistant plastic tubing. Here, flow path is any tubing set, filter, or single-use component used in bio-processing (typically downstream processing) that may be connected using an aseptic connector system; though tube welding is also employed to connect systems. The tubing diameters generally range from about 1/8” ID to 3/4” ID.

[0068] Most of this description focuses on single-use bioreactors, but the principals apply generically to any of the aforementioned single-use equipment used in bioprocessing, both in the upstream processing (USP) and downstream processing (DSP) arenas. Examples of USP units include mixers and bioreactors, and examples of DSP tools include chromatography assemblies and filtration skids which may use films similar to those used in USP. DSP tools may implement single-use sensors to replace traditional sensors and/or enable new additional analytical capability. “Smart” sensors for DSP and USP tools may have the capability of being pre-calibrated and gamma- or beta-irradiation sterilized along with the bio-process vessel itself.

[0069] Although a number of different styles of single-use bioreactors have been conceived and introduced into the marketplace, two types currently predominate: the “pillow” or “rocker” bag and the stirred tank. The first type of single-use bioreactor to become commercially popular is generally referred to as the pillow or rocker bag style, and is described, for example, in US Patent No. 6,190,913 the teaching of which is incorporated herein by this reference in its entirety. The pillow or rocker type of

single-use bioreactor utilizes a wave motion induced by movement of a bag support platform which generally rocks about a single axis to both mix and sparge (aerate) the contents of the bioreactor. Another of disposable bioreactor is a single-use implementation of the traditional (e.g.: stainless steel and/or glass) stirred tank reactor and utilizes an impeller and a sparger just as its traditional counterpart. The single-use stirred tank implementations include single-use polymeric hard shell bioreactors that functionally imitate small scale glass vessels, and also larger scale single-use versions that generally utilize plastic liner bags that fit inside rigid containers which hold the agitation motors etc. (e.g., US Patent No. 7,384,783 the teaching of which is incorporated herein in its entirety by this reference). The larger liner bags are typically constructed of multi-layer film laminates that also utilize some form of low or ultra-low density polyethylene (LDPE or ULDPE), ethylene vinyl-acetate (EVA), or similar material generally considered to be inert for the contact layer. The liner type single-use bioreactor vessels (bioreactors or mixers or holding cells for liquids) can be constructed from a variety of different polymeric materials, but as mentioned above, are constructed with an inner layer (i.e., the bag surface which is in contact with the aqueous growth medium) made of LDPE or EVA copolymers. Other materials sometimes used in the construction of the single-use bioreactor vessels include but are not limited to high density polyethylene (HDPE) and Kevlar (Poly-paraphenylene terephthalamide). By way of example, Figure 1 shows the construction of the CX-14 film used by Thermo Fisher Scientific for flexible bioreactor vessels. Figure 1 is obtained from *Thermo Scientific Hyclone BPC Products and Capabilities 2008/2009*. The figure shows Thermo Fisher CX-14 film where the layer in contact with the bio process liquid is A1 (low density polyethylene, 10.4 mil thick) followed by layer A2 (a 0.9 mil thick "tie layer" which bonds A1 and A3), and layer A3 (Ethylene- vinyl alcohol copolymer "EVOH", 1.0 mil thick), and layer A4 (another 0.9 mil thick "tie layer" which bonds A3 and A5), and finally A5 (polyester, 0.8 mil thick).

[0070] While single-use bioreactor bags and single-use bioreactor vessels are enjoying popularity, all single use bioprocess vessels (e.g. single-use bioreactors, single-use mixing vessels, single-use liquid holding/storage vessels) are in general seeing increasing market acceptance. To date one major issue has been the lack of robust, single-use sensors that can be readily and reliably integrated into the single-

use bioprocess vessel (e.g.: including but not limited to a bioreactor or mixer). By robust, we mean accurate; gamma, beta, or x-ray radiation stable; and capable of being used for real time sensing (real time within the speeds or time responses required for bio-processing) e.g. providing samples at 1 Hz (or fractions of a Hz to a
5 few Hz) for biological process monitoring and/or control for at least 21 days without significant drift in any 24 hour period.

[0071] Single use sensors are generally introduced into these larger stirred single-use bioreactors through lateral ports or simply attached to an interior surface. Currently, a popular way to introduce optical single-use sensors into single-use
10 vessels is through the use of lateral ports. These ports rely on either a “friction fit” (surface to surface contact between their surface area and the single use sensor without a bonding agent to retain the single-use sensor element) or a classic-o-ring. The ports can be constructed from a rigid base and a flexible piece of tubing, or an entirely rigid structure which then relies on o-rings to maintain a seal, or the ports can
15 be constructed entirely from a flexible material (see, e.g., US 2009/0126515 A1, which is incorporated herein by reference in its entirety). Another way single-use sensors are introduced into single-use vessels is to simply adhere them to the innermost surface layer of the vessel and then both illuminate through bag material and collect fluorescent light the same way. Irrespective of how the optical single-use
20 sensors are introduced, they need to be exposed to the contents (therefore inner area) of the vessel in order to measure the analyte concentrations of interest.

[0072] At this point it is valuable to review in detail the process of introducing “peripherals” to a single-use bioreactor. Peripherals here mean anything introduced into the single-use vessel using a port. Typically, a peripheral directly supports a
25 function of the bio-processing vessel in which it is inserted. Common peripherals include but are not limited to single use sensors, filters, tubing, and sampling ports. As can be appreciated, peripherals come in various shapes, sizes, and materials of construction. However, many of them are designed for a single use and for insertion into a bio-processing vessel via a port, which may have a standard configuration. The
30 process of adding peripherals through a port is desirable in both rigid and semi-rigid single-use vessels as well as flexible vessels. The addition of peripherals this way is useful as the materials used for the construction of sensors, filters, etc. are generally

not made of polyethylene and therefore cannot be simply welded or bonded to the vessel in the way the films themselves are welded together. In general there are a very limited number of materials which can thermally bond to polyethylene. Therefore, a method which allows a convenient aseptic connection to be made will be applicable in the construction of single-use bioprocess vessels. The aseptic peripheral assembly apparatus and methods described herein apply generally to any peripheral device for bioprocessing vessels. In other words, the disclosed assemblies are not limited to sensors.

[0073] For a rigid single-use bio-process vessel there are similar material issues and therefore a port; typically a port with an o-ring is utilized to introduce single-use sensors. When starting the construction of a flexible film based single-use vessel, the films are cut or punched according to design drawings and are thermally bonded (e.g. melted together). However, as many of the elements are of dissimilar materials and cannot be readily bonded together, therefore the single-use industry has almost universally adopted the use of ports or flanges in order to add tubing, sensors, sampling sites etc. Figure 2 shows a picture of a flexible port 21 used by Thermo Fisher to introduce peripherals. It is instructive to note that the base 25 of this port 21 is thermally bonded to the inner layer of the flexible bag 26. The probe, 24, is inserted through the port body, 23, and a seal is made by a molded in flange, 22, which replaces an o-ring. Typically the ports that are welded to the bags are also some form of polyethylene and can be constructed in flexible formats or rigid formats depending on the exact application. As the ports are polyethylene based or compatible material, they can be thermally welded to the film material or over-molded on a flexible single-use vessel. Other vendors utilize rigid ports with a similar overall design but utilizing one or more physical o-rings make the seal.

[0074] Peripherals like the vent filters for example are added by connected tubing between the hose-barb ends of the rigid ports and the vent filters. Tubing is often very difficult to slip over the hose barb and requires a lubricant, or more specifically a substance to temporarily change the surface tension such that the tubing can be slipped over the hose barbs. The tubing is then typically secured with 2 tie-wraps that are put on in opposite directions to ensure that the tubing is not pinched and therefore forms a hermetic seal. The substances used to provide the temporary change in

surface tension allowing the construction to occur is typically isopropyl alcohol and water (70%/30%) mix or pure isopropanol. Pure isopropanol is typically viewed less by manufacturing organizations as it is flammable, must be transported on the ground (e.g. trucking) and the vapors can be dangerous making it difficult to deal with in a controlled environment. The use of isopropanol water mixes and/or isopropanol in the construction of single-use bioprocess containers is a standard and ubiquitous practice. There is often a sizable accumulation (tens of milliliters or more) of these substances inside the container when the construction process is complete. As single-use (flexible or not) vessels are then double wrapped within 2 or more bags for use in cGMP (Good Manufacturing Practice) qualified applications, the bags are clearly hermetically sealed and the isopropanol/water mix or isopropanol is locked into the vessel during the gamma sterilization process.

[0075] The subsequent gamma sterilization of the closed container with these substances locked inside creates other more chemically active substances. For example, when water is exposed to Gamma radiation (Gamma Hydrolysis) it is broken down and forms hydrogen, hydroxyl radicals, and H_2O_2 (hydrogen peroxide) and peroxide radicals (LaVerne, J.A., Radiation Research 153, 196-200, (2000)). It is also clear that the exposure of isopropanol (C_3H_8O or C_3H_7OH) to Gamma radiation will lead to an even greater formation of highly reactive $OH^\cdot$ radicals (J. Environmental Eng. Management, 20/30, 151-156 (2010)). This means that there are reactive agents (e.g.: solvents) inside the single-use vessel that were never considered, let alone tested in the USP Class VI (United States Pharmacopia) testing regimen or considered by the BPSA, Bio-processing Systems Alliance, (bpsalliance.org) sub-committees on single-use vessels or sensors. Neither H_2O_2 nor the post gamma isopropanol compounds are conducive to cell growth and are likely detrimental to any active element (e.g. an opto-chemical sensor) inside the bioprocess container.

[0076] These reactive compounds have implications to sensing elements that are introduced into the single-use vessels. Potentially detrimental reactive compounds are formed in all or the majority of the currently employed single-use vessels where water vapor, oxygen, and plastics are present. While the concentration of these reactive compounds created in each material and how these levels scale with surface area and

content (e.g. water, isopropanol, etc.) may not be known, clearly the problem exists in many contexts.

[0077] Most, if not all, single-use components for bioprocessing are currently sterilized using Gamma radiation or Beta radiation. The requirements for sterilization are in part mandated by the International Standards Organization (ISO) release number 11137-2, which is incorporated herein by reference in its entirety, (ISO 11137: iso.org/iso/catalogue_detail.htm?csnumber=51238).

[0078] This standard stipulates the radiation level and requires a particular reduction in the number of colony forming units (CFU) of bacteria. Typical levels of gamma radiation for cGMP manufacturing applications are 25 kGy to 40 kGy and statistical studies on the number of CFU's performed each quarter. However, many in the bioprocessing industry are calling for, or have called for, higher levels of gamma radiation in order to be certain that the level of bacteria and adventitious agents is low enough. The calls for increased levels of gamma radiation have often been for assemblies of single-use components so that there is increased margin on meeting standards such as ISO 11137-2 mentioned above.

[0079] While Gamma radiation is a very convenient and effective method of sterilization especially for single-use components in the pharmaceutical, medical, and biotech fields, it has quite a few unintended side effects aside from those mentioned above. Even at 25 kGy there are many fundamental deleterious effects on the materials used to construct single-use components. These include but are not limited to cross-linking of polymers so that they become brittle and can break or leak during use, the creation of color centers or other material level defects that affect the color (e.g. the absorption spectrum is modified), and fundamental changes in material properties (*Structural Modifications of Gamma Irradiated Polymers: AN FTIR Study, Advances in Applied Science Research*, D. Sinha, 2012, 3, (3):1365-1371, incorporated herein by reference in its entirety). Additionally, recent research has shown that materials like low density polyethylene which were heretofore considered completely safe and "pristine" for applications as containers for materials or for bioreactor liners are not as inert after exposure to gamma radiation as had been previously considered. In fact, a recent publication (*Identification of a Leachable Compound Detrimental to Cell Growth in Single-Use Bioprocess Containers*,

Hammond et al., PDA Journal of Pharmaceutical Science and Technology, Vol. 67, No. 2, March-April 2013, incorporated herein by reference in its entirety) shows without equivocation that the aforementioned CX-514 film exhibits detrimental characteristics to cell growth following exposure to gamma radiation. The paper identifies compounds such as anti-oxidants (e.g. tris(2,4-di-tert-butylphenyl) phosphite) which are present in many formulations of polyethylene as responsible for at least some of the issues noted in the field. The paper identifies byproducts (e.g: (bis(2,4-di-tert-butylphenyl)phosphate (bDtBPP)) of these anti-oxidants that are created by gamma radiation process. The gamma radiation breaks down the anti-oxidant that remains in the LDPE; often the anti-oxidant is not completely consumed during the rolling of the films. Hammond et al. showed that bDtBPP can inhibit cell growth for many lines of commonly used cells in biotech drug development. Another byproduct of the gamma radiation exposure of the anti-oxidant noted in Hammond's paper is Phosphoric Acid. It is clear that depending on the surface area of the vessel and the sensitivity of the active elements introduced, many deleterious effects will be endured. However, even before this paper by Hammond et al. was published, it was commonly known that polyethylene films can give off hydrogen when gamma irradiated and that this could interact with ozone that is also created when oxygen is present during the gamma process. As most bio-process containers/bioreactors are closed vessels that are not under vacuum during the sterilization process it is quite possible for Phosphoric Acid, Hydrogen, ozone, and Hydrogen Peroxide to exist simultaneously in the same volume. It is also likely that aggressive free radicals are also present in the single-use vessels immediately following gamma sterilization depending on the exact materials used and the conditions in the vessel during gamma exposure.

[0080] In addition to the titer or viability of the cells being directly reduced by the presence of free radicals and radiation byproducts, these byproducts can also affect active elements (e.g. sensors) introduced into the single-use bags prior to gamma exposure. In fact, has been noted by many in the field that single-use optical sensors for analytes such as dissolved oxygen (DO), pH, dissolved CO₂ (dCO₂), and others are deleteriously affected by exposure to gamma radiation in single-use bioprocess containers. The gamma exposure affects the sensors via two distinct readily delineated mechanisms. The first is the chain scissioning

(sterigenics.com/crosslinking/crosslinking.htm) of molecules when exposed to the gamma radiation and the second appears to be a surface mediated effect which impacts the spots during construction of the container and during gamma radiation exposure inside the container.

- 5 [0081] With this background information it is instructive to review the usage case for single-use sensors in this arena. The most common currently deployed single-use sensors for analytes in the single-use vessels are optical sensors for measuring dissolved oxygen (DO), pH, and dissolved CO₂ (dCO₂). These quantities are often measured using sensors which are based upon the principles of phase fluorimetry.
- 10 The sensors spots are most commonly constructed using dyes and/or metal inorganic compounds that behave optically like dyes (Lakowicz, *Principles of Fluorescence of Spectroscopy*, 3rd edition, Springer 2006, incorporated herein by reference in its entirety). These substances are coated on small (typically ~ 3-7 mm diameter) optically transparent disks of polyester or mylar or similar inert USP Class VI, ADCF
- 15 materials, and positioned inside the vessel in the area where the analyte in question is to be measured. The coatings are sometimes referred to as "spots."

- [0082] This sensing modality has evolved rapidly over the last 20 years as the telecom revolution provided inexpensive optical sources (LEDs) and detectors. This type of sensing is popular in bio-processing, and more generally in, e.g., medicine and
- 20 biotechnology as it can be either miniaturized, or made non-invasive, or both. It is also important to note that the sensor elements can and are produced in USP Class VI, animal component derived free (ADCF) formats. With advances in electronics and light sources, the use of information gathering in the frequency domain has become an attractive approach to the art of fluorescent sensing. Sensors that utilize the phase
- 25 delay of the fluorescence signal relative to the modulated excitation signal are based on fluorescence lifetime. Phase fluorometric systems work by detecting a change in the phase lag of the emitted fluorescent signal as a function of analyte concentration. In most cases, this approach has been found to be a more efficacious basis for a sensor than monitoring the quenching of fluorescent intensity in the time domain. In general,
- 30 an optical excitation source is modulated at a frequency, f , and the light impinges upon an analyte sensitive dye. The dye re-emits light at a longer wavelength (a fluorescent signal) with the same modulation frequency, but with a delay in phase, as

shown in Figure 3. The phase delay is caused by the fact that the energy levels of the fluorescent material have finite time constants associated with them. In many ways, one can analogize the fluorescent material to a classical electrical low pass filter to understand the origin of the delay. The fluorescent states can be thought of as a capacitor which has a capacitance that is a function of the environment. At a given frequency, the phase of the signal passed by the low pass filter is mediated by the capacitor's value. In a similar way, the phase delay between the excitation signal and the emitted fluorescent signal is a function of the analyte concentration. An example of this delay is represented in Figure 3. (See C. M. McDonagh et al., *Phase fluorimetric dissolved oxygen sensor*, Sensors and Actuators B 74, (2001) 124-130, incorporated herein by reference in its entirety). Figure 3 shows the excitation wave 31 and the emitted or fluorescent wave 32 as well as the phase delay, 33, between the two. The relationship describing the phase delay, ϕ , and its relationship to the modulation frequency, f , and the fluorescent life time τ can be represented by the following equation:

$$\phi = \text{ArcTan}(2 \pi f \tau)$$

where τ will change as the analyte concentration changes, which means that ϕ will also change as the analyte concentration changes. Methods and suitable data processing equipment which allow one to calculate the phase delay between the excitation signal and the fluorescence signal are known to one skilled in the art (Lakowicz, *Principles of Fluorescence of Spectroscopy*, 3rd edition, Springer 2006, incorporated herein by reference in its entirety).

[0083] The construction of phase fluorimetric based sensors has generally favored using fiber-optic based illumination and collection geometries, though this is not the only method. A fiber optic based design is shown in Figure 4. An aspect of the system pictured in Figure 4 is that the excitation light 41 is remotely filtered by 42 and coupled into a fiber 43, while the collected fluorescent signal from 46 that returns through 44 is delivered through a filter 47 to a photo-detector 49 which is also located remotely from the dye 46 and lens 45. This allows a system where the optical sources and coupling all occur at the same location as the data processing electronics. While this simplifies some of the design and implementation issues and allows the use of fiber-optic for both delivery of the excitation light and collection of the fluorescence

signal from a remote location, it is also limiting in several ways. First, the fiber's ability to withstand bending and other mechanical perturbation is limited. Leakage of both illumination light and signal light caused by bending the fiber or fiber bundle results in the excitation light actually impinging on the fluorophore being of lower
5 than optimal power, and loss of the collected fluorescent signal can significantly reduce the signal to noise ratio. The use of multiple fibers or fiber bundles can help, but dramatically increases the cost and complexity of the system. The collection of the fluorescent signal is often the most vexing problem, and the ability of the fiber (or fiber bundle) to collect light is limited so that most systems of this type collect
10 substantially less than 10% of the light emitted by the fluorophore. This usually results in using a much higher intensity excitation beam than required. This is important because the fluorophores are subject to photo-degradation and their useful sensing lifetimes are limited by this fact. The photo-degradation of the fluorophore and even of the host matrix on which the fluorophore is immobilized in is evidenced
15 by drift in the readings and a lower fluorescence efficiency.

[0084] Because of the issue with photo-degradation, another method of constructing a phase fluorimetric sensor using free space optics has gained market acceptance. Free space optics allows for a large increase in the efficiency of collection of the fluorescence emission light and thereby a large decrease in the excitation light. The
20 decrease in the excitation light allows a corresponding decrease in the rate of photo-degradation and therefore a longer sensing lifetime of the fluorophore. This method is described in US Patent Nos. 7,489,402 B2 and 7,824,902 B2 and the teachings of these are incorporated herein, in their entirety, by this reference. Figure 5 shows an example of this type of free space optical phase fluorimetric sensing system. In
25 Figure 5, element 50 is the excitation source (typically an LED), element 51 is a filter to shape the spectrum of 50. 53 is a lens which helps focus the light 52 onto 59B, the fluorophore. Element 54 can be a single-use element on which the fluorophore 59B is mounted in, or it can be the inner wall 55 of the bioprocess container/bioreactor. The emitted fluorescent signal, 56, is sent through a filter 57 which filters out the light not
30 at the wavelengths (color) of the fluorescent signal before it impinges upon a photo-detector 58. The entire optical assembly is typically enclosed in an opaque housing, 59, which blocks ambient light from both the fluorophore and the detector.

[0085] Figure 6 shows the physical platform on which the fluorophore or opto-chemical sensor spot is mounted. Other parameters (e.g. temperature) can be measured through this physical platform or "carrier." The particular carrier shown in Figure 6 is configured to measure dissolved oxygen and pH through the use of opto-chemical sensor spots mounted in the recessed "cups" 62 and 63 and temperature via cup 61 having a 316L electro-polished plate that is molded into the carrier. Figure 7 shows this carrier sealingly affixed to a mounting flange. In Figure 7 the prior art carrier is shown as 72 and the mounting flange is shown as 73. The mounting flange is typically made of some form of low density polyethylene or compatible material so that it can be thermally welded to the inner layer of a flexible bio-process container. This is described in US Patent Application No. 2012/0244609 A1 and the teaching is herein incorporated by reference in its entirety.

[0086] All of the phase fluorimetric optical sensors share a common characteristic, which is that the fluorescent dye that enables the sensing must be inside the bioprocessing vessel. As previously noted, persons of skill in this field of bioprocessing have noted that the sensing spots do not work the same after gamma radiation as before gamma radiation. As mentioned, gamma radiation can chain scission the dyes and matrix including the sensor spot, as it does all materials. The details of the interaction of the gamma radiation with the materials used to construct single-use bioreactors and their contents were not well understood as evidenced by the paper by Hammond et al. Irrespective of this, some effects have been noted and attempts made to overcome them. This is evidenced by the following series of 3 patents that attempt to address the aforementioned issues. These patent applications are all aimed at methodologies to shelter or isolate the sensor spot from the substances created in the single-use vessels during and immediately following gamma radiation. These patent applications (WO 2010/001457 A1, WO 2011/066901) are aimed at minimizing the volume of the compartment the spot is contained in during the gamma process by creating a type of pocket or housing that shelters the spot. While this minimizes the interaction with some of the volatiles created by the gamma sterilization process, it does not eliminate the effects entirely. The third patent application (WO 2011/015270 A1) focuses on coating the spots with a substance that dissolves when the vessel is filled with liquid (e.g. media for cell growth). The patent teaches uses glycerin or glucose as the coating. There is, however, no clear study

showing the degree of protection these coatings offer no evidence that other issues are not created by these coatings during the gamma sterilization process.

[0087] Finally, none of these approaches address the fact that the spots themselves are still subject to chain scissioning by the gamma radiation. The chain scissioning of the dye molecules and the host material in which it is embedded (the spot) is evidenced by the fact that the time response of the sensor spot is generally significantly slower and the phase response of the spots altered during the gamma process. The time response for a pH sensor when gamma radiated at 40 kGy can be significantly slower (2x or more) than before gamma radiation, and the phase response is typically altered such that the usable measuring range is significantly reduced. This change in the phase response of the spot also causes the calibration of the spot to be dramatically different than before gamma radiation and therefore can render the spot useless. Given the ambiguity in the dose level inherent in gamma sterilization, it is exceedingly difficult to pre-calibrate the spots. The typical sterilization range is 25 kGy to 40 kGy (with many using 25 kGy to 50 kGy), and the calibration is significantly different between these two extrema dose levels. Clearly one can try to provide a calibration in the middle of the range and hope that this is representative of the gamma dose. Given that most vendors of single-use vessels try to minimize their sterilization costs by loading large pallets with their product and having them sterilized simultaneously in a chamber, the spread across the pallet in the run is often 25 kGy to 40 kGy or more. This means that some of the single-use vessels might receive close to the minimum dose and some close to the maximum and therefore the single-use sensors are required to work across a large range of gamma radiation values. Unfortunately, this is often very difficult if not impossible.

[0088] A current typical process for implementing and using single-use opto-chemical sensors is outlined in the flow chart shown in Figure 27. In the boxes 27-A, the manufacturing and calibration of the sensor are described; this activity is generally performed in a particulate controlled environment (e.g., a clean room). Typically the opto-chemical sensors (e.g. spots) are produced and either coated onto the surface to be exposed to the analyte or coated onto a material like, but not limited to, polycarbonate and then the spots are punched out of this sheet of material. Either way, some of the sensors, in whatever form is therefore convenient, are calibrated for

their use conditions (e.g. sterilization conditions, analyte details), and the calibration is made available by memory chip, barcode, or simple manual data entry. Once the sensor calibration is complete and verified, the sensors are packaged. This packaging is typically optically opaque and often with 2-3 layers of bagging for entry into cGMP receiving, and clean rooms. In the next box, 27-B, the activities which happen at the single-use bioprocessing vessel vendor's location are described. Typically the opto-chemical sensor spots are positioned and/or oriented in the aforementioned vessel. This can be through a port, through sealingly affixing to the vessel, or by simply adhering the opto-chemical sensor to the inner surface. When the single-use bioprocess vessel construction is completed, the vessels are packaged by double or triple bagging for application in end-user cGMP environments and then packaged in placed in boxes. The boxes are stacked on pallets and sent out for gamma radiation. The size of the pallets and the varying density make reliable and consistent radiation dosing difficult despite the fact that frequently many dosimeters are placed around the pallet during sterilization. Additionally, spatial hot spots in the dosing are common and often difficult to avoid leading to further variability and ambiguity in the radiation dose the opto-chemical sensors receive. This can lead to decreased accuracy of the calibration now associated with the opto-chemical sensors even under ideal conditions. The box, 27-C, describes the end-user actions once the single-use bioprocess vessel is received. The end-user typically unpacks and sets up the vessel, and then either enters the calibration data by scanning, manual entry, or automated readout (e.g. Finesse Solutions' system). The user then performs a 1 point standardization of the sensor or a 2 point re-calibration (Finesse Solutions' system capability). Generally speaking, the 1 point standardization is not sufficient to allow opto-chemical pH sensors to function well enough for use.

[0089] When phase fluorescent sensor chemistry (the spot) is exposed to the gasses or other byproducts created in the single-use vessels during gamma radiation and simultaneously to the gamma radiation itself, the two effects may combine to render the sensor inaccurate or simply not usable. One way to mitigate some of the effects of gamma irradiation is to develop a detailed pre-calibration method and a sophisticated user based calibration scheme. If the conditions in the single-use vessel during the gamma or e-beam sterilization process include significant amounts of water, isopropyl alcohol, air, and/or organic phosphates, the results of the radiation process are sensor

spots which are not functional even with intervention provided by the aforementioned sophisticated calibration algorithms. Even if the sensors are shielded somewhat from the larger environment of the vessel, the fluorescence properties of the spots is still often significantly compromised such that they are not reliable or accurate enough for use in the intended application (e.g.: controlling the single-use vessel for cell growth, buffer preparation, etc.).

[0090] One way around this issue is to provide a method and apparatus for separating the sensor spots and the vessels during their respective sterilization periods. Clearly both the sensor and the single-use vessel must be sterilized and the sterility of both the single-use vessel and the sensor spots must be maintained when the sensors are introduced into the single-use vessel. This requirement can be met with an appropriately designed aseptic connector and the appropriate method of handling of the sensor carrier. Here, "carrier" means the physical element that the sensor elements are mounted on; here "appropriate" means that it is constructed of materials known to not outgas significantly or which can be proven to not outgas significantly during sterilization or other process, including normal end use. These materials include but are not limited to suitable grades (USP Class VI/ISO 1993, animal component derived free, latex free, phthalate free, gamma and e-beam stable) of polycarbonates, polysulfone, Kynar, or co-polyester during either gamma or beta e-beam sterilization and which can be constructed to meet the form factor required for the carrier for the single-use sensor spots. Two such designs are shown in Figures 6, 7, and 8. The carriers are meant to work with free space optical systems and they are detailed in US Patent Application No. 2012/0244609 A1 and US Patent No. 7,824,902 respectively.

[0091] Aseptic connectors such as those described in US Patent No. 3,865,411 (incorporated herein by reference in its entirety) have been used in the bio-processing industry. They have typically been used to make aseptic connections in sets of tubing or anywhere fluid transfer occurs. The basic concepts involve the ability to have two components that can be individually sterilized and maintain their individual sterility before and after being connected. Additionally, when connected they allow for communication between the passageways in the two components. An example from the aforementioned design is shown in Figure 9. In this figure, 91 is a flexible tube or

conduit leading from the receptacle, 92 is an annular flange attached to tubing, 93 is a compressible gasket, 94 is a diaphragm which has a draw tab 95. The conduit 91 must clearly be closed before sterilization and the removable draw tab 95 covers the other opening. Therefore the system can be sterilized and maintain sterility until use.

5 The connectors are used together as shown in Figure 10. Here the units shown in Figure 9 are connected and their respective diaphragms 6 are brought together and the tabs 5 are brought together. As described in US Patent No. 3,865,411, "The assembly of the two fittings is then clamped together by an appropriate mechanical means 7, i.e.: a spring-loaded clamp or a snap-fitting ball, so that the two gaskets are

10 compressed against each with a predetermined amount of pressure to form a tight seal." In practical use, any such implementation that maintains a sterile barrier after the sterilization process can be readily employed in this arena.

[0092] This concept of an aseptic connection may be applied to the problem described here. The separation of the opto-chemical sensor spots and the single-use

15 bio-process vessel allows the single-use vessel to be gamma radiated (or more generally stated – sterilized) according to whatever standards are required to meet ISO 11137-2 and the end-users' sterility requirements. It also allows the opto-chemical sensors to be separately sterilized such that they are not exposed to the contents of the vessel during the sterilization process. Additionally, it allows for the decision of

20 which sensors and how many of each to be used in a process to be decided just prior to the run as opposed to months or years before when the single-use vessel is designed and built.

[0093] An aseptic connector that can be used for this purpose is shown in Figure 13. The opto-chemical or optical sensor spot and its carrier 13-1 are held in the top

25 section, while the bottom section sleeve 13-2 is attached to the single-use bio-processing vessel (not shown). The sensor may be configured or designed to detect an optical response. In some embodiments, an electrochemical sensor, temperature sensor, pH sensor, oxygen sensor, or single use sensor is held in the carrier 13-1. In some embodiments, two or more sensors are held in the carrier 13-1.

30 In various embodiments, an oxygen sensor, pH sensor, and temperature sensor are held in carrier 13-1. If this were a flexible bag it would typically be thermally welded to the inside surface, but clearly can be attached to any surface of a container through

use of welding or suitable adhesives or retention/attachment processes. In a conventional approach, if this were to be welded to a flexible single-use bio-processing vessel, the material on the base of sleeve 13-2 would either have to be compatible with the inner layer of the bag or sealingly affixed to a flange made of such a material. The carrier 13-1 may be disk shaped with the sensor having an exposed sensing surface on the flat side of the carrier 13-1. As described below, in some embodiments, the carrier may be generally sheath shaped with the sensor having an exposed sensing surface on an end of the carrier. In some embodiments, two or more sensors in carrier 13-1 have exposed sensing surfaces on a flat side of the carrier 13-1.

[0094] While the following examples show sensor-containing assemblies, this disclosure is not limited to sensors. Other peripherals such as filter flow tubes, sample ports, etc. may be used in place of sensors in the following examples. Figure 14 shows an aseptic sensor connection assembly for installing a sterilized sensor in a bio-processing vessel via an aseptic vessel connector or aseptic connector 14-3 affixed to the vessel. The Figure depicts a full cross-sectional view of both halves of the aseptic sensor connection assembly prior to the connection being made. Shown is an example of an optical carrier 13-1 as shown in Figure 6. Here the carrier depicted in Figure 13 as 13-1 is labeled as 14-1 and as before, holds the opto-chemical sensors spots and the stainless steel window for temperature sensing.

[0095] The sleeve 14-2 is a substantially tubular part of the top portion of the aseptic connector that encloses the carrier 14-1 and its sealing flange 14-10; this entire assembly holding the sensor will be referred to as the sensor applicator or applicator, having an applicator connector 14-4. The applicator may be composed of a polycarbonate, polysulfone, polyvinylidene fluoride, co-polyester, or a combination of any of these. The sleeve 14-2 may be constructed of rigid material. The applicator may be configured to be removed from the carrier 14-1 and the bio-processing vessel after the carrier is installed in the bio-processing vessel. The lower half or aseptic connector 14-3 is attached to the single-use bio-processing vessel so that the sensor carrier 14-1 can be aseptically connected to the vessel at a later date and processed independently and will be referred to as the aseptic vessel connector or aseptic connector 14-3. The two halves are sterilized independently prior to the time when

they are connected. In some cases, the applicator may be configured to remain on the bio-processing vessel after the carrier is installed in the bio-processing vessel. It should be noted that the parts are all required to be made of gamma, beta, or x-ray stable materials and that all wetted materials will be required to meet ISO 10993/USP Class VI, requirements as well as be animal component derived free, latex free, and phthalate free. The applicator connector 14-4 may be configured to temporarily or permanently connect to the aseptic vessel connector 14-3. The applicator connector 14-4 may be a clip or retaining device which, when the two components (applicator and aseptic vessel connector 14-3), are united by compressing o-rings 14-7, can latch around the ledge or sealing member 14-5 thereby forming a hermetically sealed unit. While the applicator connector 14-4 is depicted as a clip or latch here, any mechanical device element that allows the o-rings or equivalent sealing devices to be engaged and locked together can be utilized such that the applicator connector 14-4 is configured to provide a hermetic seal with the aseptic vessel connector 14-3 while the plunger 14-8 inserts the carrier 14-1 into the aseptic vessel connector 14-3. In some embodiments, the applicator connector 14-4 is an o-ring.

[0096] Once the two halves are hermetically locked together, the temporary removable hermetic sealing tabs 14-6 are removed, simultaneously allowing an opening between the two halves. Prior to being removed, the removable hermetic sealing tabs 14-6 cover a vessel-facing opening of the sleeve 14-2 to maintain the sterilized sensor in aseptic condition prior to installing the sensor in the bio-processing vessel. The hermetic sealing tabs 14-6 may be configured to be removed from the aseptic sensor connection assembly after connecting the applicator to the aseptic vessel connector 14-3 on the bio-processing vessel at a location where the sterilized sensor is to be installed, and before plunging the carrier 14-1 from a position within the sleeve 14-2 to a position engaged with the aseptic vessel connector 14-3. These tabs can be made of a simple film for low pressure situations, or can include a USP Class VI, latex free, phthalate free, ADC plate. The tabs can be coated with an adhesive to allow them to stick together (or any other technique can be employed that allow the surfaces of the tabs to naturally bond) to facilitate their simultaneous removal.

[0097] The opening allows the depression of the plunger 14-8 through 14-2 pushing the O-rings 14-11 of the carrier/flange 14-1/14-10 into the body of aseptic vessel connector 14-3. The plunger 14-8 may include the O-rings 14-11 to form the hermetic seal while the plunging occurs. As these O-rings 14-11 are compressed forming a hermetic seal, the retaining clips 14-13 latch around the ledge or sealing member 14-12 holding carrier 14-1 into aseptic vessel connector 14-3. The sealing member 14-12 is configured to form a leak-tight seal with the aseptic connector 14-3 where the carrier 14-1 is installed. As mentioned above, O-rings 14-11 provide for a hermetic seal as a sealing member 14-12, but any similar means (e.g. gasket, clip, etc.) can be utilized for this purpose. It should also be noted that the O-rings 14-9 maintain a hermetic seal between sleeve 14-2 and plunger 14-8 through the sterilization process and as plunger 14-8 is depressed. The preferred embodiment of the hermetic seal will be maintained without the use of a lubricant or similar material, as minimizing the amount of materials that can be wetted minimizes the risk of any contamination. However, USP Class VI/ISO10993, animal component derived free, latex free, phthalate free materials do exist and can be used to enhance or enable a hermetic seal. For example, silicone products like the materials offered by NuSil (nusal.com/Products/Healthcare/Restricted/Documents/Restricted%20Healthcare%20Materials%20Selection%20Guide.pdf). Clearly using this aseptic connector 14-3, there is no need for a port or other coupling method to introduce the optical carrier of Figure 6 or similarly purposed opto-chemical sensor mounting plate into a single-use bio-processing vessel.

[0098] The process of attaching and deploying the connecting of the applicator and aseptic vessel connector are shown in Figure 15 where C1 shows the pre-sterilized separate parts. C2 shows the two parts connected and locked together. C3 shows the locked parts with the removable hermetic sealing tabs removed so that there is communication between the two halves of the assembly and therefore between the single-use bio-processing vessel and the optical sensor carrier. C4 shows the plunger depressed and the carrier locked into place in the vessel-mounted portion of the aseptic vessel connector. In some embodiments, the sensor is ready to use even with a non-removable applicator such that the assembly remains in the position shown in C4 during operation of the bio-processing vessel. In some embodiments, the applicator is detachable or removable. C5 shows an example including the entire

detached plunger, and applicator where the carrier used to be, and the optical carrier and sensors deployed in the vessel, leaving the sensor ready to use.

[0099] In certain embodiments, the single-use bio-processing vessel will not be filled with fluid or used in such a manner here that is a large (greater than approximately 1 psig) pressure internally. A non-functioning carrier or blank can be constructed that can be utilized instead of the optical carrier with active elements. This blank would be employed to seal the aseptic vessel connector so that the vessel can be filled with liquid without the potential for leakage around the aforementioned removable hermetic sealing tab.

[0100] This concept can be used with other forms of carriers, including the optical sensor carrier shown in Figure 8. In Figure 8, the cylindrical body of a sheath-type carrier, 81, is opaque, and there is an optically transparent lens or window 82 on which the opto-chemical sensor spot is glued or deposited. An optional light shield 84 is used for spots that cannot support an opaque coating. Additionally, a stainless steel plate 83 is molded in and acts as a thermal window through which temperature can be sensed. Of course, other sensor configurations can be used in sheath-type carriers.

[0101] Figure 16 shows a sheath-shaped carrier utilizing a similar aseptic sensor connection assembly as described above, thereby negating the need for a port as shown in Figure 2. The term “sheath-shaped” or “sheath-type” as described herein is defined as a covering that encircles the carrier and optionally affiliated structures before and sometimes during installation. The sheath may be close-fitting with respect to the carrier. Carrier 16-1 of the applicator (referred to above as the “sensor applicator”) holds the sensor, while aseptic vessel connector 16-2 is sealingly affixed to the single-use bioprocessing vessel.

[0102] Figure 17 shows a full cross-sectional view of the sheath-shaped optical carrier as it is adapted for use in an aseptic sensor connection assembly. The aseptic sensor connection assembly here includes a sheath-shaped carrier 17-1 and has a sealing flange or sleeve 17-2, a plunger 17-5, all of which is housed in an applicator 17-3 having an applicator connector 17-7. The sleeve 17-2 is affixed to carrier 17-1 such that there is no leakage around the seal (e.g., leak-tight seal). This seal can be

carried out with epoxy or similar method, or it can be designed into the mold of the sheath style optical carrier 17-1 so that the entire aseptic sensor connection assembly is one piece (monolithic).

[0103] In use, the carrier 17-1 and sleeve 17-2 are inserted into sensor applicator 17-3, while the bottom half 17-4 of aseptic vessel connector 17-8 is sealingly affixed to the single-use bio-processing vessel as discussed above. These operations may be performed at different times and/or by different entities. The openings to applicator 17-3 are hermetically sealed by the plunger 17-5 and its O-rings 17-6 on the top and by one of the removable hermetic sealing tabs 17-9. The top half of the applicator 17-3 is connected to the aseptic vessel connector 17-8 such that the locking mechanism 17-11 engages with the applicator connector 17-7 allowing the O-rings 17-10 to form a hermetic seal pushing the removable hermetic sealing tabs 17-9 together. The removable hermetic sealing tabs 17-9 are removed, providing an opening between the two halves of the assembly. The comments for the utilization of alternative locking mechanism, the removable hermetic sealing tabs, and the seals mentioned in association with Figure 14 apply equally here. The plunger 17-5 is depressed pushing the sensor optical carrier 17-1 and sleeve 17-2 into the bio-processing vessel through aseptic vessel connector 17-8 and allowing the locking mechanisms 17-11 to retain the ridge on the sleeve 17-2. The hermetic seal is maintained by O-rings 17-10 and 17-12. With the optical carrier locked in to the aseptic vessel connector 17-8 on the single-use bio-processing vessel, the top half applicator 17-3 and the plunger 17-5 can be removed by depressing or pinching the locking mechanism on the applicator connector 17-7. In some cases, the applicator 17-3 remains attached to the aseptic vessel connector 17-8 during use of the bio-processing vessel and is not removed.

[0104] In some cases, the peripheral connection assembly may need to be modified from the structure shown in Figures 14-17 to accommodate the size and shape of the peripheral. Further, some peripherals will require a custom carrier that is neither a disk nor a sheath.

[0105] An overview of the process for installing a peripheral is shown in Figure 18 and Figure 19. In Figure 18, D1 shows the top part of the aseptic container with the optical sensor carrier contained inside and sealed and the bottom half which would be attached to a vessel. Both halves have been sterilized. D2 shows the two halves

connected with the sealing tabs still in place. D3 shows the system with the sealing tabs removed allowing communication between the two halves. In Figure 19, D4 shows the plunger depressed and the optical carrier pushed through the vessel side connector with the flange locked into place. In some embodiments, the sensor is ready to use when installed with a non-removable applicator such that the assembly remains in the position shown in D4 during operation of the bio-processing vessel. In other embodiments, the applicator is detachable or removable. D5 shows an example where the optical carrier side connector and plunger are removed leaving the system ready to use. Again, it should be noted that though it is not necessary, the aseptic connector system shown here may be used before the single-use bioprocessing vessel is filled with any liquid or under pressure. With careful design of the sealing tabs and how they are retained, higher pressures can be accommodated.

[0106] Figures 20A-20C show views of an example aseptic sensor connection assembly for installing a sterilized sensor in a bio-processing vessel (not shown) via an aseptic vessel connector 20-8 affixed to the vessel. The Figures depict full cross-sectional views of both halves of the aseptic sensor connection assembly prior to the carrier, with sensor, being inserted. The aseptic connector 20-8 is affixed to the bio-processing vessel which may be separately shipped to, for example, a customer or a downstream distributor.

[0107] The aseptic sensor connection assembly includes a sheath-shaped carrier 20-1 which includes a sealing member 20-2 and is configured to carry a sensor. In some embodiments, the sensor inside the carrier 20-1 is affixed to the carrier 20-1. The carrier 20-1 is contained in an applicator 20-3 including an applicator sleeve 20-13 and an applicator connector 20-7. The sealing member 20-2 is affixed to carrier 20-1 such that there is no leakage around the seal (e.g., it forms a leak-tight seal when installed in the bio-processing vessel). This seal between elements 20-2 and 20-1 can be formed with epoxy or similar adhesive component, or it can be designed into the mold of the sheath carrier 20-1 so that the carrier 20-1 with sealing member 20-2 is one piece (monolithic). The carrier 20-1 and all of its components, including the sealing member 20-2 are configured to move together axially with respect to the applicator sleeve 20-13 to insert the carrier 20-1 into the bio-processing vessel. Prior to installation in the bio-processing vessel, the carrier 20-1 and applicator 20-3 may

be gamma or beta-radiated at a low level in an oxygen and moisture free container to sterilize the components. In certain embodiments, the radiation is provided at a relatively low dose such as about 15 kGy or lower.

[0108] The carrier 20-1 and its associated sealing member 20-2 may be inserted into applicator 20-3, while the aseptic vessel connector 20-8 is sealingly affixed to the single-use bio-processing vessel (not shown) as discussed above. These operations may be performed at different times and/or by different entities. In some embodiments, the aseptic connector 20-8 is assembled by a manufacturer prior to packaging and sterilization. Figure 20B shows a depiction of the assembly prior to inserting the carrier. Viewed from outside and prior to mating with the aseptic connector of the bio-processing vessel, the assembly includes carrier 20-1, which includes sealing member 20-2 having clips 20-11 for mechanically engaging with apertures 20-12 on applicator sleeve 20-13. Applicator sleeve 20-13 is part of applicator 20-3, which further includes applicator connector 20-7. During fabrication, an opening on the vessel-facing side of applicator 20-3 is hermetically sealed by removable hermetic sealing tabs 20-9 (shown in Figures 20B and 20C). Hermetic sealing tabs 20-9 may also cover the opening to the aseptic connector 20-8. In the depicted embodiment, the opening to applicator 20-3 is adjacent to the applicator connector 20-7. A view from the bio-processing vessel toward the assembly is shown in Figure 20C. As can be seen through hermetic sealing tab 20-9, the carrier 20-1 with sensor sits inside the applicator 20-3 prior to connecting the applicator connector 20-7 and aseptic connector 20-8. After installation, o-rings 20-10 (shown in Figure 20A) form a leak tight seal between the applicator sleeve 20-13 and sealing member 20-2 associated with the carrier.

[0109] Figure 21 shows an example of the assembly whereby the carrier 20-1 is inserted into the bio-processing vessel via the aseptic vessel connector 20-8. In some embodiments, the carrier 20-1 including the sensor 20-5 and sealing member 20-2 is first inserted into the applicator sleeve 20-13. To insert the carrier 20-1, the applicator connector 20-7 may then be aligned with the aseptic vessel connector 20-8. The removable hermetic sealing tabs (not shown) are removed, providing an opening between the two halves of the assembly. A clamp 20-4 may be used to hold the applicator connector 20-7 and the aseptic vessel connector 20-8 in place at least while

the carrier 20-1 is installed in the bioprocessing vessel. For example, as shown in Figure 21, ring clamp 20-4 holds the assembly together. In some embodiments, the applicator connector 20-7 may include an adhesive ring and antiseptic, which may keep the inside of the connector 20-7 sterile, while the clamp 20-4 is used to provide mechanical support for maintaining the seal. In some embodiments, a clamp is not used to hold the applicator connector 20-7 and the aseptic vessel connector 20-8 together while the carrier 20-1 is installed. In some embodiments employing hermetic sealing tabs, the tabs are removed before a clamp 20-4 is applied. In other embodiments, the sealing tabs are removed after a clamp 20-4 is applied or after an adhesive connection is made, but such clamp or connection must permit sufficient flexibility to permit removal of the tabs while maintaining an aseptic condition.

[0110] The utilization of alternative locking mechanisms, removable hermetic sealing tabs, and seals mentioned in association with Figures 14 and 17 apply equally here. The carrier 20-1 with its sealing member 20-2 are inserted into the bio-processing vessel through applicator sleeve 20-13, applicator connector 20-7, and aseptic vessel connector 20-8 such that the clips 20-11 on the sealing member 20-2 and the applicator sleeve 20-13 are coupled to apertures 20-12 to form a mechanical coupling. Of course, other mechanisms such as clamps, pins, tabs, friction couplings, and the like may be used in place of the clips 20-11 and apertures 20-12. In the depicted embodiment, the hermetic seal is established by O-rings 20-10. The carrier 20-1 is inserted into the bio-processing vessel by pushing it axially downwards such that the carrier 20-1 and sealing member 20-2 move relative to applicator sleeve 20-13 until the clips 20-11 at sealing member 20-2 and the apertures 20-12 at the top of the applicator sleeve 20-13 match and lock to form a mechanical seal.

[0111] At the same time, sealing member 20-2 in the carrier 20-1 will form a leak-tight seal using O-rings 20-10 and the sensor 20-5 is inserted into the bio-processing vessel. With the optical carrier 20-1 locked to the applicator 20-3 connected to the aseptic vessel connector 20-8 on the single-use bio-processing vessel via applicator connector 20-7, cables and other external components may be inserted into the carrier 20-1 to be coupled with the sensor 20-5, which is now within the bio-processing vessel. Alternatively, the cables and/or other external components are applied to the carrier 20-1 prior to installation in the bio-processing vessel. During operation of the

bio-processing vessel, in some embodiments, the aseptic peripheral connection assembly remains in the position shown in Figure 21 such that none of the components are removed until the peripheral or sensor 20-5 needs to be replaced.

5 [0112] Figure 22 shows the clamp 20-4 holding the applicator connector 20-7 and applicator 20-3 to hold the applicator connector 20-7 and the aseptic connector 20-8 together. Example clamps include spring-loaded clamp or a snap-fitting ball, C-clamps, axial clamps, snaps, and ring-clamps. Note in this figure, carrier 20-1 is fully inserted through the aseptic connector 20-7, as depicted both at the top and bottom of the figure, with the sealing member 20-2 inside the applicator 20-3.

10 [0113] Disclosed embodiments are suitable for convenient and efficient delivery to the user. The fabrication, sterilization, and insertion of the applicator assembly shown in Figures 20A-22 may include the following operations. As described above with respect to Figures 20A-20C, the aseptic connector is affixed to a bioreactor. Separately, the applicator assembly is fabricated, sterilized, and separately delivered
15 to an end user site. For example, the applicator may include the applicator sleeve, and the applicator connector; the hermetic sealing tab sealing the opening of the applicator connector; and the carrier including the sealing member and the sensor. The applicator may then be sterilized using any suitable approved method of sterilization, such as by gamma irradiation at a particular dose level (e.g., about 15 kGy or lower).
20 The sterilized applicator may then be delivered to an end user's site, whereby the user may attach the sterilized applicator to the applicator connector by aligning the applicator connector to the aseptic connector, removing the hermetic sealing tabs, applying a clamp around the flange portion of the aseptic connector and applicator connector to form the aseptic peripheral connection assembly, and plunging the
25 carrier into the bio-processing vessel while forming a leak-tight seal. The system may then be ready to use by operating the bio-processing vessel and obtaining sensor readings.

[0114] In some cases, the peripheral connection assembly may need to be modified from the structure shown in Figures 14-22 to accommodate the size and shape of the
30 peripheral. Further, some peripherals will require a custom carrier that is neither a disk nor a sheath.

[0115] As mentioned above, there are a multitude of ways to create an aseptic connection between the carrier/flange and the aseptic vessel connector. Figure 23 shows a method where the retaining clips in Figure 14 (14-13) have been replaced with a set of threads 23-4. The sensor carrier and flange assembly 23-1/23-2 are held
5 into the aseptic vessel connector 23-3 by threading or screwing the components together. The plunger (not shown) in this case would be used to screw 23-1/23-2 into 23-3. Care needs to be taken with the materials so post sterilization dimensions still allow for a hermetic seal to be created by the threaded components. A USP Class
10 VI/IS10993 animal component derived free, latex free, phthalate free gel or adhesive can also be applied to the threads 23-4 during construction of the parts and prior to mating of the parts to enhance the hermeticity of the seal. In this system, o-rings are the primary method used to ensure the sealing, but can also be replaced with gaskets or alternative methods of creating a seal between surfaces.

[0116] In Figure 24 another variation of the connection method is shown where the
15 sensor carrier and sealing flange 24-1/24-2 are shown connected and with a textured edge 24-4. This edge is press fit into the edge 24-5 on the aseptic vessel connector 24-3 and retains the sensor carrier and flange 24-1/24-2 in the aseptic vessel connector. As with the previous methods of making a seal between the sensor carrier and flange 24-1/24-2, the aseptic vessel connector o-rings are shown here but other
20 methods can be employed.

[0117] In Figure 25 a bayonet mounting method of retaining the sensor carrier and flange into the aseptic vessel connector is similar to what is used in many cameras with changeable lenses. In Figure 25 the system with the connection made is shown. Figure 26 shows a cross-sectional view of the components before the sensor
25 carrier/flange (26-1/26-2) assembly is retained in the aseptic vessel connector assembly 26-3. The flange 26-2 has mounting segments 26-4 which are inserted into the openings 26-5 in the aseptic vessel connector. The flange with the mounting segment would be rotated clockwise as seen in Figure 26 such that the mounting segment 26-4 is retained in the groove 26-6. There can be several such segments and
30 grooves around the perimeter of the flange and aseptic vessel connector respectively. The mounting segments will be friction fit into the groove such that the sensor carrier

and flange are retained in the aseptic vessel connector with the o-rings maintain the seal as described before.

[0118] Other variations on this design can be considered including, but not limited to a retaining clips, a continuously retaining clip etc.

5 [0119] Above a system of separating the sensors and sensor carrier from the single-use bioprocessing vessel has been detailed. In this case the single-use vessel can be gamma sterilized exactly as before without such an aseptic vessel connector attached. If the carrier side aseptic connector is constructed using truly inert materials as described before and the assembly is subjected to gamma radiation in a bag or
10 container constructed of equally inert materials without liquids present, the assembly can be gamma radiated with minimized effect. Specifically, there should be very few free radicals, hydrogen peroxide, organic phosphates or other substances which can affect the performance of the opto-chemical sensor spots. However, there is still the effect of gamma radiation and its potential to compromise the performance of the
15 sensor spots.

[0120] While ISO 11137 describes the requirements to minimize colony forming units (CFU) of bacteria and adventitious agents with gamma radiation, the end goal is simply the reduction in the number of CFU's. Several method of implementing the testing are stipulated in ISO 11137-2 including the number of samples to be tested and
20 how they are to be prepared, and further testing if there are failures based on the given criteria.

[0121] We describe here a method for minimizing the required level of gamma radiation, beta radiation, or x-ray radiation required to give the same reduction in CFU's. It has been well documented for the treatment of waste water, hospital
25 maintenance, and general surface disinfection, that ultra-violet (UV) light is very effective at reducing the number of CFUs of bacteria and spores. For example, the US Environmental Protection Agency has issued EPA 815-R-06-007, *Ultraviolet Disinfection Guidance Manual for the Final Long Term 2 Enhanced Surface Water Treatment Rule*, which is incorporated herein by reference in its entirety. Figure 11
30 shows a chart giving dosages in fluence (energy per unit area) required for the log reduction in various types of unwanted bacteria etc. Additionally, the guidance

details that the optimal wavelengths of the radiation are between approximately 235 nm and 290 nm with a peak effectiveness around 265 nm. This fluence and wavelength can be provided by pulsed UV lamps as described in the guidance, and by newer technologies such as UV LEDs (e.g. s-ct.com).

5 [0122] Another method for disinfecting and sterilizing is atmospheric plasma based sterilization. A recent publication, "Cold Atmospheric Air Plasma Sterilization against Spores and Other Microorganisms of Clinical Interest", Klampfl et al., Applied and Environmental Microbiology, 78, 15 p. 5077, Aug. 2012, which is incorporated herein by reference in its entirety, describes a study that showed a
10 substantial reduction in CFU's of both Gram-negative and Gram-positive bacteria as well as types of fungus after exposure of samples to a cold atmospheric plasma (CAP). This means that the plasma was created using air at essentially room temperature (under 40°C); specifically not at high temperature (120°C) and not with toxic gases like formaldehyde or ethylene oxide. Figure 12 shows the log reduction in
15 various strains of bacteria and fungus created by 60 seconds or less of exposure to CAP.

[0123] If the number of CFU's on the optical carrier and associated aseptic connector can be minimized before being packaged, a lower dose of gamma, beta, or x-ray radiation will be used to meet the acceptable limit on CFU's for use in the
20 pharmaceutical production arena. A method for preparing and installing a sensor while reducing the number of CFU's is outlined in Figure 28. In box 28-A, the activities carried out during the manufacturing and packaging of the sensors are described in detail. In certain embodiments, all of the activities are performed in a class 10,000 or better clean room. The overall bacterial and adventitious agent level
25 in the clean room will be minimized by vigilant cleaning and exposure to UV light. In certain embodiments, all of the components (except, in general, the opto-chemical sensor materials which can be rapidly degraded by exposure to intense UV radiation) in the aseptic connection system need to be UV cleaned with UV radiation at ≥ 20 mJ/cm² and/or plasma cleaned as described in the references provided. The opto-
30 chemical sensor spots are affixed to the carrier. The calibration has assumed to have already been carried out and this calibration information is encoded in the associated memory chip or provided in some other fashion to be available to the end-user. The

carrier is assembled with the sensor applicator with its aseptic connection mechanism and plunger. In some embodiments, the carrier is assembled with the applicator connector and sleeve without a plunger. The hermetic sealing tabs are sealed onto the applicator connector prior to sterilization. In some embodiments, the applicator connector includes the hermetic sealing tabs already sealed onto the applicator connector. This entire sensor assembly (e.g: carrier, connector, and, in some embodiments, plunger) and the aseptic vessel connector portion are also plasma cleaned, if necessary, to further reduce the number of colony forming units. Both of the matching components/assemblies are then vacuum-packaged in the clean room as per the requirements of the end-user. In certain embodiments, the base packaging material is not light transmitting and meets all the aforementioned requirements of being USP Class VI/ISO10993, animal component derived free, latex free, phthalate free, and sterilizing radiation (e.g.: gamma, beta, x-ray) stable. Stable here means that it does not release agents that have a deleterious effect on the sensors or are toxic, and it maintains the integrity as packaging materials. Both components/assemblies are now sent for sterilization. A suitable source is e-beam (Beta) radiation, which can be easily controlled. In certain embodiments, the dose will be $\leq$ about 15 kGy as the sensors are minimally affected by this level. However, if higher doses are required, this method of preparation and radiation may minimize the required dose and will decouple of the radiation/sterilization process of the opto-chemical sensor from radiation/sterilization process of the single-use bioprocess vessel. This also eliminates the exposure of the opto-chemical sensor to the free radicals and chemicals produced during the radiation of the single-use bioprocess vessel. If these free radicals are present long (e.g., days to weeks) after the sterilization of the single-use bioprocess vessel, the vessels can be flushed with air, nitrogen, or even water for injection before the aseptic connection between the two components. Box 28-B shows the actions of the end-user who receives both the single-use bioprocess container and the sensor applicator assembly. The single-use bioprocess vessel will have the aseptic bioprocess vessel connector attached during its construction. For a flexible film based single-use bioprocess vessel this would likely have entailed having the aseptic vessel connector side be equipped with a plate or flange that was welded to the inner layer during its construction. The end user will connect the sensor applicator as described before, but can orient the sensor if desired. Once the aseptic

connection is made, the user can perform a 1 point standardization or a 2 point calibration as desired and according to the requirements of the sensor being used.

5 [0124] Depending on the rates of infection (base level of CFU's) found on the components, the UV radiation may not be necessary or not at the above stipulated levels.

[0125] This hermetically packaged sensor and carrier can then be exposed to the minimal amount of gamma, beta, or x-ray radiation required to meet the number of CFU as stipulated by ISO 11137. In certain embodiments, a goal is to use ≤ 15 kGy of radiation to ensure compliance with ISO 11137, at which dose level the effect on
10 opto-chemical sensors may be negligible. Generally, speaking there exist several possible combinations of treatment that can allow compliance with ISO 11137. These include:

1. Gamma/beta/x-ray sterilization of the sealed aseptic connector packages at ≥ 25 kGy
- 15 2. Use of UV Sterilization of the optically insensitive components as described above
3. Use of cold atmospheric plasma of the assemblies in the package or before packaging
4. Exposure of components to ethylene oxide sterilizing compound before
20 packaging if the sensor components are not deleteriously affected by this
5. Any combination of 2, 3, 4, or any combination including 1 but without the stipulation that the sterilizing radiation exceed 25 kGy.

[0126] The sensor carriers from Finesse Solutions, Inc. already come with a base
25 level of calibration already programmed into the attached memory chip. This calibration is developed by detailed testing of similar opto-chemical sensors after running them through the exact same process and often contemporaneously. This calibration may be applied to the balance of the carriers and sensors in the lot that is

being processed. This calibration may be applied prior to UV or plasma sterilization and carriers, as they need to be connected to a programming device.

[0127] With the sterilization of the single-use vessel and the sensor(s) and its carrier separated, the process for installing of the sensor by the end-user is different than when the two components are sterilized together. In the case of separated sterilization, the sensors may be put on their carriers and processed as described above. The packaged sensors and carriers may be sent out for gamma, e-beam, or x-ray sterilization at ~ 15kGy and stored by the vendor (e.g.: Finesse Solutions, Inc.) or sent out for sterilization as orders arrive. As the sensor applicator assembly may be dimensionally small (e.g., < ~15 cm) the use of Beta radiation/e-beam is applicable. All of the components can be placed flat, once component deep, such that the sterilizing radiation can be delivered quickly, uniformly, and consistently. Meanwhile, the end-user receives their single-use bio-process vessel from their preferred vendor already gamma irradiated according to that vendor's standards. The single-use bioprocess vessel aseptic connector is already in-place and sterilized with the single-use bio-processing vessel. The single-use vessel is then set-up according to the vendor's instructions. At this point, before filling the vessel with media, the sensor, applicator, and aseptic single-use vessel connector are connected to the vessel as described above. The media is then added and the single-use bio-processing vessel is prepared for initial use. As part of the set-up it is typical for offline samples to be taken so that the dissolved oxygen and pH probes can be standardized to the probes that the process was developed with. At this point, a one-point standardization against the offline standard is typically performed and the sensors are calibrated and ready to use. The temperature is often checked or standardized in a similar manner against a known temperature standard using a port designed into the bag for this purpose.

[0128] For example, the carrier for the spots can be "pre-sterilized" using a plasma cleaner and/or using ultra-violet light before and/or after attaching the spots. The ultra-violet light can be supplied by a variety of different high pressure lamps and/or UV LEDs as mentioned above. The aseptic connector including the carrier may be sterilized by any suitable method. For example, the sensor may be sterilized using ethylene oxide (ETO). The choice of sterilization procedure depends on the spots'

sensitivity to these sterilization processes. Pre-sterilization may reduce the required gamma or beta dose levels while ensuring that the level of CFU's is acceptable to meet the needs for bioprocessing or similar activity. Typical gamma sterilization facilities use CO₆₀ to provide the gamma radiation, and providing a uniform dose across a palette is not possible, nor is delivering a precise dose.

[0129] Due to this fact, x-ray (e.g. Rhodotron), or beta radiation may be a suitable alternative to gamma radiation for sterilization of the optical spot and carrier. As mentioned before, beta radiation does not penetrate as far into materials as gamma radiation does, but an accelerator source typically allows for far less ambiguity in dosing than gamma sources. Due to the lack of penetration depth, beta radiation is rarely used for sterilizing single-use vessels in a commercial setting. It simply is not economical to sterilize each individual single-use vessel separately and if the container exceeds 12"-18", the beta radiation will not thoroughly or uniformly sterilize the vessel.

[0130] A typical free space optical "carrier" or component for the spots would not reach a size that cannot be uniformly sterilized by with beta radiation, and they can be packaged such that they are in a thin (less than 5 inches) layer which can be quickly and economically sterilized. Additionally, the dose of radiation required to meet ISO11137-2 standards for the reduction in the number of colony forming units (CFU's) of bacteria can be dramatically reduced by adherence to the correct process. An example process allowing this reduction carries out all of the work in a 1000 clean room or better and pre-sterilizing the carrier using both Ultra Violet (UV) radiation and plasma cleaning. UV radiation is widely documented and is widely used for disinfection (e.g.: *Ultraviolet Disinfection Guidance Manual For The Final Long Term 2 Enhanced Surface Water Treatment Rule*, US EPA, Office of Water (4601), EPA 815-R-06-007, incorporated herein by reference). Plasma cleaning has also been found to be an effective agent for sterilization (e.g.: *Cold Atmospheric Air Plasma Sterilization against Spores and Other Microorganisms of Clinical Interest*, Klampfl et al. Applied and Environmental Microbiology, 78, 15, 5077, Aug 2012, previously incorporated by reference).

[0131] We have found experimentally that if the spots are affixed to a carrier made of a suitable material (e.g., a material that does not outgas even when exposed to

gamma radiation) and packaged in a suitable material, radiation and the dose of sterilizing radiation can be maintained at $\leq$ about 15 kGy there is little to no damage done to the spots. At this level of radiation, the change in the phase response of the spots is minimized and is very repeatable. In our experiments using suitable materials
5 we have found that there is no outgassing during the sterilization process or at least none that affects the spots. Therefore many of the negative effects that sensors endure during gamma sterilization can be avoided.

[0132] The pre-sterilization of the carrier can be carried out with UV light, for example with a pulsed Xenon lamp or other source that has the peak of its radiation
10 between 254 nm and 280 nm with sufficient intensity. Other sources as mentioned above are high power UV LEDs, and other high pressure metal vapor lamps (e.g. Mercury) or laser sources.

[0133] Additional applications of this type of aseptic connector can be found for radiation (gamma, beta, x-ray) sensitive electronics. Many types of digital (and
15 analog) chips/circuits cannot be utilized in single-use bioprocess vessels due to their incompatibility with the sterilization process. Specifically, most integrated circuits are not compatible with sterilization by the aforementioned ionizing radiation. If the circuits, potentially, containing other types of sensors (e.g. pressure, temperature) which have their signal conditioned by integrated circuit electronics are desired to be
20 implemented on a single-use bio-process vessel, they can also be mounted on a carrier and sterilized with UV radiation, atmospheric plasma, or chemical processing (e.g. ethylene oxide). Similarly, the peripheral components mentioned earlier such as sampling ports, temperature sensing wells, or additional spargers can all be added to a single-use bioprocessing vessel post sterilization this way, leading to a far more
25 flexible component for the end user.

CONCLUSION

[0134] Although the foregoing embodiments have been described in some detail for purposes of clarity of understanding, it will be apparent that certain changes and modifications may be practiced within the scope of the appended claims. It should be
30 noted that there are many alternative ways of implementing the processes, systems, and apparatus of the present embodiments. Accordingly, the present embodiments are

to be considered as illustrative and not restrictive, and the embodiments are not to be limited to the details given herein.

CLAIMS

What is claimed is:

1. An aseptic peripheral connection assembly for installing a sterilized peripheral in a bio-processing component via an aseptic connector affixed to the component, the peripheral connection assembly comprising:
 - a. a carrier disposed within an applicator sleeve, the carrier comprising the sterilized peripheral and a sealing member configured to form a leak-tight seal with the aseptic connector on the bio-processing component at a location where the carrier is to be installed, the carrier configured to be inserted from a position within the applicator sleeve to a position engaged with the aseptic connector to form the leak-tight seal;
 - b. an applicator comprising the applicator sleeve and an applicator connector, the applicator connector comprising a component-facing opening configured to connect to the aseptic connector on the bio-processing component; and
 - c. a removable hermetic sealing tab covering the component-facing opening of the applicator sleeve to maintain the sterilized peripheral in aseptic condition prior to installation in the bio-processing component.
2. The aseptic peripheral connection assembly of claim 1, wherein the bio-processing component is a self-contained container or a flow path.
3. The aseptic peripheral connection assembly of claim 1, wherein the bio-processing component is a bioreactor or a filter flow path.
4. The aseptic peripheral connection assembly of claim 1, wherein the peripheral is a single use sensor.
5. The aseptic peripheral connection assembly of claim 1, wherein the carrier comprises two or more sensors.

6. The aseptic peripheral connection assembly of claim 1, wherein the carrier comprises one or more sensors having exposed sensing surfaces on a side of the carrier.
- 5 7. The aseptic peripheral connection assembly of claim 1, wherein the carrier is generally sheath shaped and comprises a sensor having an exposed sensing surface on an end of the carrier.
8. The aseptic peripheral connection assembly of claim 1, wherein the applicator comprises a clip or a ledge configured to engage with an aperture on the applicator sleeve.
- 10 9. The aseptic peripheral connection assembly of claim 1, wherein the sealing member comprises an o-ring.
10. The aseptic peripheral connection assembly of claim 1, wherein the applicator sleeve is constructed of a rigid material.
11. The aseptic peripheral connection assembly of claim 1, wherein the applicator sleeve has a tubular shape and has a substantially circular shape that forms a seal with an interior surface of the applicator.
- 15 12. The aseptic peripheral connection assembly of claim 1, further comprising a clamp configured to clamp the applicator connector and the aseptic connector.
13. The aseptic peripheral connection assembly of claim 1, wherein the applicator comprises one or more apertures and the carrier comprises one or more clips configured to be inserted into the one or more apertures when the carrier is engaged with the aseptic connector to form the leak-tight seal.
- 20 14. The aseptic peripheral connection assembly of claim 1, wherein the peripheral is installed in the bio-processing component such that the carrier in its installed position has one end extending through the aseptic connector and within the bio-processing component.
- 25

15. The aseptic peripheral connection assembly of claim 1, wherein the applicator connector is configured to provide a hermetic seal with the aseptic connector while the carrier is inserted into the aseptic connector.
- 5 16. The aseptic peripheral connection assembly of claim 1, wherein the applicator is made from a material selected from the group consisting of polycarbonate, polysulfone, polyvinylidene fluoride, co-polyester, and any combination thereof.
- 10 17. The aseptic peripheral connection assembly of claim 1, wherein the applicator is composed of USP Class VI material that is animal derived component free, latex free, phthalate free, and gamma and e-beam stable.
18. The aseptic peripheral connection assembly of claim 1, wherein the carrier comprises an o-ring for forming a hermetic seal with the applicator sleeve when the carrier is inserted into position with the aseptic connector.
- 15 19. The aseptic peripheral connection assembly of claim 1, wherein the removable hermetic sealing tab is configured to be removed from the aseptic peripheral connection assembly after connecting the applicator to the aseptic connector on the bio-processing component at a location where the sterilized peripheral is to be installed, and before inserting the carrier from a position within the applicator to a position engaged with the aseptic connector.
- 20 20. The aseptic peripheral connection assembly of claim 1, wherein the removable hermetic sealing tab comprises a film or sheet having a thickness of between about 1 mil and 20 mil.
- 25 21. The aseptic peripheral connection assembly of claim 1, wherein the removable hermetic sealing tab comprises a USP Class VI, latex free, phthalate free animal derived component free polymeric plate.
22. The aseptic peripheral connection assembly of claim 1, wherein the removal hermetic sealing tab is coated with an adhesive.

23. A bio-processing component kit comprising:
- a. the aseptic peripheral connection assembly of claim 1; and
 - b. a bio-processing component housing with the aseptic connector affixed.
- 5 24. The bio-processing component kit of claim 23, wherein the bio-processing component housing is a single use bio-processing vessel housing or a flow path.
- 10 25. The bio-processing component kit of claim 23, wherein the bio-processing component housing is configured or designed as a bioreactor or a filter with a flow path.
- 15 26. The bio-processing component kit of claim 25, wherein the bio-processing component housing comprises a container packed with material for product separation.
- 20 27. A method of fabricating an aseptic peripheral connection assembly for installing a sterilized peripheral in a bio-processing component via an aseptic connector affixed to the component, wherein the aseptic peripheral connection assembly comprises (i) a carrier disposed within the applicator sleeve, the carrier comprising the sterilized peripheral and a sealing member configured to form a leak-tight seal with the aseptic connector on the bio-processing component at a location where the carrier is to be installed; and (ii) an applicator comprising an applicator sleeve and an applicator connector, the applicator connector comprising a component-facing opening configured to connect to the aseptic connector on the bio-processing component, the method comprising:
- 25 a. placing the carrier in the applicator;
 - b. packaging the aseptic peripheral connection assembly in a hermetically sealed package; and

- c. sterilizing the aseptic peripheral connection assembly, wherein the sterilizing does not employ exposure to radiation at a level of greater than about 15 kGy.

5 28. The method of claim 27, further comprising, prior to packaging, plasma cleaning the carrier in the applicator.

29. The method of claim 27, further comprising adhering a hermetic sealing tab on the component-facing opening of the applicator connector.

10 30. A method of using an aseptic peripheral connection assembly for installing a sterilized peripheral in a bio-processing component via an aseptic connector affixed to the component, wherein the aseptic peripheral connection assembly comprises (i) a carrier comprising the peripheral and a sealing member configured to form a leak-tight seal with the aseptic connector on the bio-processing component at a location where the carrier is to be installed; and
15 (ii) an applicator comprising a sleeve and an applicator connector comprising a component-facing opening, the method comprising:

- a. connecting the applicator connector of the applicator to the aseptic connector on the bio-processing component;
- b. removing a hermetic sealing tab covering the component-facing opening of the applicator connector; and
- 20 c. plunging the carrier from a position within the applicator sleeve to a position engaged with the aseptic connector, and forming the leak-tight seal therewith.

25 31. The method of claim 30, further comprising clamping the applicator connector to the aseptic connector.